



May 14, 2020

Skeletal Dynamics, Inc.
Ana M. Escagedo
President
7300 N. Kendall Drive, Suite 400
Miami, Florida 33156

Re: K200367

Trade/Device Name: Distal Humerus Plating System
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories
Regulatory Class: Class II
Product Code: HRS, HWC
Dated: February 13, 2020
Received: February 14, 2020

Dear Ana Escagedo:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, MPH
Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K200367

Device Name

Distal Humerus Plating System

Indications for Use (Describe)

The Skeletal Dynamics Distal Humerus Plating System is indicated for intra-articular fractures of the distal humerus, comminuted supracondylar fractures, osteotomies, and non-unions of the distal humerus.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human
Services Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY
Skeletal Dynamics Inc.'s
Distal Humerus Plating System

Submitter

Skeletal Dynamics, Inc.
7300 N. Kendall Drive
Suite 400
Miami, FL 33156
Phone: 305-596-7585
Facsimile: 305-596-7591
Contact Person: Ana M. Escagedo
Date Prepared: March 29, 2020

Name and Classification

Name: Distal Humerus Plating System
Common Name: Plate, Fixation, Bone
Classification: 21 CFR §888.3030
Regulatory Class: Class II
Product Code: HRS

Predicate Device

Synthes Variable Angle LCP Elbow System (Medial and Posterolateral Distal Humerus Plates), K120717

Additional Predicate:

Synthes 3.5mm LCP Distal Humerus System, K033995

Device Description

The Skeletal Dynamics Distal Humerus Plating System consists of five titanium bone plates in different shapes (medial direct, lateral direct, lateral helical, lateral capitellar, and supracondylar), in multiple lengths, and in left and right configurations to address distal humerus fractures. The plates are made of medical grade titanium alloy. Locking and compression screws in 2.7mm, 3.5mm and 4.5mm and 3.0mm polyaxial locking screws are provided in various lengths. Also included in the system is specialized instrumentation. The System is provided non-sterile for sterilization in the user facility.

The Distal Humerus Plating System is comprised of:

- Titanium alloy plates, screws and fasteners
- CoCr Cannulated Polyaxial Locking Screws (PLS)
- Stainless steel K-wires (for provisional fixation; not for implantation)
- System specific instrumentation

Indications for Use

The Distal Humerus Plating System is indicated for intra-articular fractures of the distal humerus, comminuted supracondylar fractures, osteotomies, and non-unions of the distal humerus.

Summary of Technological Characteristics

The substantial equivalence of the Distal Humerus Plating System to the predicate device is demonstrated by similarities in intended use, indications for use, materials, design (fundamental scientific technology), performance, sterility and packaging and does not present any new issues of safety or effectiveness.

Performance Testing

Engineering analysis and mechanical testing demonstrated that the Skeletal Dynamics Distal Humerus Plating System is equivalent to predicate devices currently marketed. Mechanical testing which established equivalency included ASTM F382, Standard Specification and Test Methods for Metallic Bone Plates, and ASTM F543, Standard Specification and Test Methods for Metallic Bone Screws. Therefore, the subject device is as safe and effective as the legally marketed predicate device.

Conclusions

The Skeletal Dynamics Distal Humerus Plating System is substantially equivalent to the predicate device identified in this premarket notification.