



October 14, 2022

ZS Fab Inc.
% Karen Warden, PhD
President
BackRoads Consulting Inc.
PO Box 566
Chesterland, Ohio 44026

Re: K221858

Trade/Device Name: ZSFab Lumbar Interbody System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: MAX
Dated: September 14, 2022
Received: September 15, 2022

Dear Dr. Warden:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for
Brent Showalter, Ph.D.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K221858

Device Name

ZSFab Lumbar Interbody System

Indications for Use (Describe)

The ZSFab Lumbar Interbody System is intended for lumbar interbody fusion. The devices are indicated for use at one or two contiguous levels in the lumbar spine from L2-S1, in skeletally mature patients who have had at least six months of non-operative treatment. The ZSFab Lumbar Interbody System is indicated to treat lumbar degenerative disc disease (DDD) with up to Grade 1 spondylolisthesis or retrolisthesis at the involved spinal level(s). DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by imaging studies (radiographs, CT, MRI). Additionally, the ZSFab Lumbar Interbody System can be used as an adjunct to fusion in patients diagnosed with degenerative scoliosis. The ZSFab lumbar Interbody System is designed for use with autogenous and/or allogeneic bone graft comprised of cancellous and/or corticocancellous bone and/or demineralized allograft bone with bone marrow aspirate to facilitate fusion and to be used with supplemental fixation cleared for use in the lumbosacral spine.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

| | |
|---|---|
| Date: | 14 Spetember 2022 |
| Sponsor: | ZSFab Inc. 96 Clematis Ave, Suite 2F Waltham, MA 02453 Office: 617.468.8665 |
| Sponsor Contact: | Xuewei Ma, R&D Director |
| 510(k) Contact: | Karen E. Warden, PhD BackRoads Consulting Inc. PO Box 566 Chesterland, OH 44026 Office: 440.729.8457 |
| Proposed Trade Name: | ZSFab Lumbar Interbody System |
| Common Name: | Lumbar interbody fusion device |
| Device Classification: | Class II |
| Regulation Names, Regulation Numbers, Product Codes: | Intervertebral fusion device with bone graft, lumbar, 888.3080, MAX |
| Device Description: | The ZSFab Lumbar Interbody System includes additively manufactured interbody fusion devices for lumbar implantation. The implants are designed as a solid frame that includes lattice structures to provide surgical stabilization of the spine. The lattices have near-elliptical pores with axis lengths of 610-1000µm. The endplates are featured with teeth design and stochastic lattice structures with average pore size of 620-710µm. Each lumbar interbody has a central cavity for bone graft material. The implants are available in a variety of height, length, width and lordotic angulation combinations to accommodate the patient specific anatomy and clinical circumstances. The implants are sold sterile. |
| Indications for Use: | The ZSFab Lumbar Interbody System is intended for lumbar interbody fusion. The devices are indicated for use at one or two contiguous levels in the lumbar spine from L2-S1, in skeletally mature patients who have had at least six months of non-operative treatment. The ZSFab Lumbar Interbody System is indicated to treat lumbar degenerative disc disease (DDD) with up to Grade 1 spondylolisthesis or retrolisthesis at the involved spinal level(s). DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by imaging studies (radiographs, CT, MRI). Additionally, the ZSFab Lumbar Interbody System can be used as an adjunct to fusion in patients diagnosed with degenerative scoliosis. The ZSFab lumbar Interbody System is designed for use with autogenous and/or allogeneic bone graft comprised of cancellous and/or corticocancellous bone and/or demineralized allograft bone with bone marrow aspirate to facilitate fusion and to be used with supplemental fixation cleared for use in the lumbosacral spine. |
| Materials: | The ZSFab Lumbar Interbody System implants are manufactured from Ti-6Al-4V ELI titanium alloy (ASTM F3001). |
| Primary Predicate: | Cascadia™ Interbody System (K2M Inc. – K172009) |
| Additional Predicates: | ZSFab Cervical Interbody System (ZSFab Inc. – K202488), Medtronic ARTiC-L™ 3D Ti Spinal System with TiONIC™ Technology (Medtronic Sofamor Danek USA, Inc – K190959) |

| | |
|---------------------------------------|---|
| Performance Data: | <p>Mechanical testing of the worst case ZSFab Lumbar Interbody System implant included static and dynamic axial compression and static and dynamic compression shear according to ASTM F2077. In addition, subsidence according to ASTM F2267 and expulsion tests were performed. The mechanical test results demonstrate that the ZSFab Lumbar Interbody System performance is substantially equivalent to the predicate devices.</p> |
| Technological Characteristics: | <p>The ZSFab Lumbar Interbody System possesses the same technological characteristics as one or more of the predicate devices. These include:</p> <ul style="list-style-type: none">• intended use (as described above)• basic design (additively manufactured structure),• material (titanium alloy) and• sizes (dimensions are comparable to those offered by the predicate systems) <p>The ZSFab Lumbar Interbody System is the same as previously cleared devices.</p> |
| Conclusion: | <p>The ZSFab Lumbar Interbody System possesses the same intended use and technological characteristics as the predicate devices. Therefore the ZSFab Lumbar Interbody System is substantially equivalent for its intended use.</p> |