



August 25, 2022

Hebei Astro Medical Supply Co., Ltd.  
% Diana Hong  
General Manager  
Mid-Link Consulting Co.,Ltd.  
P.O.box 120-119  
Shanghai, 200120  
China

Re: K221143

Trade/Device Name: Nitrile Exam Gloves (Blue, Black)  
Regulation Number: 21 CFR 880.6250  
Regulation Name: Non-Powdered Patient Examination Glove  
Regulatory Class: Class I, reserved  
Product Code: LZA  
Dated: July 25, 2022  
Received: July 25, 2022

Dear Diana Hong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Clarence W. Murray III, Ph.D.  
Assistant Director  
DHT4B: Division of Infection Control  
and Plastic Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K221143

Device Name

Nitrile Exam Gloves (Blue, Black)

Indications for Use (Describe)

Nitrile Exam Gloves (Blue, Black) is a disposable device intended for medical purpose that is worn on the examiner's hands to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

This 510(k) Summary is being submitted in accordance with requirements of Title 21, CFR Section 807.92.

The assigned 510(k) Number: K221143

1. Date of Preparation: 8/23/2022

2. Sponsor Identification

**Hebei Astro Medical Supply Co., Ltd.**

Address: East of Xiaoxixian, West of Jingsan Street, South of Weiwu Road, North of Weiqi Road,  
Jinzhou Economic Development Zone, Hebei Province, P.R, China, 052260

Establishment Registration Number: 3015537296

Contact Person: Ning Zheng

Position: General Manager

Tel: +86-311-85125369

Email: [erin@wallyplastic.net](mailto:erin@wallyplastic.net)

3. Designated Submission Correspondent

Ms. Diana Hong (Primary Contact Person)

Ms. Jinlei Tang (Alternative Contact Person)

**Mid-Link Consulting Co., Ltd**

P.O. Box 120-119, Shanghai, 200120, China

Tel: +86-21-2281-5850,

Fax: 360-925-3199

Email: [info@mid-link.net](mailto:info@mid-link.net)

4. Identification of Proposed Device

Trade Name: Nitrile Exam Gloves (Blue, Black)

Common Name: POWDER FREE NITRILE EXAMINATION GLOVES

**Regulatory Information**

Classification Name: polymer patient examination glove

Classification: I;

Product Code: LZA;  
 Regulation Number: 21CFR 880.6250  
 Review Panel: General Hospital;

Indication for Use:

Nitrile Exam Gloves (Blue, Black) is a disposable device intended for medical purpose that is worn on the examiner's hands to prevent contamination between patient and examiner.

Device Description

The proposed device is a power free medical glove. The device is provided in blue and black. The device meets the requirements of *ASTM D6319-19: Standard specification for Nitrile Examination Gloves for Medical Application*. The proposed gloves are available in six sizes, which are XS, S, M, L, XL, XXL, it could be selected by the user depended on size of hand. The different between each size is just in the dimension. The proposed device is provided in non-sterile.

5. Identification of Predicate Device

510(k) Number: K211457

Product Name: Powder Free Nitrile Examination Gloves, (Blue, Black, Indigo)

6. Summary of Technological characteristics

Table 1 Comparison of Technology Characteristics

ITEM	Proposed Device	Predicate Device K211457	Comparison
Product Code	LZA	LZA	Same
Regulation Number	21 CFR 880.6250	21 CFR 880.6250	Same
Class	I	I	Same
Indication for use	Nitrile Exam Gloves (Blue, Black) is a disposable device intended for medical purpose that is worn on the examiner's hands to prevent contamination between patient and examiner.	The Powder Free Nitrile Examination Gloves is a disposable device intended for medical purpose that is worn on the examiner's hands to prevent contamination between patient and examiner.	Same
Material	Nitrile	Nitrile	Same
Color	Blue, Black	Blue, Black, Indigo	Different
Sterility	Non-sterile	Non-sterile	Same
Single-use	Yes	Yes	Same
Surface Treatment /Powder or Powder Free	Powder Free	Powder Free	Same

Size	XS, S, M, L, XL, XXL		XS, S, M, L, XL		Different	
Dimensions (ASTM D6319-19)	Width					
	XS	75±5mm	XS	70±10mm		
	S	85±5mm	S	80±10mm		
	M	95±5mm	M	95±10mm		
	L	105±5mm	L	110±10mm		
	XL	115±5mm	XL	120±10mm		
	XXL	125±5mm	/	/		
	Length					
	XS	230mm min	XS	230mm min		
	S	230mm min	S	230mm min		
	M	230mm min	M	230mm min		
	L	230mm min	L	230mm min		
	XL	230mm min	XL	230mm min		
	XXL	230mm min	/	/		
	Thickness					
Palm	0.05mm min	Palm	0.08-0.10 mm			
Finger	0.05mm min	Finger	0.10-0.12 mm			
Physical Properties (ASTM D6319-19 and ASTM D412-16)	Before Aging					Same
	Tensile Strength	14MPa min	Tensile Strength	14MPa min		
	Ultimate Elongation	500% min	Ultimate Elongation	500% min		
	After Aging					
	Tensile Strength	14MPa min	Tensile Strength	14MPa min		
	Ultimate Elongation	400% min	Ultimate Elongation	400% min		
Power free residue (ASTM D6319-19 and ASTM D6124-17)	Less than 2mg per glove		Less than 2mg per glove		Same	
Freedom from Holes (ASTM D5151-19)	Meet AQL 2.5 with G1		Meet AQL 1.5 with G1		Different	
Biocompatibility						
Sensitization	Under the conditions of study, not a sensitizer		Under the conditions of study, not a sensitizer		Different	
Intracutaneous Reactivity	Under the conditions of study, not an irritant		Under the conditions of study, not an irritant			
System Toxicity	Under the conditions of study, non-system toxicity		Not tested			
In Vitro Cytotoxicity	Not tested		Under the conditions of study, noncytotoxic			

Different - Color

The proposed device is provided in two colors, blue and black, and the predicate device is provided

in three colors, blue, black and indigo. The color of the proposed device can be covered by the predicate device.

#### Different - Size & Dimensions

The size and dimension of the proposed device is not exactly same as the predicate device. The user can select appropriate model depended on size of user's hand.

#### Different - Freedom from Holes

The freedom from holes of the proposed device is different from predicate device. The proposed device meets AQL 2.5 with G1, while the predicate device meets AQL 1.5 with G1.

#### Different - Biocompatibility

The biocompatibility test item of the proposed device is different from the predicate device. The proposed device was conducted for systemic toxicity and not for cytotoxicity. While the predicate device was conducted for cytotoxicity and not for systemic toxicity.

### 7. Summary of Non-Clinical Test

Non-clinical tests were conducted to verify that the proposed device met all design specifications as was same/similar to the predicate device. The test results demonstrated that the Nitrile Exam Gloves (Blue, Black) complies with the following standards:

- ASTM D6319-19 Standard Specification for Nitrile Examination Gloves for Medical Application
- ASTM D5151-19 Standard Test Method for Detection of Holes in Medical Gloves
- ASTM D3767-03 (2020) Standard Practice for Rubber-Measurement of Dimensions
- ASTM D412-16 Standard Test Methods for Vulcanized Rubber and Thermoplastic Elastomers-Tension
- ASTM D6124-06 (Reapproved 2017) Standard Test Method for Residual Powder on Medical Gloves
- ISO 10993-10:2010 Biological Evaluation of Medical Devices-Part 10: Tests for Irritation and Skin Sensitization;
- ISO 10993-11:2017 Biological evaluation of medical devices-Part 11: Tests for systemic

toxicity;

Table 2 Summary of Performance Testing

Name of the Test Methodology/ Standard	Purpose	Acceptance Criteria	Results																								
ASTM D5151-19 Standard Test Method for Detection of Holes in Medical Gloves	The test was performed in accordance with ASTM D5151-19 Standard Test Method for Detection of Holes in Medical Gloves to evaluate the detection of holes in medical gloves.	Gloves are free from water droplets, stream, or other types of water leakage.	Requirement met																								
ASTM D3767-03 (2020) Standard Practice for Rubber – Measurement of Dimensions	The test was performed in accordance with ASTM D3767-03 (2020) Standard Practice for Rubber – Measurement of Dimensions to evaluate the geometrical dimension of rubber products and specimens for physical tests.	The measurement results shall conform to the minimum values specified in the table below. <table border="1" data-bbox="853 1099 1262 1608"> <thead> <tr> <th colspan="2">Width</th> </tr> </thead> <tbody> <tr> <td>XS</td> <td>75±5mm</td> </tr> <tr> <td>S</td> <td>85±5mm</td> </tr> <tr> <td>M</td> <td>95±5mm</td> </tr> <tr> <td>L</td> <td>105±5mm</td> </tr> <tr> <td>XL</td> <td>115±5mm</td> </tr> <tr> <td>XXL</td> <td>125±5mm</td> </tr> <tr> <td>Length</td> <td>230mm min</td> </tr> <tr> <th colspan="2">Thickness</th> </tr> <tr> <td>Palm</td> <td>0.05mm min</td> </tr> <tr> <td>Finger</td> <td>0.05mm min</td> </tr> <tr> <td></td> <td></td> </tr> </tbody> </table>	Width		XS	75±5mm	S	85±5mm	M	95±5mm	L	105±5mm	XL	115±5mm	XXL	125±5mm	Length	230mm min	Thickness		Palm	0.05mm min	Finger	0.05mm min			Requirement met
Width																											
XS	75±5mm																										
S	85±5mm																										
M	95±5mm																										
L	105±5mm																										
XL	115±5mm																										
XXL	125±5mm																										
Length	230mm min																										
Thickness																											
Palm	0.05mm min																										
Finger	0.05mm min																										
ASTM D412-16 Standard Test Methods for Vulcanized Rubber and Thermoplastic Elastomers—Tension	The test was performed in accordance with ASTM D412-16 Standard Test Methods for Vulcanized Rubber and Thermoplastic Elastomers—Tension to evaluate the	The measurement results shall conform to the minimum values specified in the table below Before Aging: <table border="1" data-bbox="853 1787 1257 1951"> <tbody> <tr> <td>Tensile Strength</td> <td>14MPa min</td> </tr> <tr> <td>Ultimate Elongation</td> <td>500% min</td> </tr> </tbody> </table> After Aging:	Tensile Strength	14MPa min	Ultimate Elongation	500% min	Requirement met																				
Tensile Strength	14MPa min																										
Ultimate Elongation	500% min																										



	tensile (tension) properties of vulcanized thermoset rubbers and thermoplastic elastomers.	Tensile Strength	14MPa min	
		Ultimate Elongation	400% min	
ASTM D6124-06 Standard Test Method for Residual Powder on Medical Gloves	The test was performed in accordance with ASTM D6124-06 Standard Test Method for Residual Powder on Medical Gloves to evaluate the amount of residual powder (or filter- retained mass) found on medical gloves.	Less than 2mg per glove		Requirement met

Table 3 Summary of Biocompatibility Testing

Test Methodology	Purpose	Acceptance Criteria	Result
System Toxicity	The test was performed in accordance with ISO 10993-11 Biological evaluation of medical devices — Part 11: Tests for systemic toxicity to evaluate the systemic toxicity of the test sample.	Non-system toxicity	Under the conditions of the study, the proposed device was non-system toxicity.
Sensitization	The test was performed in accordance with ISO 10993-10 Third Edition 2010-08-01 Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization to evaluate the	Non-sensitizing	Under the conditions of the study, the proposed device was non-sensitizing.

	sensitization of the test sample.		
Intracutaneous Reactivity	The test was performed in accordance with ISO 10993-10 Third Edition 2010-08-01 Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization to evaluate the irritation of the test sample.	Non-irritating	Under the conditions of the study, the proposed device was non-irritating.

8. Clinical Test Conclusion

No clinical study is included in this submission.

9. Conclusion

The conclusions drawn from the nonclinical tests demonstrate that the proposed subject device is as safe, as effective, and performs as well as or better than the legally marketed predicate device K211457.