



June 28, 2022

Xingyu Medical Tech Co., Ltd
% Jarvis Wu
Consultant
Shanghai Sungo Management Consulting Company Limited
14th Floor, 1500# Central Avenue
Shanghai, 200122
China

Re: K221273

Trade/Device Name: Medical Face Mask (M001)
Regulation Number: 21 CFR 878.4040
Regulation Name: Surgical Apparel
Regulatory Class: Class II
Product Code: FXX
Dated: May 2, 2022
Received: May 2, 2022

Dear Jarvis Wu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent Showalter, Ph.D.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K221273

Device Name

Medical Face Mask (M001)

Indications for Use (Describe)

The Medical Face Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This a single use, disposable device(s), provided non-sterile.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Xingyu Medical Tech Co., Ltd.
No. 2189 Yaoqian Road, Gaomi Economic Development Zone, Weifang City, Shandong Province, China

510(k) Summary

K221273

Document Date Prepared: 2022/4/21

A. Applicant:

Name: Xingyu Medical Tech Co., Ltd.

Address: No. 2189 Yaoqian Road, Gaomi Economic Development Zone, Weifang City, Shandong Province, China

Contact Person: Carrie Li

Tel: 0086-18263618867

Mail: carrie@xingyugloves.com

Submission Correspondent:

Primary contact: Mr. Jarvis Wu

Shanghai SUNGO Management Consulting Co., Ltd.

14th floor, 1500# Century Ave., Shanghai 200122, China

Tel: +86-21-58817802

Email: haiyu.wang@sungoglobal.com

Secondary contact: Mr. Raymond Luo

14th floor, 1500# Century Ave., Shanghai 200122, China

Tel: +86-21-68828050

Email: fda.sungo@gmail.com

B. Device:

Proprietary Name: Medical Face Mask

Common Name: Surgical Face Mask

Model(s): M001

Regulatory Information

Classification Name: Surgical Face Mask

Classification: Class II

Product code: FXX

Xingyu Medical Tech Co., Ltd.

No. 2189 Yaoqian Road, Gaomi Economic Development Zone, Weifang City, Shandong Province, China

Regulation Number: 878.4040

Review Panel: Surgical Apparel

C. Predicate device:

510K	Device name	ASTM F2100-19 level	Manufacturer
K211631	Surgical Face Mask	Level 2 & 3	Macheng thimble Technology Investment Co., Ltd

(Note: Predicate device has NOT been subject to any Medical Device Recalls, including design-related recall.)

D. Indications use of the device:

The Medical Face Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This a single use, disposable device(s), provided non-sterile.

E. Device Description:

The Proposed Device Medical Face Mask (Model: M001) is available in white, blue and black colors, and flat pleated type mask, utilizing ear loops way for wearing, and it has Nose clips design for fitting the face mask around the nose.

The proposed device is manufactured with three layers, the inner and outer layers are made of spun-bond polypropylene, and the middle layer is made of melt blown polypropylene filter. The model of proposed device, ear loops, is held in place over the user's mouth and nose by two elastic ear loops welded to the face mask. The elastic ear loops are made of Polyester and Spandex. The nose piece contained in the proposed device is in the layers of face mask to allow the user to fit the face mask around their nose, which is made of Metal Core Plastic (Iron wire & Polypropylene). The proposed device(s) are sold non-sterile and are intended to be single use, disposable device.

F. Comparison with predicate device

Table 1 General Comparison

Device	Proposed Device	Predicate device	Comparison
Manufacturer	Xingyu Medical Tech Co., Ltd.	Macheng thimble Technology Investment Co., Ltd	-

Xingyu Medical Tech Co., Ltd.

No. 2189 Yaoqian Road, Gaomi Economic Development Zone, Weifang City, Shandong Province, China

510K number	K221273	K211631	-	
Device name	Medical Face Mask	Surgical Face Mask	-	
Classification	Class II Device, FXX (21 CFR878.4040)	Class II Device, FXX (21 CFR878.4040)	Same	
Indications for use	The Medical Face Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This a single use, disposable device(s), provided non-sterile.	The Surgical Face Masks are intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This a single use, disposable device(s), provided non-sterile.	Same	
Material	Outer layer	Non-woven fabric	Spun-bond polypropylene	Same
	Middle layer	Melt-blown Non-woven fabric	Melt blown polypropylene filter	Same
	Inner layer	Non-woven fabric	Spun-bond polypropylene	Same
	Nose clip	Metal Core Plastic (Iron wire & Polypropylene)	Malleable aluminum wire	Different
	Ear loops	Polyester and Spandex	Polyester	Different
Color	White, Blue, Black	Blue	Different	
Dimension (length)	175mm+/-5mm	17.5cm+/-0.5cm	Same	
Dimension (Width)	95mm+/-5mm	9.5cm+/-0.5cm	Same	
OTC use	Yes	Yes	Same	

Xingyu Medical Tech Co., Ltd.

No. 2189 Yaoqian Road, Gaomi Economic Development Zone, Weifang City, Shandong Province, China

Sterility	Non-Sterile	Non-Sterile	Same
Use	Single Use, Disposable	Single Use, Disposable	Same
ASTM F2100-19 level	Level 2 & 3	Level 2 & 3	Same
Biocompatibility	Meet ISO10993 ,proved non-cytotoxicity, non-irritating and non-sensitizing	Meet ISO10993 ,proved non-cytotoxicity, non-irritating and non-sensitizing	Same

From the comparison we found the material of proposed device's nose clip and Ear loops, as well as its color, are different from the predicate device. The biocompatibility tests were conducted to both components to ensure their compliance to the ISO10993-5 and ISO10993-10. There is no new risk generated from the difference of the material.

G. Non-Clinical Test Conclusion

Non-clinical tests were conducted to verify that the proposed device met all design specifications as was similar to the predicate device. The test results demonstrated that the proposed device complies with the following standards and the requirements stated in the Guidance for Industry and FDA Staff: Surgical Masks – Premarket Notification [510(k)] Submission issued on March 5, 2004:

- ISO 10993-5: 2009 Biological Evaluation of Medical Devices -- Part 5: Tests For In Vitro Cytotoxicity
- ISO 10993-10: 2010 Biological Evaluation of Medical Devices - Part 10: Tests For Irritation And Skin Sensitization
- ASTM F2100, Standard Specification for Performance of Materials Used In Medical Face Masks
- ASTM F1862, Standard Test Method for Resistance of Medical Face Masks To Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume At A Known Velocity);
- EN 14683, Medical Face Masks—Requirements and Test Methods;
- ASTM F2101, Standard Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) Of Medical Face Mask Materials, Using A Biological Aerosol of Staphylococcus Aureus;
- ASTM F2299, Stand test method for determining the initial efficiency of materials used in medical face masks to penetration by particulates using latex spheres;
- 16 CFR 1610, Standard for the Flammability of clothing textiles;

Table 2 - Performance Testing

Item	Purpose	Proposed device	Acceptance Criteria	Result
------	---------	-----------------	---------------------	--------

<p>Fluid Resistance Performance ASTM F1862</p>	<p>Assess the performance of a mask to resistance to a synthetic blood preparation targeted toward the mask at a set pressure</p>	<p>32 out of 32 per lot pass at 120 mmHg for 3 non-consecutive lots tested;</p> <p>32 out of 32 per lot pass at 160 mmHg for 3 non-consecutive lots tested;</p>	<p>29 out of 32 pass at 120 mmHg for Level 2</p> <p>29 out of 32 pass at 160 mmHg for Level 3</p>	<p>PASS</p>
<p>Particulate Filtration Efficiency ASTM F2299</p>	<p>Assess the performance of a mask to penetration by sub-micron polystyrene latex particles of 0.1 micron</p>	<p>3 non-consecutive lots tested, using a sample size of 32/lot.</p> <p>Average Result- Level 2 : 99.2% Level 3 : 99.4%</p>	<p>$\geq 98\%$</p>	<p>PASS</p>
<p>Bacterial Filtration Efficiency ASTM F2101</p>	<p>Assess the performance of a mask to penetration by a prepared solution with known concentration of an indicator bacterial organism</p>	<p>3 non-consecutive lots tested, using a sample size of 32/lot.</p> <p>Average Result- Level 2 : 99.2% Level 3 : 99.4%</p>	<p>$\geq 98\%$</p>	<p>PASS</p>
<p>Differential Pressure (Delta P) EN 14683 Annex C</p>	<p>Assess the performance of a mask for resistance to air movement through the</p>	<p>3 non-consecutive lots tested, using a sample size of 32/lot.</p> <p>Average Result-</p>	<p>$< 6.0\text{mmH}_2\text{O}/\text{cm}^2$</p>	<p>PASS</p>

Xingyu Medical Tech Co., Ltd.

No. 2189 Yaoqian Road, Gaomi Economic Development Zone, Weifang City, Shandong Province, China

	materials of the face of the mask	Level 2 : 4.4 mmH ₂ O/cm ² Level 3 : 5.1 mmH ₂ O/cm ²		
Flammability 16 CFR 1610	Assess the resistance of a mask to ignition	3 non-consecutive lots tested, using a sample size of 32/lot. Average Result- Level 2 : Class 1 Level 3 : Class 1	Class 1	PASS

Table 3 – Biocompatibility Testing

Test Method	Purpose	Acceptance Criteria	Result
Cytotoxicity	Assess the potential risk of Cytotoxicity of mask material	Non-Cytotoxic	PASS Under the conditions of the study, the device is non-cytotoxic.
Irritation	Assess the potential risk of Irritation of mask material	Non-Irritating	PASS Under the conditions of the study, the device is non-irritating.
Sensitization	Assess the potential risk of Sensitization of mask material	Non-Sensitizing	PASS Under the conditions of the study, the device is non-sensitizing

H. Clinical Test Conclusion

No clinical study is included in this submission.

I. Conclusion

The conclusion drawn from the nonclinical tests demonstrates that the subject device in 510(K) submission, the Medical Face Mask (Model: M001) is as safe, as effective, and performs as well as or better than the legally marketed predicate device cleared under K211631.