



May 25, 2021

Boston Scientific Corporation  
Melissa Klamerus  
Principal Regulatory Affairs Specialist  
4100 Hamline Ave North  
St. Paul, Minnesota 55112

Re: K210608

Trade/Device Name: LUX-Dx Insertable Cardiac Monitor

Regulation Number: 21 CFR 870.1025

Regulation Name: Arrhythmia Detector And Alarm (Including ST-Segment Measurement And Alarm)

Regulatory Class: Class II

Product Code: MXD

Dated: February 24, 2021

Received: March 1, 2021

Dear Melissa Klamerus:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Jennifer Shih Kozen  
Assistant Director  
Division of Cardiac Electrophysiology,  
Diagnostics and Monitoring Devices  
Office of Cardiovascular Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K210608

Device Name

LUX-Dx™ Insertable Cardiac Monitor

Indications for Use (Describe)

The LUX-Dx Insertable Cardiac Monitor (ICM) is intended to monitor and record subcutaneous electrocardiogram (S-ECG). The recorded S-ECG is used for the clinical evaluation and diagnosis of cardiac arrhythmias. The LUX-Dx is indicated for use in patients that have a known heart condition and are at risk of developing an abnormal heart rhythm, or have symptoms that may suggest a cardiac arrhythmia such as dizziness, palpitations, syncope, chest pain, and/or shortness of breath.

The LUX-Dx has not been tested specifically for pediatric use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



4100 Hamline Avenue North  
St Paul, MN 55112-5798  
651.582.4000 Tel  
651.582.5134 Fax  
www.bostonscientific.com

## 510(k) Summary

### 1. Submitter

Boston Scientific Corporation  
4100 Hamline Avenue North  
St. Paul, Minnesota 55112-5798

Contact: Melissa Klamerus  
Principal Regulatory Specialist  
Phone: (612) 247-6878  
Email: melissa.klamerus@bsci.com

Date Prepared: 23 February 2021

### 2. Device

Trade Name: LUX-Dx™ Insertable Cardiac Monitor  
Common Name: Arrhythmia detector and alarm  
Product Code/Panel: MXD, Cardiovascular  
Device Class and Panel: Class II  
Classification Regulation: 21 CFR 870.1025

### 3. Predicate Device

Trade Name: LUX-Dx™ Insertable Cardiac Monitor  
Manufacturer: Boston Scientific Corp  
Clearance Number: K193473, 26 June 2020  
Common Name: Arrhythmia detector and alarm  
Product Code/Panel: MXD, Cardiovascular  
Device Class and Panel: Class II  
Classification Regulation: 21 CFR 870.1025

### 4. Indication for Use

The LUX-Dx Insertable Cardiac Monitor (ICM) is intended to monitor and record subcutaneous electrocardiogram (S-ECG). The recorded S-ECG is used for the clinical evaluation and diagnosis of cardiac arrhythmias. The LUX-Dx is indicated for use in patients that have a known heart condition and are at risk of

developing an abnormal heart rhythm, or have symptoms that may suggest a cardiac arrhythmia such as dizziness, palpitations, syncope, chest pain, and/or shortness of breath.

The LUX-Dx has not been tested specifically for pediatric use.

## 5. Device Description

The LUX-Dx ICM sensor device (Model M301) evaluates S-ECG waveform data for indications of cardiac arrhythmias and “marks” the S-ECG signal for clinical presentation and evaluation when the algorithm criteria are met. The device is inserted into the subcutaneous layer of the fourth intercostal space of the left chest wall. The sensor device is powered by an integrated battery. The electrodes used for detecting the S-ECG signal are located on each end of the sensor device, in the header and at the base of the battery. The LUX-Dx system includes the following main components:

- LUX-Dx™ sensor device - a subcutaneously-implanted sensor device for cardiac arrhythmia event data collection and transmission. In addition, symptom events are collected and transmitted from the sensor device.
- Mobile Monitor (MM) - mobile applications (myLUX™ Patient app and LUX-Dx™ Clinic Assistant app) running on an OTS mobile device that communicates with the LUX-Dx sensor device (using Bluetooth Low Energy (BLE)) and the LATITUDE Clarity™ server (using cellular/Wi-Fi) for collection and transmission of event, patient, and device data.
- LATITUDE Clarity™ server - a server that communicates with the Mobile Monitor for bidirectional data transmission and provides web access for clinicians to perform remote monitoring activities and manage general patient and system parameters and workflow activities.
- System Accessories- for insertion of the sensor device, an insertion tool and incision tool are provided. In addition, a magnet is provided to initiate sensor/MM app communication.

## 6. Substantial Equivalence

Characteristic	Model M301 LUX-Dx ICM	Model M301 LUX-Dx ICM Patient Patch v1.00
<b>Indications for use</b>	<p>The LUX-Dx Insertable Cardiac Monitor (ICM) is intended to monitor and record subcutaneous electrocardiogram (S-ECG). The recorded S-ECG is used for the clinical evaluation and diagnosis of cardiac arrhythmias. The LUX-Dx is indicated for use in patients that have a known heart condition and are at risk of developing an abnormal heart rhythm, or have symptoms that may suggest a cardiac arrhythmia such as dizziness, palpitations, syncope, chest pain, and/or shortness of breath.</p> <p>The LUX-Dx has not been tested specifically for pediatric use.</p>	Identical
<b>Usage</b>	Single use	Identical

Characteristic	Model M301 LUX-Dx ICM	Model M301 LUX-Dx ICM Patient Patch v1.00
<b>Sterility</b>	Supplied sterile	Identical
<b>Volume Dimension</b>	1.2 cc	Identical
<b>MR Conditional</b>	1.5 T and 3T	Identical
<b>Energy Source</b>	Battery powered - LiMnO <sub>2</sub>	Identical
<b>Longevity</b>	3 years	Identical
<b>Device Software Version</b>	FW v 1.00.00	FW v 1.00.00 Patient Patch v1.00
<b>Remote Monitoring</b>	Yes	Identical
<b>Packaging</b>	Sterile kit including ICM device and custom implant tools. Accessories: External, non-sterile devices are packaged separately.	Identical

## 7. Summary of Performance Testing

Boston Scientific performed safety risk management activities and design verification and validation testing to demonstrate that the LUX-Dx sensor device software is substantially equivalent to the predicate device software. The system conforms to user needs and intended use and software requirements and design specifications.

Performance Testing - Bench:

- Software Verification
- System Verification
- Confirmatory Validation

## 8. Conclusion

Based on the intended use, fundamental technological characteristics, and performance testing, the proposed LUX-Dx sensor device software has been shown to be appropriate for its intended use and is considered to be substantially equivalent to the predicate device software.