



August 19, 2021

WYZ Medical Supplies LLC
% Ivy Wang
Consultant
Shanghai Sungo Management Consulting Company Limited
13th Floor, 1500# Central Avenue
Shanghai, Shanghai 200122
China

Re: K211576
Trade/Device Name: WYZ Surgical Mask
Regulation Number: 21 CFR 878.4040
Regulation Name: Surgical Apparel
Regulatory Class: Class II
Product Code: FXX
Dated: May 21, 2021
Received: May 21, 2021

Dear Ivy Wang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Clarence W. Murray, III, PhD
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K211576

Device Name

WYZ Surgical Mask

Indications for Use (Describe)

The WYZ Surgical Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This a single use, disposable device(s), provided non-sterile.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) Summary

K211576

Document prepared date: 7/30/2021

A. Applicant:

Name: WYZ Medical Supplies LLC
Address: 5913 Imperial Loop, College Station, TX 77845, U.S.A.
Contact Person: Daniel Chen
Tel: +1 512-942-7893
Mail: daniel@wyzmedical.com

Submission Correspondent:

Primary contact: Ivy Wang

Shanghai SUNGO Management Consulting Co., Ltd.

Room 1409, Dongfang Building, 1500# Century Ave., Shanghai 200122, China

Tel: +86-21-58817802

Email: haiyu.wang@sungoglobal.com

Secondary contact: Mr. Raymond Luo

1500# Century Ave., Shanghai 200122, China

Tel: +86-21-68828050

Email: fda.sungo@gmail.com

B. Device:

Trade Name: WYZ Surgical Mask

Common Name: Surgical Face Mask

Model(s): WYZ201

Regulatory Information

Classification Name: Surgical Face Mask

Classification: Class II

Product code: FXX

Regulation Number: 878.4040

Review Panel: Surgical Apparel

C. Predicate device:

510K	Device name	Manufacturer
K153496	Disposable Surgical Face Masks	Xiantao Rayxin Medical Products Co., Ltd.

D. Indications for use of the device:

The WYZ Surgical Masks (model: WYZ201) are intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and

body fluids. This a single use, disposable device(s), provided non-sterile.

E. Device Description:

The proposed device (model: WYZ201) is blue color, and Flat Pleated type mask, utilizing Ear Loops way for wearing, and it has Nose clips design for fitting the face mask around the nose.

The proposed devices are manufactured with three layers, the inner and outer layers are made of Polypropylene (PP) spun-bond, and the middle layer is made of Polypropylene (PP) melt-blown. The WYZ201 model of proposed device, ear loops, is held in place over the users’ mouth and nose by two elastic ear loops welded to the face mask. The elastic ear loops are made of Elastic film based on styrene block copolymer (SBC) with polyolefin (PO) outer layer. Hotmelt based on styrene-isoprene block copolymers (SIS) Nonwoven on both sides based on polypropylene (PP) spun-bond. The nose piece contained in the proposed device(s) is in the layers of face mask to allow the user to fit the face mask around their nose, which is made of Galvanized iron wire with PP+PE coating.

The proposed devices are sold non-sterile and are intended to be single use, disposable device.

F. Comparison with predicate device/

Table 1 General Comparison

Device	Proposed Device	Predicate Device 1	Comparison
Manufacturer	WYZ Medical Supplies LLC	Xiantao Rayxin Medical Products Co., Ltd.	-
510K number	-	K153496	-
Classification	Class II Device, FXX (21 CFR878.4040)	Class II Device, FXX (21 CFR878.4040)	Same
Indications for use	The WYZ Surgical Mask (model: WYZ201) are intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This a single use, disposable device(s), provided non-sterile.	The Disposable Surgical Face Masks are intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single use, disposable device(s), provided non-sterile.	Similar

Style		Ear Loops, Flat Pleated, 3 layers	Ear Loops, Flat Pleated, 3 layers	Same
Material	Outer layer	Polypropylene (PP) spun-bond	Spun-bond polypropylene	Same
	Middle layer	Polypropylene (PP) melt-blown	Melt blown polypropylene filter	Same
	Inner layer	Polypropylene (PP) spun-bond	Spun-bond polypropylene	Same
	Nose clip	Galvanized iron wire with PP+PE coating	Malleable aluminum wire	Analysis
	Ear loops	Elastic film based on styrene block copolymer (SBC) with polyolefin (PO) outer layer. Hotmelt based on styrene-isoprene block copolymers (SIS) Nonwoven on both sides based on polypropylene (PP) spun-bond	Polyester	Analysis
Color		Blue	Blue	Same
Dimension (Length)		17.5 cm +/- 1 cm	17.5 cm +/- 0.1 cm	Same
Dimension (Width)		10 cm +/- 1 cm	9.5 cm +/- 0.1 cm	Analysis
OTC use		Yes	Yes	Same
Sterility		Non-Sterile	Non-Sterile	Same
Use		Single Use, Disposable	Single Use, Disposable	Same
ASTM F2100 level		Level 2	Level 2	Same
Biocompatibility		ISO10993	ISO10993	Same

G. Non-Clinical Test Conclusion

Non-clinical tests were conducted to verify that the proposed device met all design specifications as was similar to the predicate device. The test results demonstrated that the proposed device complies with the following standards and the requirements stated in the Guidance for Industry and FDA Staff: Surgical Masks – Premarket Notification [510(k)] Submission issued on March 5, 2004:

- ISO 10993-5: 2009 Biological Evaluation of Medical Devices -- Part 5: Tests For In Vitro Cytotoxicity
- ISO 10993-10: 2010 Biological Evaluation of Medical Devices - Part 10: Tests For Irritation And Skin Sensitization
- ASTM F2100, Standard Specification for Performance of Materials Used In Medical Face Masks
- ASTM F1862, Standard Test Method for Resistance of Medical Face Masks To Penetration by

Synthetic Blood (Horizontal Projection of Fixed Volume At A Known Velocity);

- EN 14683, Medical Face Masks—Requirements and Test Methods;
- ASTM F2101, Standard Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) Of Medical Face Mask Materials, Using A Biological Aerosol of Staphylococcus Aureus;
- ASTM F2299, Stand test method for determining the initial efficiency of materials used in medical face masks to penetration by particulates using latex spheres;
- 16 CFR 1610, Standard for the Flammability of clothing textiles;

Table 2 - Performance Testing

Test Item	Purpose	Acceptance Criteria (level 2)	Result
Fluid Resistance Performance ASTM F1862	The purpose of the performance testing is to demonstrate the functionality of the subject device.	29 out of 32 pass at 120 mmHg	PASS 3 non-consecutive lots tested, using a sample size of 32/lot. 32 out of 32 pass at 120 mmHg
Particulate Filtration Efficiency ASTM F2299		≥ 98%	PASS 3 non-consecutive lots tested, using a sample size of 32/lot. Lot1: 99.82% Lot2: 99.78% Lot3: 99.72%
Bacterial Filtration Efficiency ASTM F2101		≥ 98%	PASS 3 non-consecutive lots tested, using a sample size of 32/lot. Lot1: 99.92% Lot2: 99.84% Lot3: 99.86%
Differential Pressure (Delta P) EN 14683 Annex C		< 6.0mmH ₂ O/cm ²	PASS 3 non-consecutive lots tested, using a sample size of 32/lot. Lot1: 3.79 mmH ₂ O/cm ² Lot2: 4.03 mmH ₂ O/cm ² Lot3: 4.05 mmH ₂ O/cm ²
Flammability 16 CFR 1610		Class 1	PASS 3 non-consecutive lots tested, using a sample size of 32/lot. Class 1

Table 3 Biocompatibility Comparison

Item	Purpose	Acceptance Criteria	Result
Cytotoxicity	The purpose of the performance testing is to demonstrate the functionality	Non-Cytotoxic	PASS Under the conditions of the study, the device is non-cytotoxic.

Irritation	of the subject device.	Non-Irritating	PASS Under the conditions of the study, the device is non-irritating.
Sensitization		Non-Sensitizing	PASS Under the conditions of the study, the device is non-sensitizing

H. Clinical Test Conclusion

No clinical study is included in this submission.

I. Conclusion

Based on the nonclinical tests performed, the subject device (Model: WYZ201) is as safe, as effective, and performs as well as the legally marketed predicate device, K153496, Xiantao Rayxin Medical Products Co., Ltd, Single-use Surgical Masks.