



December 22, 2022

SI-BONE, Inc.
Meirav Harsat
Director of Regulatory Affairs
471 El Camino Real, Suite 101
Santa Clara, California 95050

Re: K222774

Trade/Device Name: iFuse Bedrock Granite® Implant System
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: OUR, NKB, OLO
Dated: December 9, 2022
Received: December 12, 2022

Dear Meirav Harsat:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,


Colin O'neill -S

Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K222774

Device Name

iFuse Bedrock Granite® Implant System

Indications for Use (Describe)

The iFuse Bedrock Granite® Implant System is intended for sacroiliac joint fusion for the following conditions:

- Sacroiliac joint dysfunction that is a direct result of sacroiliac joint disruption and degenerative sacroiliitis. This includes conditions whose symptoms began during pregnancy or in the peripartum period and have persisted postpartum for more than 6 months.
- To augment immobilization and stabilization of the sacroiliac joint in skeletally mature patients undergoing sacropelvic fixation as part of a lumbar or thoracolumbar fusion.
- Acute, non-acute, and non-traumatic fractures involving the sacroiliac joint

When connected to compatible pedicle screw systems with 5.5- or 6.0-mm posterior rods made from either titanium alloy or cobalt chrome alloys the iFuse Bedrock Granite Implant System is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to thoracolumbosacral fusion for the following acute and chronic instabilities or deformities of the thoracic, lumbar, and sacral spine:

- Degenerative disc disease (DDD) as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies
- Spondylolisthesis
- Trauma (i.e., fracture or dislocation)
- Spinal stenosis
- Deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis)
- Spinal tumor
- Pseudarthrosis
- Failed previous fusion

Please refer to the additional information section in the Instructions for Use on compatible pedicle screw system rods.

The iFuse Bedrock Granite Navigation instruments are intended to be used with the iFuse Bedrock Granite Implant System to assist the surgeon in precisely locating anatomical structures in iFuse Bedrock Granite Implant System procedures, in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as the pelvis or vertebra, can be identified relative to the acquired image (CT, MR, 2D fluoroscopic image or 3D fluoroscopic image reconstruction) and/or an image data-based model of the anatomy. iFuse Bedrock Granite Navigation instruments are intended to be used with the Medtronic StealthStation System.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

iFuse Bedrock Granite® Implant System

I. 510(k) SUBMITTER

SI-BONE, Inc.

471 El Camino Real, Suite 101,

Santa Clara, CA 95050

Phone: 408-207-0700

Fax: 408-557-8312

Date Prepared: December 19, 2022

Contact Person: Meirav Harsat, Director Regulatory Affairs

FDA Establishment

Registration No.: 3007700286

II. DEVICE

Trade Name of Device iFuse Bedrock Granite® Implant System
Classification Name Sacroiliac Joint Fixation
Classification II
Regulation Number 21 CFR 888.3040; metallic bone fastener Smooth or threaded
21 CFR 888.3070, Thoracolumbosacral pedicle screw system
21 CFR 882.4560, Stereotaxic instrument
Product Code OUR, NKB, OLO

III. PREDICATE DEVICE

Predicate Device	Manufacturer	510(k)#	Clearance Date
iFuse Bedrock Granite® Implant System	SI-BONE, Inc.	K220195	May 26, 2022

IV. DEVICE DESCRIPTION

The iFuse Bedrock Granite® Implant System consists of implants of various lengths and diameters, and associated instruments sets (for both open and minimally invasive [MIS] approaches). The titanium (Ti-6Al-4V ELI) implant consists of a porous fusion sleeve with threaded length attached to a solid post that has connection and implant placement features of a typical pedicle fixation screw. It is intended to provide sacroiliac joint fusion in the sacral alar iliac (SAI) trajectory (when used with commercially available sacroiliac joint fusion promoting devices), and foundational stabilization when connected to compatible pedicle screw fixation systems in both the SAI and the Iliac trajectories. It is designed for connection to compatible commercially available pedicle screw systems via Ø5.5 mm or Ø6.0 mm diameter circular titanium alloy or cobalt chrome alloy spinal fixation rods. Please refer to the additional information section in the Instructions for Use on compatible pedicle screw system rods.

V. INDICATIONS FOR USE

The iFuse Bedrock Granite® Implant System is intended for sacroiliac joint fusion for the

following conditions:

- Sacroiliac joint dysfunction that is a direct result of sacroiliac joint disruption and degenerative sacroiliitis. This includes conditions whose symptoms began during pregnancy or in the peripartum period and have persisted postpartum for more than 6 months.
- To augment immobilization and stabilization of the sacroiliac joint in skeletally mature patients undergoing sacropelvic fixation as part of a lumbar or thoracolumbar fusion.
- Acute, non-acute, and non-traumatic fractures involving the sacroiliac joint

When connected to compatible pedicle screw systems with 5.5- or 6.0-mm posterior rods made from either titanium alloy or cobalt chrome alloys the iFuse Bedrock Granite Implant System is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to thoracolumbosacral fusion for the following acute and chronic instabilities or deformities of the thoracic, lumbar, and sacral spine:

- Degenerative disc disease (DDD) as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies
- Spondylolisthesis
- Trauma (i.e., fracture or dislocation)
- Spinal stenosis
- Deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis)
- Spinal tumor
- Pseudarthrosis
- Failed previous fusion

Please refer to the additional information section in the Instructions for Use on compatible pedicle screw system rods.

The iFuse Bedrock Granite Navigation instruments are intended to be used with the iFuse Bedrock Granite Implant System to assist the surgeon in precisely locating anatomical structures in iFuse Bedrock Granite Implant System procedures, in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as the pelvis or vertebra, can be identified relative to the acquired image (CT, MR, 2D fluoroscopic image or 3D fluoroscopic image reconstruction) and/or an image data-based model of the anatomy. iFuse Bedrock Granite Navigation instruments are intended to be used with the Medtronic StealthStation System.

VI. SUMMARY OF SUBSTANTIAL EQUIVALENCE

The subject device is substantially equivalent to its predicate in terms of intended use and indications for use, technological characteristics, materials, manufacturing methods, and principles of operation.

VII. PERFORMANCE DATA

SI-BONE performed the following mechanical testing of the connection of the Bedrock Granite tulip to commercially available representative rods.

- Static Axial Grip per ASTM F1798 (axial gripping capacity)

- Static Axial Torsion Grip per ASTM F1798 (torsional gripping capacity)
- Dynamic Axial Compression per ASTM F1717 (dynamic compression bending)

Testing was performed on a range of rod diameters and manufacturing methods to demonstrate compatibility with the iFuse Bedrock Granite[®] Implant System

The test results demonstrate that the device is substantially equivalent to the legally marketed predicate device and does not raise different questions of safety and effectiveness compared to the predicate device.

VIII. CONCLUSION

The iFuse Bedrock Granite[®] Implant System is substantially equivalent to its predicate in terms of intended use and indications for use, technological characteristics, materials, manufacturing methods, and principles of operation. Testing has shown compatibility with standard 5.5 and 6.0 mm diameter circular rods made from titanium alloy or cobalt chrome alloys found in commercially available cleared pedicle screw systems. The test results demonstrate that the device is substantially equivalent to the legally marketed predicate device and does not raise different questions of safety and effectiveness compared to the predicate device.