



December 9, 2022

Instrumentation Laboratory Co.
Nikita Malladi
Regulatory Affairs Manager I
180 Hartwell Road
Bedford, Massachusetts 01730

Re: K223402

Trade/Device Name: HemosIL von Willebrand Factor Antigen
Regulation Number: 21 CFR 864.7290
Regulation Name: Factor Deficiency Test
Regulatory Class: Class II
Product Code: GGP
Dated: November 8, 2022
Received: November 9, 2022

Dear Nikita Malladi:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Takeesha Taylor-bell -S

Deputy Director
Division of Immunology
and Hematology Devices
OHT7: Office of In Vitro Diagnostics
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K223402

Device Name
HemosIL von Willebrand Factor Antigen

Indications for Use (Describe)

Automated latex enhanced immunoassay for the quantitative determination of von Willebrand Factor Antigen (VWF:Ag) in human citrated plasma on IL Coagulation Systems.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

This 510(k) Summary of Safety and Effectiveness is being submitted in accordance with the requirements of the Safe Medical Device Act of 1990 and 21 CFR 807.92.

Submission Type	Special 510(k)	
Submitter's Information	Instrumentation Laboratory Company 180 Hartwell Road Bedford, MA 01730-2443 (USA)	
Contact Person	Nikita Malladi Regulatory Affairs Manager I Phone: 781-674-3245 Fax: 781-861-4207 Email: nmalladi@werfen.com	
Preparation Date	November 08, 2022	
Device Trade Name	HemosIL von Willebrand Factor Antigen	
Predicate Device and 510(k) Number	HemosIL von Willebrand Factor Antigen	K200033
Regulatory Information	Regulation Section No.	21 CFR 864.7290
	Regulation Description	Factor Deficiency Test
	Classification	Class II
	Product Code	GGP
	Classification Panel	Hematology (81)

Device Description

The VWF:Ag kit is a latex particle enhanced immunoturbidimetric assay to quantify VWF:Ag in plasma. When a plasma containing VWF:Ag is mixed with the Latex Reagent and the Reaction Buffer included in the kit, the coated latex particles agglutinate. The degree of agglutination is directly proportional to the concentration of VWF:Ag in the sample and is determined by measuring the decrease of transmitted light caused by the aggregates.

Indications for Use / Intended Use

Automated latex enhanced immunoassay for the quantitative determination of von Willebrand Factor Antigen (VWF:Ag) in human citrated plasma on IL Coagulation Systems.

Reason Submission Qualifies as Special 510(k)

This Special 510(k) is being submitted to modify the open vial stability labeled claim for HemosIL von Willebrand Factor Antigen from 3 months to 14 days based on testing to the current CLSI EP25-A guideline.

The submission meets the criteria for a Special 510(k) based on the following:

- The proposed change is submitted by the manufacturer legally authorized to market the existing device.
- There is a well-established method to evaluate the change: CLSI EP25-A.
- The data can be reviewed in a summary or risk analysis format.

In addition, the change in this submission **does not** introduce:

- Changes to indications for use or intended use
- Changes to operating principle
- Changes to formulation
- Changes to labeled performance claims, *except* to reduce the open vial stability claim to 14 days
- Changes to assay algorithms or data reduction software

Design Control Activities

The open vial stability study was performed in accordance to an established plan and protocol and design control procedures. Testing verified all acceptance criteria were met.

Comparison to Predicate Device		
<i>Similarities</i>		
Item	Predicate (K200033)	Subject Device
Trade Name	HemosIL von Willebrand Factor Antigen	Same
Manufacturer	Instrumentation Laboratory Company	Same
Intended Use/ Indications for Use	Automated latex enhanced immunoassay for the quantitative determination of von Willebrand Factor Antigen (VWF:Ag) in human citrated plasma on IL Coagulation Systems.	Same
Measurand	von Willebrand Factor Antigen	Same
Type of Test	Latex immunoassay	Same
Methodology	The VWF:Ag kit is a latex particle enhanced immunoturbidimetric assay to quantify VWF:Ag in plasma. When a plasma containing VWF:Ag is mixed with the Latex Reagent and the Reaction Buffer included in the kit, the coated latex particles agglutinate. The degree of agglutination is directly proportional to the concentration of VWF:Ag in the sample and is determined by measuring the decrease of transmitted light caused by the aggregates.	Same
Sample Type	Citrated plasma	Same
Kit Composition	Latex Reagent: 2 vials x 3 mL of a suspension of polystyrene latex particles coated with a rabbit polyclonal antibody directed against VWF containing bovine serum albumin, buffer, stabilizer and preservative. Reaction Buffer: 2 vials x 4 mL of HEPES buffer containing bovine serum albumin, stabilizers and preservative.	Same
Instrumentation	ACL Elite / ACL Elite Pro K060162 ACL TOP Family K160276 ACL TOP Family 50 Series K150877	Same

Comparison to Predicate Device (Cont.)		
<i>Differences</i>		
Item	Predicate (K200033)	Subject Device
Reagent Open Vial Stability	Current Insert Claim	Modified Insert Claim
	Opened reagents are stable 3 months at 2-8°C in the original vial.	Opened reagents are stable 14 days at 2-8°C in the original vial.

Conclusion	HemosIL von Willebrand Factor Antigen, with the modified open vial labeled claim, is substantially equivalent to the legally marketed predicate device, HemosIL von Willebrand Factor Antigen, FDA cleared under K200033.
-------------------	---