



May 2, 2023

Nexus Spine, LLC  
% Christine Scifert  
Partner  
MRC Global  
9085 E. Mineral Cir., Suite 110  
Centennial, Colorado 80112

Re: K223529

Trade/Device Name: PressON Spinal Fixation System  
Regulation Number: 21 CFR 888.3070  
Regulation Name: Thoracolumbosacral Pedicle Screw System  
Regulatory Class: Class II  
Product Code: NKB  
Dated: March 30, 2023  
Received: March 31, 2023

Dear Christine Scifert:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin  
O'Neill -S 

Colin O'Neill, M.B.E.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K223529

Device Name

PressON Spinal Fixation System

Indications for Use (Describe)

The PressON Spinal Fixation System is a posterior, non-cervical pedicle screw system intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar, and sacral spine: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, fracture, dislocation, scoliosis, kyphosis, lordosis, spinal tumor, and failed previous fusion (pseudarthrosis).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(k) Summary**

PressON Spinal Fixation System

March 30, 2023

**Company:** Nexus Spine, LLC  
2825 East Cottonwood Parkway Suite 330  
Salt Lake City, UT 84121

**Primary Contact:** Christine Scifert  
Partner, MRC Global  
9085 E. Mineral Cir., Suite 110, Centennial, Colorado 80112  
Phone: (901) 831-8053  
Email: christine.scifert@AskMRCGlobal.com

**Company/Secondary Contact:** Jared Crocker  
Vice President of Quality and Regulatory Affairs  
Nexus Spine, LLC  
Phone: (801) 702-8592  
jared.crocker@nexusortho.com

**Trade Name:** PressON Spinal Fixation System

**Common Name:** Thoracolumbosacral pedicle screw system

**Classification:** Class II

**Regulation:** 21 CFR 888.3070 - Thoracolumbosacral Pedicle Screw System

**Panel:** Orthopedic

**Product Code:** NKB

**Primary Predicate:** K160820 Nexus Spine, LLC, PressON Pro Spinal Fixation System

**Device Description:**

The PressON Spinal Fixation System is composed of pedicle screws and rods. These components can be assembled and implanted using associated instruments via a posterior approach into the pedicles of the noncervical vertebral bodies. The system is composed of pedicle screws, cortical screws, and a variety of couplers. All components of the PressON Spinal Fixation System are made from Ti-6Al-4V ELI (ASTM F-136).

**Indications for Use:**

The PressON Spinal Fixation System is a posterior, non-cervical pedicle screw system intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar, and sacral spine: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, fracture, dislocation, scoliosis, kyphosis, lordosis, spinal tumor, and failed previous fusion (pseudarthrosis).

**Substantial Equivalence:**

The subject components are substantially equivalent to the following predicate device: PressON Pro Spinal Fixation System, Nexus Spine, LLC (K160820)

**Statement of Technological Comparison:**

The PressON Spinal Fixation System is substantially equivalent to the above listed predicate devices in terms of materials, design, indications for use and operational principles.

**Performance Testing:**

The following mechanical performance tests were conducted:

- ASTM F1717 Dynamic Compression Bending
- ASTM F1798 Axial Load Gripping Capacity
- ASTM F1798 Axial Torsion Gripping Capacity
- ASTM F1798 Flexion-Extension Gripping Capacity
- ASTM F1798 Anterior-Posterior Gripping Capacity

Performance testing was performed, and results demonstrated that the subject device is substantially equivalent to the predicate device.

**Conclusion:**

The subject device, PressON Spinal Fixation System, is substantially equivalent to the PressON Pro Spinal Fixation System, Nexus Spine, LLC (K160820).