



April 12, 2023

AMO Manufacturing USA, LLC  
Laarni Ricafort  
Manager, Regulatory Affairs  
510 Cottonwood Drive  
Milpitas, CA 95035

Re: K223838  
Trade/Device Name: Catalys™ Precision Laser System  
Regulation Number: 21 CFR 886.4390  
Regulation Name: Ophthalmic Laser  
Regulatory Class: Class II  
Product Code: OOE  
Dated: March 10, 2023  
Received: March 14, 2023

Dear Laarni Ricafort:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling

(21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

J Angelo Green -S



for Tieuvi Nguyen, Ph.D.

Director

DHT1A: Division of Ophthalmic Devices

OHT1: Office of Ophthalmic, Anesthesia,

Respiratory, ENT and Dental Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K223838

Device Name  
CATALYST™ Precision Laser System

### Indications for Use (Describe)

The CATALYST™ Precision Laser System is indicated for use in patients undergoing cataract surgery for removal of the crystalline lens. Intended uses in cataract surgery include anterior capsulotomy, phacofragmentation, and the creation of single plane and multi-plane arc cuts/ incisions in the cornea, each of which may be performed either individually or consecutively during the same procedure.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) SUMMARY

The following 510(k) Summary is being submitted in accordance with the requirements of 21 CFR 807.92

**510(k) Summary:** K223838

### [807.92(a)(1)] Submitter Information

**Sponsor/Submitter:** AMO Manufacturing USA, LLC  
510 Cottonwood Drive  
Milpitas, CA 95035 USA  
Phone 408-273-5166  
Fax:408 273-5966

**Contact Person:** Laarni Ricafort  
Manager, Regulatory Affairs  
Email: [lricafor@its.jnj.com](mailto:lricafor@its.jnj.com)  
Phone: 408-273-4016  
Fax: :408 273-5966

**Date Summary Prepared:** April 10, 2023

### [807.92(a)(2)] Name of Device

**Device Trade Name:** CATALYS™ Precision Laser System  
**Common Name:** Ophthalmic laser  
**Device Classification:** Class II  
**Regulation Number:** 21 CFR 886.4390  
**Classification Name:** Ophthalmic Femtosecond Laser  
**Product Code:** OOE

### [807.92(a)(3)] Legally Marketed Devices

**Predicate Device:** CATALYS™ Precision Laser System  
K200056; cleared May 18, 2020

### [807.92(a)(4)] Device Description

#### Device Description:

Catalys™ Precision Laser System ophthalmic surgical laser system used in healthcare facilities such as hospitals, Ambulatory Surgery Centers (ASCs) and surgeon office settings. The System is an electromedical device which contains software. System components include a single-use Liquid Optics™ Interface and optional Mobile Patient Bed.

The Catalys™ Precision Laser System (also referred to as the Catalys™ System or System) is an ophthalmic surgical laser system indicated for use in cataract surgery to create a precise anterior capsulotomy and/or to effect lens fragmentation, thus facilitating efficient lens removal. The System also creates single plane and multi-plane arc cuts/incisions in the cornea, each of which may be performed either individually or consecutively during the same procedure. The System employs femtosecond (“FS”) laser technology with integrated Optical Coherence Tomography (“OCT”), all of which are controlled and monitored by dedicated electronics. The System utilizes a common optical path for the OCT and femtosecond treatment laser (including the three-dimensional scanner and Liquid Optics™ [patient] Interface). As such, the beams are intrinsically co-registered and provide for precise overlap between imaging and treatment beams. Ocular surfaces recognized by the system software include anatomy within the anterior chamber, such as the anterior and posterior corneal surfaces and the anterior and posterior surfaces of the crystalline lens. Detailed axial or sagittal cross-sectional views are available via OCT, to demarcate proposed incisions versus adjacent ocular structures (for example, iris, pupil and limbus).

The Catalys™ Precision Laser System laser classification per 21 CFR 1040.10 and 1040.11 is Class 4.

### [807.92(a)(5)] Intended Use

#### Indications for Use:

The Catalys™ Precision Laser System is indicated for use in patients undergoing cataract surgery for removal of the crystalline lens. Intended uses in cataract surgery include anterior capsulotomy, phacofragmentation, and the creation of single plane and multi-plane arc cuts/incisions in the cornea, each of which may be performed either individually or consecutively during the same procedure.

**Difference in Indications from Predicate Device**

The Catalys™ Precision Laser System is indicated for use in patients undergoing cataract surgery for removal of the crystalline lens. Intended uses in cataract surgery include anterior capsulotomy, phacofragmentation, and the creation of single plane and multi-plane arc cuts/incisions in the cornea, each of which may be performed either individually or consecutively during the same procedure. The subject device and the predicate device have the same indications for use.

**[807.92(a)(6)] Technical Characteristics**

**Technological Characteristics:**

The subject Catalys Precision Laser System is unchanged in regard to its technological characteristics, indications for use, and intended uses. The software revisions for the subject device include updates to the graphical user interface and host to improve the efficiency of the workflow as well as the addition of software modules to support the import of patient exams and iris registration. The revisions in the modified device are constrained to only software changes, of which most are related to advanced astigmatism management and additional updates to currently implemented features.

Software verification and validation testing in addition to bench testing was performed to verify the ability of the software to meet its intended use and to ensure that no adverse effects have been introduced due to the software change. This testing included subsystem level verification and regression testing as well as system validation using the latest software, Mobile Patient Bed, and Liquid Optics Interface.

The following table provides a comparison of the predicate device and subject device for the purpose of demonstrating substantial equivalence to the predicate device for its indication for use, intended use, technological characteristics and added new features.

**Similarities and Differences Between Cleared Predicate Device and Subject Device**

Attribute	Predicate Device Catalys Precision Laser System	Subject Device Catalys Precision Laser System
<b>510(k) Number</b>	K200056	TBD
<b>Regulation Number</b>	886.4390	Same

Attribute	Predicate Device Catalys Precision Laser System	Subject Device Catalys Precision Laser System
<b>Regulation Name</b>	Ophthalmic Laser	Same
<b>Regulatory Class</b>	Class II	Same
<b>Product Code</b>	OOE	Same
<b>Indications for Use</b>	The Catalys™ Precision Laser System is indicated for use in patients undergoing cataract surgery for removal of the crystalline lens. Intended uses in cataract surgery include anterior capsulotomy, phacofragmentation, and the creation of single plane and multi-plane arc cuts/incisions in the cornea, each of which may be performed either individually or consecutively during the same procedure.	Same
<b>System Type</b>	Ophthalmic Femtosecond Laser with Spectral Domain OCT	Same
<b>Laser Mechanism of Action</b>	Plasma, Cavitation	Same
<b>Treatment Laser Wavelength (nm)</b>	1030 ±5	Same
<b>Output Power, Max</b>	Per ISO 15004-2:2007 limits	Same
<b>Maximum Pulse Energy (µJ)</b>	10	Same
<b>Repetition Rate (kHz)</b>	9-120	Same
<b>Pulse Duration (fs)</b>	< 600	Same
<b>Spot Size; diameter (µm)</b>	5	Same
<b>System controls</b>	Microprocessor with Graphical User Interface	Same
<b>Patient Contact Interface</b>	Suction-ring type interface devices (marketed as Liquid Optics™ Interface) Sterile and Single-use	Same

Attribute	Predicate Device Catalys Precision Laser System	Subject Device Catalys Precision Laser System
<b>LOI Suction Ring Seal Diameters (mm)</b>	LOI External (mm): 21.6 Internal (mm): 14.1  LOI-12 External (mm):19 Internal (mm):12	Same
	0180-1401 External (mm): 21.6 Internal (mm): 14.1  0180-1201 External (mm): 19 Internal (mm): 12	Same
<b>OCT Axial Resolution (µm)</b>	30	Same
<b>OCT transverse Resolution (µm)</b>	15	Same
<b>Scan speed (A-scans/sec)</b>	1000	Same
<b>A-scan depth (nm)</b>	2	Same
<b>Optical Source</b>	820-930	Same
<b>Optical Power</b>	ANSI Class 1 < 3.48mW at cornea	Same
<b>Iris Imaging</b>	Live iris view	Same
<b>Trajectory Timing Synchronization</b>	FPGA coordinates from one non-reentrant VI	Same
<b>Communication Method for Watchdog for Host PC with the Mobile Patient Bed Pairing</b>	Direct FPGA Interface	Same



Attribute	Predicate Device Catalys Precision Laser System	Subject Device Catalys Precision Laser System
<b>Software Features</b>	Built-in Nomogram Toric Alignment Marks Lens Fragmentation Alignment Pre-op Manual Entry Pre-op Import (from Cassini pre-op device) Iris Registration (built by Cassini Technologies)	Addition of Stevens nomogram Toric Mark Axis Calculator Additional Nomogram customizations Pre-op Import (from IOLMaster 700 pre-op device) Iris Registration (built by Johnson and Johnson)

**[807.92(b)(1)] Determination of Substantial Equivalence**

The cOS 7.0 software within the subject device was subjected to hardware and software bench tests, in conjunction with simulated use testing.

Software-specific bench testing of the Catalys™ System was conducted to demonstrate the System’s ability to meet all intended design specifications related to the software design changes.

**Non-Clinical Performance Data:**

Bench testing of the predicate device is directly applicable to the subject device as there are no significant changes to the subject device other than the design changes resident in the software.

Bench testing, when coupled with software regression testing, verification and validation testing presented for the subject device, including regression testing provides reasonable assurance that the System remains safe and effective for its intended use and furthermore, that it is substantially equivalent to the identified predicate device

Clinical Data was deemed not necessary for the Catalys™ Precision Laser System. The performance data demonstrated that the device performs as intended.

**Clinical Performance Data:**

The subject device does not include any changes to the indications for use or intended use of the predicate device. It does not introduce any new harms or unacceptable risks, and therefore does not require clinical testing to assess safety and performance or to demonstrate equivalence.

**[807.92(b)(3)] Conclusion**

**Conclusions from Non-Clinical and Clinical Tests:**

The subject Catalys Precision Laser System is substantially equivalent to the currently cleared Catalys Precision Laser System. The subject Catalys Precision Laser System is substantially equivalent to the predicate in terms of indications for use, technological characteristics and fundamental scientific technology. The mechanism of laser cutting is the same for both systems, in that the ultra-short laser pulses create a highly localized plasma and subsequent cavitation event that, when controlled by a computerized scanning system, direct the laser beam through a three-dimensional pattern to produce a precise capsulotomy, fragment the crystalline lens and create arc cuts/incisions in the cornea.