



February 23, 2023

Stryker
Michelle Stephens
Senior Staff Regulatory Affairs Specialist
5900 Optical Ct
San Jose, California 95138

Re: K230216

Trade/Device Name: 1688 4K Camera Control Unit with Advanced Imaging Modality
Regulation Number: 21 CFR 876.1500
Regulation Name: Endoscope And Accessories
Regulatory Class: Class II
Product Code: GCJ, GWG
Dated: January 25, 2023
Received: January 26, 2023

Dear Michelle Stephens:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,


Jianting Wang -S

Jianting Wang
Acting Assistant Director
DHT4A: Division of General Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K230216

Device Name

1688 4K Camera System with Advanced Imaging Modality (AIM)

Indications for Use (Describe)

The 1688 Video Camera is indicated for use in general laparoscopy, nasopharyngoscopy, ear endoscopy, sinuscopy, neurosurgery and plastic surgery whenever a laparoscope/ endoscope/ arthroscope/ sinuscope is indicated for use. The 1688 Video Camera is indicated for adults and pediatric patients.

A few examples of the more common endoscope surgeries are Laparoscopic cholecystectomy, Laparoscopic hernia repair, Laparoscopic appendectomy, Laparoscopic pelvic lymph node detection, Laparoscopically assisted hysterectomy, Laparoscopic and thorascopic anterior spinal fusion, Anterior cruciate ligament reconstruction, Knee arthroscopy, Small joint arthroscopy, Decompression fixation, Wedge resection, Lung biopsy, Pleural biopsy, Dorsal sympathectomy, Pleurodesis, Internal mammary artery dissection for coronary artery bypass, Coronary artery bypass grafting where endoscopic visualization is indicated and Examination of the evacuated cardiac chamber during performance of valve replacement.

The users of the 1688 Video Camera are general and pediatric surgeons, gynecologists, cardiac surgeons, thoracic surgeons, plastic surgeons, orthopedic surgeons, ENT surgeons, neurosurgeons and urologists.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(k) Summary

Submitter:

Applicant:	Stryker Endoscopy 5900 Optical Court San Jose, CA 95138
Contact Person:	Michelle Stephens Senior Staff Regulatory Affairs Specialist Email: michelle.stephens@stryker.com Phone: (408) 754-2473
Date Prepared:	February 16, 2023

Subject Device:

Name of Device:	1688 4K Camera System with Advanced Imaging Modality
Common or Usual Name	3-chip Video Camera
Classification Name:	Laparoscope, General and Plastic Surgery (21 C.F.R. §876.1500) Endoscope, Neurological (21 C.F.R. §882.1480)
Regulatory Class:	II
Product Code:	G CJ G WG
510(k) Review Panel:	General & Plastic Surgery Neurology

Predicate Device(s):

780 nm 1688 4K Camera System with Advanced Imaging Modality (primary predicate)	K214046
1688 4K Camera System with Advanced Imaging Modality (reference device)	K222130

Device Description:

The 1688 4K Camera System with Advanced Imaging Modality is an endoscopic camera system that produces live video in the surgical field during surgical endoscopic procedures. The system is sensitive in the visible and infrared spectrum. The optical image is transferred from the surgical site to the camera head by a variety of rigid and flexible endoscopes, which are attached to the camera head. The 1688 4K Camera System consists of three main components: (1) a camera control unit (CCU); (2) a camera head with an integral cable that connects to the CCU; and (3) a coupler for attaching an endoscope to the camera head.



Indications for Use:

Subject Device	Predicate Devices	
<p align="center">1688 4K Camera System <i>This Submission</i></p>	<p align="center">780 nm 1688 4K Camera System (K214046 – primary predicate)</p>	<p align="center">1688 4K Camera System with Advanced Imaging Modality (K222130 – reference device)</p>
<p>Intended Use: Endoscopic white light and near-infrared illumination and imaging during endoscopic procedures.</p>	<p>Intended Use: Same as subject device</p>	<p>Intended Use: Same as subject device</p>
<p>Indications for Use: The 1688 Video Camera is indicated for use in general laparoscopy, nasopharyngoscopy, ear endoscopy, sinuscopy, neurosurgery and plastic surgery whenever a laparoscope/ endoscope/ arthroscope/ sinuscope is indicated for use. The 1688 Video Camera is indicated for adults and pediatric patients.</p> <p>A few examples of the more common endoscope surgeries are Laparoscopic cholecystectomy, Laparoscopic hernia repair, Laparoscopic appendectomy, Laparoscopic pelvic lymph node detection, Laparoscopically assisted hysterectomy, Laparoscopic and thorascopic anterior spinal fusion, Anterior cruciate ligament reconstruction, Knee arthroscopy, Small joint arthroscopy, Decompression fixation, Wedge resection, Lung biopsy, Pleural biopsy, Dorsal sympathectomy, Pleurodesis, Internal mammary artery dissection for coronary artery bypass, Coronary artery bypass grafting where endoscopic visualization is indicated and Examination of the evacuated cardiac chamber during performance of valve replacement.</p> <p>The users of the 1688 Video Camera are general and pediatric surgeons, gynecologists, cardiac surgeons, thoracic surgeons, plastic surgeons, orthopedic surgeons, ENT surgeons, neurosurgeons and urologists.</p>	<p>Indications for Use: Same as subject device</p>	<p>Indications for Use: Same as subject device</p>



Comparison of Technological Characteristics with the Predicate Device:

Item	Subject Device	Predicate Devices	
	1688 4K Camera System (This Submission)	780 nm 1688 4K Camera System (K214046-primary predicate)	1688 4K Camera System with Advanced Imaging Modality (K222130 – reference device)
Manufacturer	Stryker	Same as subject device	Same as subject device
Imaging Modes	White Light Near-infrared – fluorescence Near-infrared – transillumination	Same as subject device	Same as subject device
Camera System Components	Camera Control Unit Camera Head(s) – Standard, Integrated, Inline, Pendulum, Autoclavable Coupler(s) – AIM 4K, AIM 4K Autoclavable	Camera Control Unit Camera Head(s) – Standard, Integrated, Inline, Pendulum Coupler(s) – AIM 4K	Camera Control Unit Camera Head(s) – Standard, Integrated, Inline, Pendulum, Autoclave Coupler(s) – AIM 4K, AIM 4K Autoclave
Principles of Operations	Via an optical endoscope and coupler, light is projected from a light source onto one or more complementary metal oxide semiconductor image sensors which acquire a continuous stream of image data. The image data is processed to provide a video stream that is then sent to a display for viewing.	Same as subject device	Same as subject device
Safety Standards	IEC 60601-1 IEC 60601-1-6 IEC 60601-2-18 IEC 60601-1-2	Same as subject device	Same as subject device
Modes of Operation	Alternate Frame processing Simultaneous Frame processing	Same as subject device	Same as subject device
Image Sensor	CMOS image sensor	Same as subject device	Same as subject device
Image Processing/ Video Output	Digital	Same as subject device	Same as subject device
Resolution	4K (up to 3840 x 2160)	Same as subject device	Same as subject device
Frame Rate	60 frames per second	Same as subject device	Same as subject device
Camera Control Unit PCB Digital Board	Revision F NOTE: Modified to account for End of Life (EOL) and supply chain shortages.	Revision E	Revision E



Performance Data:

The following performance data were provided in support of the substantial equivalence determination:

Test	Method	Result
Electrical Safety	In accordance with: ANSI/AAMI ES60601-1:2005 + A1:2012 IEC 60601-2-18:2009 IEC 60601-1-6:2010 + A1:2013	Pass
Electromagnetic Compatibility (EMC)	In accordance with IEC 60601-1-2:2014	Pass
Performance - Bench	CCU Timing	Pass
	Video Compatibility	Pass
Software Verification	In accordance with IEC 62304:2015	Pass

NOTE: The 1688 4K Camera System is not patient contacting; therefore, biocompatibility testing was not required to support the determination of substantial equivalence. Additionally, the device modifications to the 1688 4K Camera System do not require clinical studies to support the determination of substantial equivalence.

Conclusions:

The 1688 4K Camera System is substantially equivalent in design, intended use, principles of operation, technological characteristics and safety features to the predicate devices. There are no different issues of safety and/or effectiveness introduced by the 1688 4K Camera System.