



July 11, 2023

iLine Microsystems S.L.
Miren Hormaeche
Regulatory Affairs and Quality Director
Paseo Mikeletegi, 69
San Sebastian-Donostia, Gipuzkoa 20009
Spain

Re: K231711

Trade/Device Name: microINR System
Regulation Number: 21 CFR 864.7750
Regulation Name: Prothrombin Time Test
Regulatory Class: Class II
Product Code: GJS
Dated: June 12, 2023
Received: June 12, 2023

Dear Miren Hormaeche:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Takeesha Taylor-bell -S

Deputy Director
Division of Immunology and Hematology Devices
OHT7: Office of In Vitro Diagnostics
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K231711

Device Name

microINR System

Indications for Use (Describe)

The microINR System measures prothrombin time (PT) expressed in International Normalized Ratio (INR), for monitoring oral anticoagulant therapy with warfarin.

The microINR System consists of a meter and chips (test strips) and uses fresh capillary whole blood from a fingerstick. The microINR System is intended for patient self-testing use as well as for healthcare professionals at Point of Care settings.

The microINR System is intended for use in patients 18 years old or older. Patients must be stable on warfarin medication for at least 6 weeks before starting to use the microINR System.

For Self-testing use: The System is intended for properly trained users under specific prescription of a physician.

Caution: The microINR System is not intended for use in patients who are transitioning from heparin treatment to VKA therapy. The microINR System is not intended to be used for screening purposes.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

This Special 510(k) summary is being submitted in accordance with the requirements of 21 CFR 807.92.

1. SUBMITTER INFORMATION

Owner iLine Microsystems, S.L.
Paseo Mikeletegi, 69
20009 Donostia, Guipúzcoa
Spain

Contact Miren Itsaso Hormaeche
ihormaeche@ilinemicrosystems.com
Tel. +34 943 005 651
Fax: +34 943 008 737

Date Summary Prepared July 10, 2023

2. DEVICE INFORMATION

Proprietary Name microINR System (microINR Chips,
microINR Kit, microINR Link Kit)

Common Name Prothrombin time test

Panel Hematology

Regulatory Information:

Classification				
Device	Regulation Section	Device Class	Product Code	Test
microINR System	21 CFR 864.7750	II	GJS	Prothrombin time test

3. SUBSTANTIAL EQUIVALENCE INFORMATION:

Element	Predicate Device
Predicate Device Name	microINR System
Common Name	Prothrombin time test
510 (k) Number	K201185
Manufacturer	iLine Microsystems S.L.

4. DEVICE DESCRIPTION:

The modified device microINR System containing the microINR Link Meter is derived from the existing device microINR Meter. The change is for the addition of Bluetooth connectivity to the meter. Bluetooth functionality stays inactive during the testing/measuring process. All the actions related to Bluetooth functionality are set before or after performing the analytical test. The fundamental scientific technology of the modified device and its performance has not changed. No performance changes to the microINR System have been implemented since its clearance on K201185.

5. INDICATIONS FOR USE/INTENDED USE:

The microINR System measures prothrombin time (PT) expressed in International Normalized Ratio (INR), for monitoring oral anticoagulant therapy with warfarin.

The microINR System consists of a meter and chips (test strips) and uses fresh capillary whole blood from a fingerstick.

The microINR System is intended for patient self-testing use as well as for healthcare professionals at Point of Care settings.

The microINR System is intended for use in patients 18 years old or older. Patients must be stable on warfarin medication for at least 6 weeks before starting to use the microINR System.

For Self-testing use: The System is intended for properly trained users under specific prescription of a physician.

Caution: The microINR System is not intended for use in patients who are transitioning from heparin treatment to VKA therapy. The microINR[®] System is not intended to be used for screening purposes.

6. SUMMARY COMPARISON OF TECHNOLOGICAL CHARACTERISTICS (PREDICATE):

The microINR System containing the microINR Link Meter uses the same test strips (the microINR Chips) and measuring algorithm as the predicate, the microINR System for self-testing use and professional use in CLIA waived settings (K201185). The only change is the addition of Bluetooth connectivity to the Meter. The performance characteristics of the microINR System remain the same.

7. TESTING IN SUPPORT OF SUBSTANTIAL EQUIVALENCE DETERMINATION

Testing on the modified device was conducted as follows: testing for Radiofrequency, Electromagnetic Compatibility, and Electrical Safety; Cybersecurity analysis; Usability testing; and Software verification and validation.

All testing results met the pre-determined acceptance criteria. Based on the testing, the microINR System including the microINR Link Meter performs as intended, with no new questions of safety or effectiveness identified during testing.

The risk analysis of the implemented changes concluded that the microINR System including the microINR Link Meter has no significant detectable risks compared to the predicate device and all residual risks have been deemed acceptable.

8. CONCLUSION

The microINR System with the microINR Meter was previously cleared for self-testing use and professional use in CLIA waived settings under premarket notification K201185.

This Special 510(k) is being submitted to obtain clearance for the microINR System including the microINR Link Meter for patient self-testing, and CLIA waiver for professional use in Point of Care settings. The verification and/or validation activities required based on the Risk Analysis were evaluated and revealed that the Bluetooth connectivity features were correctly implemented.

Based on the comparative analysis, the intended use, principles of operation, performance characteristics and technological characteristics, the microINR System including the microINR Link Meter does not introduce any new significant risks or any new questions of safety and effectiveness. The difference in terms of Bluetooth functionality between the subject and the predicate microINR System device raises no new issues of safety or effectiveness. Therefore, the microINR System containing the microINR Link Meter is substantially equivalent to the predicate devices.