

September 1, 2023

Fujifilm Healthcare Americas Corporation Chaitrali Kulkarni Sr. Regulatory Affairs Specialist 81 Hartwell Avenue Suite 300 Lexington, Massachusetts 02421

Re: K232314

Trade/Device Name: Hood (DH-106STL, DH-116STL, DH-126STL, DH-096ST)

Regulation Number: 21 CFR 876.1500

Regulation Name: Endoscope and accessories

Regulatory Class: Class II Product Code: FDS, FDF Dated: August 2, 2023 Received: August 2, 2023

Dear Chaitrali Kulkarni:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to https://www.fda.gov/medical-device-problems.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance) and CDRH Learn (https://www.fda.gov/training-and-continuing-education/cdrh-learn). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice">https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Sivakami Venkatachalam -S

for
Shanil P. Haugen, Ph.D.
Assistant Director
DHT3A: Division of Renal, Gastrointestinal,
Obesity and Transplant Devices
OHT3: Office of GastroRenal, ObGyn,
General Hospital and Urology Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration

Indications for Use

Form Approved: OMB No. 0910-0120

Expiration Date: 06/30/2023
See PRA Statement below.

Submission Number (if known)						
K232314						
Device Name						
Hood (DH-106STL, DH-116STL, DH-126STL, DH-096ST)						
Indications for Use (Describe)						
This hood is intended to be used in combination with compatible endoscopes to maintain the field of view during endoscopic procedures such as mucosal resection						
Type of Use (Select one or both, as applicable)						
Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)						
CONTINUE ON A SEPARATE PAGE IF NEEDED.						

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer Paperwork Reduction Act (PRA) Staff PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

FUJIFILM Medical Systems U.S.A., Inc.'s FUJIFILM Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST

Date: August 31, 2023

Submitter's Information

FUJIFILM Healthcare Americas Corporation 81 Hartwell ave Lexington, MA 02421

Contact Person:

Chaitrali Kulkarni Sr. Regulatory Affairs Specialist Telephone: (704) 517 4886 E-Mail: chaitrali.kulkarni@fujifilm.c om

Identification of the Proposed Device:

Proprietary/Trade Name: Fujifilm Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST

Common Name: Hood Device Class: Class II

Review Panel: Gastroenterology/Urology

Classification Information:

Classification Name	CFR Section	Product Codes
Gastroscope and Accessories (Flexible/Rigid)	21 CFR 876.1500	FDS
Colonoscope and Accessories (Flexible/Rigid)	21 CFR 876.1500	FDF

Predicate Devices

• Fujifilm Hood Model DH-28GR, DH-29CR, DH-30CR (K162749)

Intended Use / Indications for Use

The FUJIFILM Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST are intended to be used in combination with compatible endoscopes to maintain the field of view during endoscopic procedures such as mucosal resection.

Device Description

FUJIFILM Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST are intended to be used in combination with compatible endoscopes to maintain the field of view during endoscopic procedures such as mucosal resection.

The FUJIFILM Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST are comprised of three main sections: an attaching portion, a distal portion, and a drain portion. The attaching portion is a wider diameter opening which is used to connect the hood to an applicable endoscope; a distal portion is the ending portion of the hood which tapers into narrower diameter opening, the drain slits on the distal portion form drain portion which prevent the fluids lodging on the surface of the endoscope.

The subject devices are used in combination with their respective applicable Fujifilm's endoscopes as shown in the table 7.1. All the applicable endoscopes marketed in USA are cleared under respective 510(k) notices.

Technological Characteristics

A comparison of the technological characteristics between the subject and predicate devices is provided in the table below.

Table 7.1: Comparison of technological characteristics of FUJIFILM Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST with their predicate device DH-28GR, DH-29CR, DH-30CR (K162749)

		edicate Device model 3GR, DH-29CR, DH-30CR		osed Device model _, DH-116STL, DH-126STL	Proposed Device model DH-096ST	FUJIFILM Comments on Predicate Device model
	DH-28GR	DH-29CR DH-30CR	DH-106STL	DH-116STL DH-126STL	DH-096ST	
Manufacturer	DH-28GR DH-29CR DH-30CR	FUJIFILM corporation	DH-106STL DH-116STL DH-126STL	FUJIFILM corporation	FUJIFILM corporation	Same as Predicate Device model
Outer diameter	DH-28GR DH-29CR DH-30CR	11.8mm 13.0mm 14.8mm	DH-106STL DH-116STL DH-126STL	11.6mm 11.6mm 13.7mm	11.0mm	The outer diameter changes slightly due to the shape change of the resin
Maximum diameter of attaching	DH-28GR	15.5mm	DH-106STL	15.3mm	14.2mm	DH-106STL, DH-116STL, DH-126STL Since the dimensions of the rubber have been changed in line with the change in the shape of the resin, the
endoscope	DH-29CR	16.5mm	DH-116STL	15.8mm		maximum diameter will change when installed. DH-096ST
	DH-30CR	18.4mm	DH-126STL	16.8mm		Since the size of the rubber is reduced to combine with an endoscope that is thinner than the existing hood, the maximum diameter changes when attached.
Distance from the tip	DH-28GR DH-29CR DH-30CR	7.0mm	DH-106STL DH-116STL DH-126STL	6.0mm	6.0mm	In order to increase versatility, the protruding length was shortened to secure the field of view when the endoscope is attached.
Total length	DH-28GR DH-29CR DH-30CR	17.0mm	DH-106STL DH-116STL DH-126STL	16.0mm	16.0mm	The total length is shortened by changing the Distance from the tip.
Combination endoscope	DH-28GR	EG-590WR, EG-580RD, EG-600ZW, EG-760Z, EC-580RD/M, EC-580RD/L, EC-740T/M, EC-740T/L	DH-106STL	EG-760Z, EC-740T/L	EG-760R	Appropriate endoscopes are set in combination according to the shape of the hood.
	DH-29CR	EG-760CT, EG-590ZW, EG-530CT, EC-760P-V/M, EC-760P-V/L	DH-116STL	EG-760CT, EG-530CT, EC-760P-V/L		
	DH-30CR	EC-760ZP-V/L, EC-600WL v2, EC-760R-V/L, EC-530DL, ES-530WE, EC-600HL, EC-760Z-V/L, EC-760S-V/L	DH-126STL	EC-600LS, EC-600WL, EC-760R-V/L, EC-760ZP-V/L		
How to attach to the	DH-28GR	Align the objective lens of endoscope with the drain of	DH-106STL	Attach the hood to the distal end of the endoscope.	Align the objective lens of endoscope with the drain	Same as Predicate Device model
endoscope	DH-29CR DH-30CR	the hood and attach the hood to the distal end of endoscope by pressing the hood until it stops.	DH-116STL DH-126STL	Align the opening of the distal portion of the hood with the instrument channel outlet of the endoscope and press the hood until it stops.	of the hood and attach the hood to the distal end of endoscope by pressing the hood until it stops.	
Existence of Sterile	DH-28GR DH-29CR DH-30CR	EOG Sterilized	DH-106STL DH-116STL DH-126STL	EOG Sterilized	EOG Sterilized	Same as Predicate Device model
Reuse or not re-use	DH-28GR DH-29CR DH-30CR	Single Use	DH-106STL DH-116STL DH-126STL	Single Use	Single Use	Same as Predicate Device model

Intended Use	DH-28GR DH-29CR DH-30CR	These hoods are intended to be used in combination with compatible endoscopes to maintain the field of view during observation of the digestive tract.	DH-106STL DH-116STL DH-126STL	These hoods are intended to be used in combination with compatible endoscopes to maintain the field of view during observation of the digestive tract.	These hoods are intended to be used in combination with compatible endoscopes to maintain the field of view during observation of the digestive tract	Same as Predicate Device model
Material	DH-28GR DH-29CR DH-30CR	Silicone rubber Manufacturer: Shin-Etsu Chemical Co., Ltd. Model:KE- 2090-60-A/B Polycarbonate resin Manufacture: Mitsubishi Engineering-Plastics Corporation Model: IUPILON S-2001R 5355	DH-106STL DH-116STL DH-126STL	Silicone rubber Manufacturer: Shin-Etsu Chemical Co., Ltd. Model:KE-2090-60-A/B Polycarbonate resin Manufacture: Mitsubishi Engineering-Plastics Corporation Model: IUPILON S-2001R 5355	Silicone rubber Manufacturer: Shin-Etsu Chemical Co., Ltd. Model:KE-2090-60-A/B Polycarbonate resin Manufacture: Mitsubishi Engineering-Plastics Corporation Model: IUPILON S-2001R 5355	Same as Predicate Device model
Transport and Storage Environment	DH-28GR DH-29CR DH-30CR	Temperature: -20 to +60°C Humidity: 10 to 95% (No condensation) Atmosphere: 70 to 106 kPa(Within range of atmospheric pressure)	DH-106STL DH-116STL DH-126STL	Temperature: -20 to +60°C Humidity: 10 to 95% (No condensation) Atmosphere: 70 to 106 kPa(Within range of atmospheric pressure)	Temperature: -20 to +60°C Humidity: 10 to 95% (No condensation) Atmosphere: 70 to 106 kPa(Within range of atmospheric	Same as Predicate Device model

The principle of operation and intended use of the FUJIFILM Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST are identical to that of the predicate device FUJIFILM Hood Model DH-28GR, DH-29CR, DH-30CR (K162749). The modifications done to the subject device include changes in dimensions and applicable endoscopes. The distance from the tip and total length is slightly smaller in the subject device compared to the predicate device. The maximum diameter of attaching endoscope of DH-096ST and DH-126STL is slightly smaller the predicate device.

As detailed in the following sections of the 510(k) notice, these changes do not alter the intended use or fundamental technology of the subject devices neither affects their safety and effectiveness.

Performance Data

Endoscope specific testing was conducted using the following consensus standards: ISO 8600-1:2015; and ISO 8600-4:2014.

Subject devices met performance specifications in the following additional testing:

- Outer diameter
- Maximum diameter of attaching endoscope
- Distance from the tip

Substantial Equivalence

FUJIFILM Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST have substantially the same intended use and similar indications, technological characteristics, and principles of operation as their predicate device Fujifilm Hood Model DH-28GR, DH-29CR and DH-30CR (K162749) . The minor dimensional differences between the FUJIFILM Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST and their predicate device Fujifilm Hood Model DH-28GR, DH-29CR and DH-30CR (K162749) were made for the purpose of overall product enhancement and general technological advancement, and raise no new issues of safety or effectiveness. Performance data demonstrated that the FUJIFILM Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST have substantial equivalent performance to the predicate device Fujifilm Hood Model DH-28GR, DH-29CR and DH-30CR (K162749) .

Conclusions

The subject devices, Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST are substantially equivalent to their predicate device Hood Model DH-28GR, DH-29CR and DH-30CR (K162749), based on intended use/indications for use and technological characteristics. The differences in the dimensions between the subject devices and its predicate device raise no new issues of safety or effectiveness. Bench testing data demonstrated that the subject devices have substantially equivalent performance to the predicate. Thus, the subject devices FUJIFILM Hood Models DH-106STL, DH-116STL, DH-126STL, and DH-096ST are as substantially equivalent as their predicate Fujifilm Hood Model DH-28GR, DH-29CR and DH-30CR (K162749).