



December 5, 2020

Daesung Maref Co., LTD
Su Hyeon So
Assistant Researcher
298-24, Gongdan-Ro
Gunpo-si, Gyeonggi-do 15809
Korea, South

Re: K203016
Trade/Device Name: DVT-2600
Regulation Number: 21 CFR 870.5800
Regulation Name: Compressible Limb Sleeve
Regulatory Class: Class II
Product Code: JOW
Dated: September 21, 2020
Received: October 1, 2020

Dear Su Hyeon So:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for Fernando Aguel
Assistant Director
DHT2B: Division of Circulatory Support,
Structural and Vascular Devices
OHT2: Office of Cardiovascular Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K203016

Device Name
DVT-2600

Indications for Use (Describe)

A device intended to prevent Deep Vein Thrombosis/Pulmonary Embolism (DVT/PE) by increasing venous blood flow to a patient who has a risk of DVT/PE.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) Summary

1. Data Prepared [21 CFR 807.92(a)(a)]

November 02, 2020

2. Submitter's Information [21 CFR 807.92(a)(1)]

- Name of Manufacturer :
DAESUNG MAREF CO., LTD.
- Address :
298-24, Gongdan-ro Gunpo-si, Gyeonggido Republic of Korea
- Contact Name :
Su Hyeon, So
- Telephone No. :
82-31-459-7211
- Fax No. :
82-31-459-7215
- Email Address :
mdra@dsmaref.com
- Registration No. :
3004116008

3. Trade Name, Regulation Name, Classification [21 CFR 807.92(a)(2)]

Trade / Device Name	DVT-2600
Classification Name	Compressible Limb Sleeve
Regulation Number	21 CFR 870.5800
Regulation Class	II
Product Code	JOW

4. Identification of Predicate Device(s) [21 CFR 807.92(a)(3)]

Predicate Device

- 510(k) Number :
K112677
- Applicant :
DAESUNG MAREF CO., LTD.
- Trade / Device Name :
THE VENOUS ASSIST SYSTEM
- Regulation Number :
21 CFR 870.5800
- Regulation Name :
Compressible Limb Sleeve
- Regulation Class:
II
- Product Code:
JOW

Predicate device has not been subject to a design-related recall.

5. Description of the Device [21 CFR 807.92(a)(4)]

A device intended to prevent DVT/PE by increasing venous blood flow to a patient who has a risk of DVT/PE.

Operating principal of this device is that the air output from the air motor is delivered to a sleeve composed of four air chambers are sequentially inflated from the first air chamber to the third chamber through solenoid valve.

6. Indications For Use [21 CFR 807(a)(5)]

A device intended to prevent Deep Vein Thrombosis/Pulmonary Embolism (DVT/PE) by increasing venous blood flow to a patient who has a risk of DVT/PE.

7. Determination of Substantial Equivalence

Summary of technological characteristics of the device compared to the predicate device. [21CFR 807.92(a)(6)]

The DVT-2600 is substantially equivalent to legally marketed predicate device (The Venous Assist System DVT-2600) with respect to indications for use and technology characteristics.

The table below presents comparisons for device :

[Table 1. Comparison of Proposed Device to Predicate Device]

	Proposed Device	Predicate Device
Model Name	DVT-2600	DVT-2600
510(k) Number	K203016	K112677
Manufacturer	DAESUNG MAREF CO., LTD.	DAESUNG MAREF CO., LTD.
Product Code	JOW	JOW
Device Class	II	II
Regulation Number	21 CFR 870.5800	21 CFR 870.5800
Regulation Name	Compressible Limb Sleeve	Compressible Limb Sleeve
Indications For Use	A device intended to prevent Deep Vein Thrombosis/Pulmonary Embolism (DVT/PE) by increasing venous blood flow to a patient who has a risk of DVT/PE.	DVT-2600 is a system to prevent DVT (Deep Vein Thrombosis) by improving the blood velocity of patients. DVT-2600 is Circulation Enhancement, Deep Vein Thrombosis Prophylaxis, Edema - Acute, Edema - Chronic, Extremity Pain Incident to Trauma or Surgery Leg Ulcers, Venous Stasis / Venous Insufficiency
Intended Use environment	Professional healthcare environment	Professional healthcare environment

Accessories	Calf sleeve Thigh sleeve Foot sleeve Boots sleeve	Calf sleeve Thigh sleeve Foot sleeve Boots sleeve
Specifications		
Power Source	100-240VAC, 50/60Hz	100-240VAC, 50/60Hz
Pressure	LEG : 20-60mmHg FOOT : 120-140mmHg	LEG : 20-60mmHg FOOT : 120-140mmHg
Number of chamber	3	3
Battery	3200mAh, 3350mAh	-

The table also provides rationale for a little difference in support of substantial equivalence to the Predicate devices.

[Table 2. Little difference with Predicate Device]

Justification to Support Substantial Equivalence
<p>The DVT-2600(Proposed device) is hardly different from the DVT-2600(Predicate device) except for battery. However, the battery is not the main power source of this device, it is an auxiliary power source. In addition, it was confirmed with the product performance test report that it did not affect the performance when using the battery. Therefore, the differences in technological characteristics do not raise different questions of safety and effectiveness.</p>

Non-Clinical Test Summary

The DVT-2600 comply with voluntary standards for electrical safety, electromagnetic compatibility. And a biocompatibility test was conducted for the sleeves, not for device, an accessory used with DVT-2600. The following data were provided in support of the substantial equivalence determination :

1) Electrical Safety, Electromagnetic Compatibility and Performance

The DVT-2600 comply with the electrical safety and electromagnetic compatibility requirements established by the standards.

- IEC 60601 : 2005/A1:2012, Medical Electrical Equipment:Part 1: General Requirements for Basic Safety and Essential Performance
- IEC 60601 60601-1-2:2014, Medical Electrical Equipment - Part 1 1-2: General Requirements for Basic Safety and Essential Performance - Collateral Standard: Electromagnetic Disturbances -Requirements and Tests
- ISO 10993-5:2009 Third edition 2009-12-15, Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity
- ISO 10993-10 Thrid Edition 2010-08-01, Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization

Clinical Test Summary

Clinical testing was not required to demonstrate the substantial equivalence of the DVT-2600 to its predicate device.

8. Conclusion [21 CFR 807.92(b)(3)]

The DVT-2600 has same intended use and technical characteristics to the predicate device except some item. Based on that information, we conclude that the differences between the proposed device and predicate device do not introduce a new intended use and do not raise new issues of safety and effectiveness.