



April 22, 2021

Shockwave Medical, Inc.
Danica Van
Regulatory Affairs Specialist
5403 Betsy Ross Dr
Santa Clara, California 95054

Re: K203365

Trade/Device Name: Shockwave Intravascular Lithotripsy (IVL) System with the Shockwave M5+
Peripheral IVL Catheter

Regulation Number: 21 CFR 870.1250

Regulation Name: Percutaneous Catheter

Regulatory Class: Class II

Product Code: PPN

Dated: March 18, 2021

Received: March 19, 2021

Dear Danica Van:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**Gregory W.
O'Connell -S**

Digitally signed by Gregory
W. O'Connell -S
Date: 2021.04.22 05:04:26
-04'00'

Gregory O'Connell
Assistant Director
DHT2C: Division of Coronary
and Peripheral Intervention Devices
OHT2: Office of Cardiovascular Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K203365

Device Name

Shockwave Intravascular Lithotripsy (IVL) System with the Shockwave M5+ Peripheral IVL Catheter

Indications for Use (Describe)

The Shockwave Medical IVL System is intended for lithotripsy-enhanced balloon dilatation of lesions, including calcified lesions, in the peripheral vasculature, including the iliac, femoral, ilio-femoral, popliteal, infra-popliteal, and renal arteries. Not for use in the coronary or cerebral vasculature.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(k) Summary for K203365

This 510(k) Summary is submitted in accordance with the requirements of the Safe Medical Devices Act (SMDA) of 1990 and 21 CFR 807.92.

Name, Address, and Phone Number of Applicant

Shockwave Medical, Inc.
5403 Betsy Ross Drive
Santa Clara, CA 95054
Phone: (510) 279-4262

Contact Person

Danica Van

Date Prepared

November 13, 2020

Device Name and Classification

Trade Name:	Shockwave Intravascular Lithotripsy (IVL) System with the Shockwave M5+ Peripheral IVL Catheter
Common Name:	Catheter, lithotripsy, peripheral, transluminal
CFR Classification:	21 CFR 870.1250
Classification Name:	Percutaneous catheter
Product Code:	PPN

Predicate Device

The predicate device is the Shockwave Medical Intravascular Lithotripsy System, K180958, cleared by FDA on July 26, 2018.

Indications for Use / Intended Use

The Shockwave Medical IVL System is intended for lithotripsy-enhanced balloon dilatation of lesions, including calcified lesions, in the peripheral vasculature, including the iliac, femoral, ilio-femoral, popliteal, infra-popliteal, and renal arteries. Not for use in the coronary or cerebral vasculature.

Device Description

The IVL Catheter is a proprietary lithotripsy device delivered through the peripheral arterial system of the lower extremities to the site of an otherwise difficult to treat calcified stenosis. Energizing the lithotripsy device will generate pulsatile mechanical energy within the target treatment site, disrupting calcium within the lesion and allowing subsequent dilation of a

peripheral artery stenosis using low balloon pressure. The IVL Catheter is comprised of an integrated balloon with an array of integrated lithotripsy emitters for the localized delivery of pulsatile mechanical energy. The system consists of an IVL Catheter, an IVL Connector Cable and an IVL Generator.

The Shockwave M⁵⁺ Peripheral IVL Catheter shaft contains an inflation lumen, a guidewire lumen, and the lithotripsy emitters. The emitters are positioned along the length of the balloon working length for delivery of pulsatile mechanical energy. The balloon is located near the distal tip of the catheter. Two radiopaque marker bands within the balloon denote the length of the balloon to aid in positioning of the balloon during treatment. The balloon is designed to provide an expandable segment of known length and diameter at a specific pressure.

The IVL Generator and Connector Cable are used with a Shockwave Medical IVL Catheter to deliver localized, lithotripsy-enhanced, balloon dilatation of calcified, stenotic arteries. The IVL Generator, IVL Connector Cable and IVL Catheters are designed to exchange data during patient treatment.

Technological Comparison

This Traditional 510(k) Premarket Notification describes pulse rate and dimensional modifications to the Shockwave M⁵ IVL Catheter, including extending the working length to 135 cm and adding an 8.0mm balloon.

The catheter labeling was updated to reference the name of the modified device, Shockwave M⁵⁺ IVL Catheter. Additionally, the Shockwave M⁵⁺ labeling includes a sterile cable sleeve in the packaging to form a convenience kit.

The IVL System has the same intended use, principles of operation and has substantially equivalent technological characteristics including same fundamental scientific technology, design, energy source, shelf life, and sterilization as the 510(k) cleared IVL System.

Summary of Performance Data

Objective evidence demonstrating that the IVL System design output meets the product design input requirements as well as that device performance characteristics conform to user needs and intended uses as defined in the product specification was provided. Testing was conducted in accordance with Shockwave Medical's Risk Analysis procedures, applicable FDA guidance documents and relevant international standards. Testing included:

- IVL Catheter design verification and validation testing:
 - Guidewire compatibility
 - Introducer sheath compatibility
 - Nominal balloon diameter
 - Balloon diameter at Rated Burst Pressure (RBP)
 - Balloon length
 - Balloon inflation time

- Balloon deflation time
- Useable catheter length
- Balloon crossing profile
- Catheter distal tip
- Distal tip durability
- Catheter bonds tensile strength
- Catheter torsional strength
- Emitter and marker band integrity
- System leakage
- Minimum balloon RBP
- Balloon fatigue (multiple inflations)
- Sonic output
- Catheter pulse count and pulse rate
- Temperature rise
- Catheter particle count
- Catheter connector length
- Catheter connection
- Catheter identification
- Catheter sterility (visual inspection)
- Cable sleeve packaging
- Catheter compatibility with materials and accessories commonly used in Over-the-Wire (OTW) peripheral balloon angioplasty procedures
- Simulated use testing

Results demonstrated that the performance of the IVL System meets its design specifications and demonstrates substantial equivalence for its intended use; therefore, additional clinical data were not required.

Basis for Substantial Equivalence

The IVL Catheter with dimensional modifications shares the same intended use, principles of operation, overall technical and functional capabilities, and similar design and materials as the identified predicate device. Any differences between the IVL Systems were evaluated through design verification and validation testing which demonstrated device performance and confirmed that there are no new questions of safety or effectiveness. The modified IVL Catheter is therefore substantially equivalent to the predicate device.