



September 23, 2021

Nextremity Solutions, Inc.
Elise Fox
Quality and Regulatory Specialist
1195 Polk Drive
Warsaw, Indiana 46582

Re: K211996

Trade/Device Name: Nexta® PEEK Hammertoe Correction System
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener
Regulatory Class: Class II
Product Code: HWC
Dated: June 24, 2021
Received: June 28, 2021

Dear Elise Fox:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, MPH
Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K211996

Device Name

Nextra® PEEK Hammertoe Correction System

Indications for Use (Describe)

The Nextremity Solutions Nextra PEEK Hammertoe Correction System is indicated for small bone reconstruction limited to inter-phalangeal repair and fusion of the lesser toes.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Prepared: September 22, 2021

Submitter: Nextremity Solutions, Inc.
1195 Polk Drive
Warsaw, IN 46582

Contact: Elise Fox
Quality and Regulatory Specialist
elise.fox@nextremity.com
Phone: 574-376-2062
FAX: 574-966-1396

Proprietary Name: Nextra PEEK Hammertoe Correction System

Common Name: Bone Screw System

Classification: 21 CFR § 888.3040 Smooth or threaded metallic bone fixation fastener;
Class II

Product Code: HWC

**Substantially
Equivalent Devices:**

- Nextremity Solutions, Nextra CH Cannulated Hammertoe System (Cleared as DuoHex™ CH Cannulated Hammertoe System), K200840-Primary Predicate
- Smith & Nephew, Hat-Trick Lesser Toe Repair System PIP Fusion (Cleared as MTP Solutions LLC, PEEK Fusion Implant), K133515-Additional Predicate
- Arthrex, TRIM-IT Spin Pin (Cleared as Arthrex Bio-Pin), K050259-Additional Predicate

Device Description:

The Nextra PEEK Hammertoe Correction System consists of a two-part mating bone screw implant construct. The screw implants are provided in various sizes. The proximal screw implants are available in diameters of 3.2mm and 4.2mm. The middle screw implant is available in diameters of 3.5mm, 4.25mm, and 5.0mm, with 0° and 10° angled options. The proximal and distal screw implants are manufactured from polyetheretherketone (PEEK) conforming to ASTM F2026. The system includes implant specific drivers and taps, and other necessary surgical site preparation instruments. The implants and instruments of the system are provided sterile for single use.

Intended Use / Indications:

The Nextremity Solutions Nextra PEEK Hammertoe Correction System is indicated for small bone reconstruction limited to inter-phalangeal repair and fusion of the lesser toes.

Summary of Technologies/Substantial Equivalence:

The Nextra PEEK Hammertoe Correction System is substantially equivalent to the predicate devices regarding the intended use and indications, material, design, sizes, and mechanical properties. Differences between the subject device system and the predicate device systems do not raise new types of safety and effectiveness questions.

Non-Clinical Testing:

Endotoxin testing was performed using the Limulus Amebocyte Lysate (LAL) method according to AAMI ST72, USP 161 and USP 85. Results met the Endotoxin limit of ≤ 20 EU per device. To evaluate the strength of the Nextra PEEK Hammertoe Correction System and components, static 3-point bend tests and dynamic 3-point bend tests were performed on the worst case implant construct according to ASTM F2193-18a. These tests confirmed that the strength of the Nextra PEEK Hammertoe Correction System is substantially equivalent to predicate devices with similar indications and is adequate for its intended use.

Clinical Testing:

Clinical testing was not necessary to demonstrate substantial equivalence of the Nextra PEEK Hammertoe Correction System to the predicate device.

Conclusions/Substantial Equivalence:

The Nextra PEEK Hammertoe Correction System is substantially equivalent to the predicate devices regarding its intended use, material, design, sizes, and mechanical properties. Differences between the subject device system and the predicate device systems do not raise new types of safety and effectiveness questions.