



April 2, 2020

Cutting Edge Spine, LLC
Mr. Kyle Kuntz
Manger R&D
101 Waxhaw Professional Park, Suite A
Waxhaw, North Carolina 28173

Re: K200552
Trade/Device Name: EVOS Lumbar Interbody System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: MAX
Dated: February 25, 2020
Received: March 3, 2020

Dear Mr. Kuntz:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent Showalter, Ph.D.
Assistant Director (Acting)
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K200552

Device Name

EVOS Lumbar Interbody System

Indications for Use (Describe)

The EVOS Lumbar Interbody System is an intervertebral body device intended for use in skeletally mature patients with Degenerative Disc Disease (DDD) of the lumbar spine with up to Grade 1 spondylolisthesis at one or two contiguous levels from L2-S1. Degenerative disc disease is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies.

The EVOS Lumbar Interbody System is intended to be used with autograft and/or allograft comprised of cancellous and/or corticocancellous bone graft to facilitate fusion. It is to be used in patients who have had six months of non-operative treatment and is to be implanted via a direct posterior or transforaminal approach. The EVOS CURVED devices are implanted singly, while the EVOS ROTATE and EVOS STRAIGHT devices may be implanted singly or in pairs in the lumbosacral spine. The EVOS Lumbar Interbody System is intended to be used with supplemental fixation.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(k) Summary

I. SUBMITTER

Date Prepared: 03/28/2020

Applicant:

Cutting Edge Spine, LLC
101 Waxhaw Professional Park Dr., Suite A
Waxhaw, NC 28173

Contact Person: Kyle Kuntz, Manager R&D
Tel: (704) 243-0892
e-mail: k.kuntz@cuttingedgespine.com

Application Correspondents:

Contact Person: Kyle Kuntz, Manager R&D
Tel: (704) 243-0892
e-mail: k.kuntz@cuttingedgespine.com

Alternate Contact: Brad Roof, Director of Operations
Tel: (704) 243-0892
e-mail: b.roof@cuttingedgespine.com

II. DEVICE

EVOS Lumbar Interbody System

Trade Name: EVOS Lumbar Interbody System
Common or Usual Name: Intervertebral Body Fusion Device
Classification Name: Per 21 CFR as follows:
888.3080
Intervertebral Fusion Device with Bone Graft, Lumbar
Regulatory Class: II
Product Codes: MAX



III. PREDICATE DEVICES

	510(k) Number	Device	Manufacturer
Primary Predicate	K150321	EVOS Lumbar Interbody System	Cutting Edge Spine
Additional Predicate	K102957	Cutting Edge Spine Interbody Device	Cutting Edge Spine

IV. DEVICE DESCRIPTION

The EVOS Lumbar Interbody System is comprised of a variety of Implants manufactured from PEEK (Polyetheretherketone) -Optima LT1 (per ASTM F-2026) and PEEK Optima LT120HA, with Tantalum bead markers per ASTM F-560. There are three main configurations; STRAIGHT, CURVED, and ROTATE. The different configurations allow for multiple surgical technique options. The EVOS STRAIGHT and the EVOS ROTATE are generally rectangular in shape while the EVOS CURVED is curved. The EVOS CURVED devices are implanted singly, while the EVOS ROTATE and EVOS STRAIGHT devices may be implanted singly or in pairs.

The EVOS device(s) are available in a range of sizes, as well as flat and biconvex endplates, and with various degrees of lordosis to accommodate variations in patients’ anatomy. The angle of lordosis on each lordotic EVOS device is oriented in a way that will provide a true anterior/posterior lordotic orientation once the device has been implanted in its final intervertebral location. For example, the True Oblique Lordotic (TOL) devices have a lordotic orientation to the device that will provide a correct anterior/posterior lordotic orientation once they have been inserted at 30 degrees from the anterior/posterior midline.

Implant heights range from 6mm to 16mm in maximum height (minus the height of the teeth). Widths range from 8mm to 12mm and lengths range from 22mm to 30mm. The variety of implant shapes and sizes accommodate various surgical technique options. The hollow implants have holes through four sides for bone graft and an inserter instrument interface on the face. Teeth on top and bottom of the spacers improve fixation. The EVOS Lumbar Interbody System includes the instrumentation to facilitate the implantation of the implants.

The approach and the discectomy are conducted using standard instruments while subsequent steps are conducted using standard and/or custom instruments. The system is





comprised of instruments and perforated instrument cases that are generally comprised of aluminum, stainless steel and/or polymeric materials.

The EVOS rotate and non-rotate Inserter(s) are instruments intended for use with the EVOS Lumbar Interbody System. The inserter(s) are designed specifically to implant the EVOS device into a prepared disc space. The EVOS non-rotate inserters are designed to work with the EVOS straight and curved devices while the EVOS rotate inserter is designed to work with the EVOS rotate devices.

V. INDICATIONS FOR USE

The EVOS Lumbar Interbody System is an intervertebral body device intended for use in skeletally mature patients with Degenerative Disc Disease (DDD) of the lumbar spine with up to Grade 1 spondylolisthesis at one or two contiguous levels from L2-S1. Degenerative disc disease is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies.

The EVOS Lumbar Interbody System is intended to be used with autograft and/or allograft comprised of cancellous and/or corticocancellous bone graft to facilitate fusion. It is to be used in patients who have had six months of non-operative treatment and is to be implanted via a direct posterior or transforaminal approach. The EVOS CURVED devices are implanted singly, while the EVOS ROTATE and EVOS STRAIGHT devices may be implanted singly or in pairs in the lumbosacral spine. The EVOS Lumbar Interbody System is intended to be used with supplemental fixation.

VI. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH PREDICATE DEVICES

Documentation was submitted which demonstrated that the EVOS Lumbar Interbody System are substantially equivalent to the predicate devices based on a comparison of the following characteristics:

- FDA product codes
- Indications for Use
- Anatomical Region
- Implant Materials
- Product Dimensions
- Device Features
- Mechanical Performance
- Available by prescription only
- Made for single use



VII. NON-CLINICAL AND CLINICAL PERFORMANCE TESTING

No design changes were made to the existing devices, nor were any new components added to the system. Therefore, no additional testing was required or performed.

VIII. CONCLUSIONS

No design changes were made to the existing devices, nor were any new components added to the system. The sterilization method was expanded to include the use of steam sterilization. The data presented in this submission demonstrates that the devices listed above are substantially equivalent to the predicate devices.

