



June 22, 2022

Hangzhou Runheng Medical Co., Ltd.  
Yaya Lu  
QC Manager  
Room 201, Shunfeng Building, NO.109 Hongxing Road,  
Qiaonan Block  
Hangzhou, Zhejiang 311215  
China

Re: K213739

Trade/Device Name: Nitrile Examination Gloves  
Regulation Number: 21 CFR 880.6250  
Regulation Name: Non-Powdered Patient Examination Glove  
Regulatory Class: Class I, reserved  
Product Code: LZA  
Dated: May 10, 2022  
Received: May 25, 2022

Dear Yaya Lu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Bifeng Qian, M.D., Ph. D.  
Acting Assistant Director  
DHT4B: Division of Infection Control  
and Plastic Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K213739

Device Name  
Nitrile Examination Gloves

Indications for Use (Describe)

Nitrile Examination Gloves are disposable devices intended for medical purpose that are worn on the examiner's hand to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## **K213739-510(k) summary**

### **I Submitter**

Device submitter: Hangzhou Runheng Medical Co., Ltd.  
Room 201, Shunfeng Building, NO.109 Hongxing Road, Qiaonan  
Block, Xiaoshan Economic and Technological Development Zone,  
Hangzhou, Zhejiang Province, P.R.China.

Contact person: Yaya Lu  
QC Manager  
Phone: +86 18285158974  
E-mail: 1264726347@qq.com

Date of Preparation: May 16<sup>th</sup>, 2022

### **II Proposed Device**

510(k) Number: K213739  
Trade/Device Name: Nitrile Examination Gloves  
Regulation Number: 21 CFR 880.6250  
Regulation Name: Non-powdered patient examination glove  
Regulatory Class: Class I  
Product code: LZA  
Review Panel: General Hospital

### **III Predicate Devices**

510(k) Number: K211515  
Trade/Device Name: Nitrile Examination Gloves  
Regulation Number: 21 CFR 880.6250  
Regulation Name: Non-powdered patient examination glove  
Classification: Class I  
Product Code: LZA  
Manufacturer: Beijing Reagent Latex Products Co., Ltd.

### **IV Device description**

Nitrile Examination Gloves are made of Nitrile rubber and are blue in color. The device is powder free nitrile examination gloves. It can be available in five specifications: XS, S, M, L and XL.

The subject device is non-sterile.

### V Indication for use

Nitrile Examination Gloves are disposable devices intended for medical purpose that are worn on the examiner's hand to prevent contamination between patient and examiner.

### VI Non-Clinical Test Conclusion

Non-clinical tests were conducted to verify that the proposed device met all design specifications as was same/similar to the predicate devices. The test results demonstrated that the proposed device complies with the following standards:

- ASTM D5151 Standard Test Method for Detection of Holes in Medical Gloves
- ASTM D6124 Standard Test Method for Residual Powder on Medical Gloves
- ASTM D6319 Standard Specification for Nitrile Examination Gloves for Medical Application
- ISO 10993-10: 2010 Biological evaluation of medical devices-Part 10: Tests for irritation and skin sensitization
- ISO 10993-11: 2017 Biological evaluation of medical devices-Part 11: Tests for systemic toxicity

Table 1 Summary of Non-Clinical Performance Testing

Test Method	Purpose	Acceptance Criteria		Result	
ASTM D6319	Physical Dimensions Test	Extra-Small: Length: $\geq 220$ mm; Width: $70 \pm 10$ mm		Pass	
		Small: Length: $\geq 220$ mm; Width: $80 \pm 10$ mm			
	Medium: Length: $\geq 230$ mm; Width: $95 \pm 10$ mm				
	Large: Length: $\geq 230$ mm; Width: $110 \pm 10$ mm				
		Extra- Large: Length: $\geq 230$ mm; Width: $120 \pm 10$ mm			
		Thickness (mm): Finger: $\geq 0.05$ Palm: $\geq 0.05$		Pass	
ASTM D6319	Physical properties	Before Aging	Tensile Strength	$\geq 14$ MPa	Pass
			Ultimate Elongation	$\geq 500\%$	
		After Aging	Tensile Strength	$\geq 14$ MPa	Pass
			Ultimate Elongation	$\geq 500\%$	
ASTM D5151	Watertightness Test for Detection of Holes	Meet the requirements of ASTM D5151 Test for AQL 2.5		Pass	

ASTM D6124	Powder Content	Meet the requirements of ASTM D6124 < 2.0mg	Pass
ISO 10993-11	To determine if the finished device material extracts pose a systemic toxicity concern.	Non- acute systemic toxicity	Under conditions of the study, did not show acute systemic toxicity in vivo / Pass
ISO 10993-10	To determine if the finished device material is an irritant.	Non-irritating	Under the conditions of the study not an irritant/ Pass
ISO 10993-10	To determine if the finished device material is a sensitizer.	Non-sensitizing	Under conditions of the study, not a sensitizer. / Pass

### VII Clinical Test Conclusion

No clinical study is included in this submission.

### VIII Summary of Technological characteristics

Table 2 Technological Characteristics Comparison Table

Item	Subject device	Predicate device	Discussion
Product name	Nitrile Examination Gloves	Snow Lotus Nitrile Examination Gloves	NA
510(k) Number	K213739	K211515	NA
Product Code	LZA	LZA	Identical
Intended use	Nitrile Examination Gloves are disposable devices intended for medical purpose that are worn on the examiner's hand to prevent contamination between patient and examiner.	Beijing Reagent Latex Products Co., Ltd Nitrile Examination Gloves are disposable devices intended for medical purpose that are worn on the examiner's hand to prevent contamination between patient and examiner.	Identical
Dimensions	Extra-Small: Length: Minimum 220mm Width: 70 ± 10 mm	Extra-Small: None	Different 1
	Small: Length: Minimum 220mm	Small: Length: Minimum 220mm	

	Width: $80 \pm 10$ mm	Width: $80 \pm 10$ mm	
	Medium: Length: Minimum 230mm Width: $95 \pm 10$ mm	Medium: Length: Minimum 230mm Width: $95 \pm 10$ mm	
	Large: Length: Minimum 230mm Width: $110 \pm 10$ mm	Large: Length: Minimum 230mm Width: $110 \pm 10$ mm	
	Extra- Large: Length: Minimum 230mm Width: $120 \pm 10$ mm	Extra- Large: Length: Minimum 230mm Width: $120 \pm 10$ mm	
Thickness per ASTM D6319 (for all sizes)	Palm thickness: Minimum 0.05 mm	Palm thickness: Minimum 0.05 mm	Identical
	Finger thickness: Minimum 0.05 mm	Finger thickness: Minimum 0.05 mm	
Physical Properties Before Aging per ASTM D6319	Tensile Strength: Minimum 14 MPa Ultimate Elongation: Minimum 500%	Tensile Strength: Minimum 14 MPa Ultimate Elongation: Minimum 500%	Identical
Physical Properties After Aging per ASTM D6319	Tensile Strength: Minimum 14 MPa Ultimate Elongation: Minimum 400%	Tensile Strength: Minimum 14 MPa Ultimate Elongation: Minimum 400%	Identical
Water tight (hole detection) per ASTM D5151	Passes at AQL of 2.5	Passes at AQL of 2.5	Identical
Powder Residue per ASTM D6319	$\leq 2$ mg/glove	$\leq 2$ mg/glove	Identical
Biocompatibility: Primary Skin Irritation per ISO 10993-10	Not an irritant under the conditions of the study.	Not an irritant under the conditions of the study.	Identical
Biocompatibility: Dermal Sensitization per ISO 10993-10	Not a sensitizer under the conditions of the study.	Not a sensitizer under the conditions of the study.	Identical
Biocompatibility: Acute Systemic Toxicity Test per ISO 10993-11	Device extracts do not pose a systemic toxicity concern under the conditions of the study.	Device extracts do not pose a systemic toxicity concern under the conditions of the study.	Identical

Device Material	Nitrile	Nitrile	Identical
Color	Blue	Blue	Identical
Size Offering	Extra-Small, Small, Medium, Large, Extra-Large	Small, Medium, Large, Extra-Large	Different 1
Number of Uses	Single Use	Single Use	Identical

**Different 1:**

The models of proposed device are different with predicate device, but all proposed devices are meet the specifications of ASTM D 6319. So we consider this as the proposed device is similar to the predicate device.

**IX Conclusion**

The conclusion drawn from the nonclinical tests demonstrates that the subject device in 510(k) submission K213739, Nitrile Examination Gloves, is as safe, as effective, and performs as well as or better than the legally marketed predicate device, Snow Lotus Nitrile Examination Gloves, cleared under 510(k) K211515.