



October 14, 2022

Southern Medical (Pty) Ltd
% Nathan Wright
Engineer and Regulatory Specialist
Empirical Technologies
4628 Northpark Drive
Colorado Springs, Colorado 80918

Re: K221447

Trade/Device Name: Southern Anterior Screw Fixated Cage (SASCA)
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral body fusion device
Regulatory Class: Class II
Product Code: OVD
Dated: September 14, 2022
Received: September 15, 2022

Dear Nathan Wright:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpnm/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for
Brent Showalter, Ph.D.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K221447

Device Name

Southern Anterior Screw Fixated Cage (SASCA)

Indications for Use (Describe)

Southern Anterior Screw Fixated Cages are indicated for use as stand-alone anterior interbody fusion devices in the lumbar spine. The devices are designed to be used with the bone screws provided and the interior cavity must be filled with autograft. Supplementary spinal fixation is not required.

The devices are indicated for use in skeletally mature patients with degenerative disc disease (DDD) at one or two contiguous levels from L1 to S1. DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies. These DDD patients may also have up to Grade I spondylolisthesis at the involved level(s). The devices may also be used in patients with pseudarthrosis / non-union from previous unsuccessful fusion surgery. Patients should have undergone at least six months of non-operative treatment.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection including suggestions for reducing this burden to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY

Submitter's Name:	Southern Medical (Pty) Ltd
Submitter's Address:	Building 10, Southern Implants Office Park 1 Albert Road Irene, Gauteng 0062 South Africa
Submitter's Telephone:	+27 12 667 6243/4
Contact Person:	Nathan Wright MS Empirical Technologies 1-719-351-0248 nwright@empiricaltech.com
Date Summary was Prepared:	May 17, 2022
Trade or Proprietary Name:	Southern Anterior Screw Fixated Cage (SASCA)
Common or Usual Name:	Intervertebral Fusion Device With Integrated Fixation, Lumbar
Classification:	Class II per 21 CFR §888.3080
Product Code:	OVD
Classification Panel:	Orthopedic – Spinal Devices (DHT6B)

DESCRIPTION OF THE DEVICE SUBJECT TO PREMARKET NOTIFICATION:

The Southern Anterior Screw Fixated Cage (SASCA) is intended to treat pain and functional disorders of the lumbar spine by supporting, stabilizing, and immobilizing the affected segment and facilitating fusion. The SASCA devices are stand-alone devices offered in a variety of sizes to accommodate individual patient anatomy. The SASCA cages are manufactured from PEEK per ASTM F2026 with tantalum markers per ASTM F560. The cages are offered with or without titanium plasma coating per ASTM F1580. The fixation screws are manufactured from Ti-6Al-4V per ASTM F136. Note: The subject SASCA device is also labeled in markets outside the United States as the Sasca 2 Anterior Lumbar Standalone Fusion Cage.

INDICATIONS FOR USE

The Southern Anterior Screw Fixated Cages are indicated for use as stand-alone anterior interbody fusion devices in the lumbar spine. The devices are designed to be used with the bone screws provided and the interior cavity must be filled with autograft. Supplementary spinal fixation is not required.

The devices are indicated for use in skeletally mature patients with degenerative disc disease (DDD) at one or two contiguous levels from L1 to S1. DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies. These DDD patients may also have up to Grade I spondylolisthesis at the involved level(s). The devices may also be used in patients with pseudarthrosis / non-union from previous unsuccessful fusion surgery. Patients should have undergone at least six months of non-operative treatment.

TECHNOLOGICAL CHARACTERISTICS

The subject and predicate devices have nearly identical technological characteristics and the minor differences do not raise any new issues of safety and effectiveness. Specifically, the following characteristics are similar between the subject and predicates:

- Indications for Use
- Materials of manufacture
- Structural support mechanism
- Sizes

Table 5-1 Predicate Devices

510k Number	Trade or Proprietary or Model Name	Manufacturer	Product Code	Predicate Type
K173347	STALIF M FLX™	Centinel Spine, Inc.	OVD	Primary
K183426	MectaLIF Anterior Stand Alone System	Medacta International SA	OVD	Additional

PERFORMANCE DATA

The SASCA has been tested in the following test modes:

- Static and Dynamic Axial Compression per ASTM F2077
- Static and Dynamic Compression Shear per ASTM F2077
- Static Torsion per ASTM F2077
- Subsidence per ASTM F2267
- Expulsion

The results of this non-clinical testing show that the strength of the SASCA is sufficient for its intended use and is substantially equivalent to legally marketed predicate devices.

CONCLUSION

The overall technology characteristics and mechanical performance data lead to the conclusion that the SASCA is substantially equivalent to the predicate device.