



Philips Healthcare (Suzhou) Co., Lit.  
% Claire Zhang  
Advanced Regulatory Engineer  
No. 258, ZhongYuan Road, Suzhou Industrial Park  
Suzhou, Jiangsu 215024  
CHINA

July 10, 2020

Re: K201725  
Trade/Device Name: DigitalDiagnost C50  
Regulation Number: 21 CFR 892.1680  
Regulation Name: Stationary x-ray system  
Regulatory Class: Class II  
Product Code: KPR  
Dated: June 8, 2020  
Received: June 23, 2020

Dear Claire Zhang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For

Thalia T. Mills, Ph.D.  
Director  
Division of Radiological Health  
OHT7: Office of In Vitro Diagnostics  
and Radiological Health  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K201725

Device Name  
DigitalDiagnost C50

### Indications for Use (Describe)

The DigitalDiagnost C50 system is intended for use in generating radiographic images of human anatomy by qualified/trained doctor or technician. Applications can be performed with the patient sitting, standing, or lying in the prone or supine position. This device is not intended for mammographic applications.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

510(k) number: K201725

## 510(k) Summary

### SPECIAL 510(k) Summary of Safety and Effectiveness

This 510(k) summary of safety and effectiveness information is prepared in accordance with 21 CFR §807.92.

**Date Prepared:** June 08, 2020  
**Manufacturer:** Philips Healthcare (Suzhou) Co., Ltd.  
No. 258, ZhongYuan Road, Suzhou Industrial Park, 215024  
Suzhou, Jiangsu Province, PEOPLE'S REPUBLIC OF CHINA  
Establishment Registration Number: 3009529630

**Contact Person:** Claire Zhang  
Advanced Regulatory Engineer  
Phone: +86-021-24128746 (cell: +86-133-8217-7838)  
Fax: +86-512-68018677  
E-mail: Claire.Zhang\_1@philips.com

**Device Name:** **DigitalDiagnost C50**  
**Classification:** Classification Name: Stationary X-Ray System  
Classification Regulation: 21CFR §892.1680  
Classification Panel: Radiology  
Device Class: Class II  
Product code: KPR (System, X-Ray, Stationary)

**Predicate Device:** Trade Name: DigitalDiagnost C50  
Manufacturer: Philips Healthcare (Suzhou) Co., Ltd.  
510(k) Clearance: K163410-January 4, 2017  
Classification Regulation: 21 CFR, Part 892.1680  
Classification Name: Stationary X-Ray System  
Classification Panel: Radiology  
Device Class: Class II  
Product Code: KPR

**Reference Device 1:** Trade Name: MobileDiagnost WDR 2.2  
Manufacturer: SEDECAL SA  
510(k) Clearance: K191813- August 2, 2019  
Classification Regulation: 21CFR 892.1720  
Classification Name: Mobile x-ray system  
Classification Panel: Radiology  
Device Class: Class II  
Product Code: IZL, MQB

**Reference Device 2:** Trade Name: Philips Eleva Workspot with SkyFlow  
 Manufacturer: Philips Medical Systems DMC GmbH  
 510(k) Clearance: K153318- December 22, 2015  
 Classification Regulation: 21 CFR 892.1680  
 Classification Name: Stationary X-Ray System  
 Classification Panel: Radiology  
 Device Class: Class II  
 Product Code: MQB, LLZ

**Reference Device 3:** Trade Name: DigitalDiagnost C90  
 Manufacturer: Philips Medical Systems DMC GmbH  
 510(k) Clearance: K182973- January 11, 2019  
 Classification Regulation: 21CFR 892.1680  
 Classification Name: Stationary X-Ray System  
 Classification Panel: Radiology  
 Device Class: Class II  
 Product Code: MQB, KPR, LLZ

**Device description:** The **DigitalDiagnost C50 Digital Radiography System (DigitalDiagnost C50)** is a flexible digital radiography (DR) system that is designed to provide fast and smooth radiography examinations of sitting, standing or lying patients.  
 The **DigitalDiagnost C50** consist of the following components: ceiling suspension with X-ray assembly, wall stand with detector carrier, patient table with detector carrier and floating table top, high voltage generator, and an acquisition and reviewing workstation for post-processing, storage and viewing of images. Images may be transferred via a DICOM network for printing, storage and detailed review.

**Indications for use:** The **DigitalDiagnost C50** is intended for use in generating radiographic images of human anatomy by qualified/trained doctor or technician. Applications can be performed with the patient sitting, standing, or lying in the prone or supine position. This device is not intended for mammographic applications.

**Fundamental scientific technology:** The fundamental scientific technology utilized in the **DigitalDiagnost C50** and the currently marketed and

predicate DigitalDiagnost C50 (K163410, January 4, 2017) is equivalent with regards to the functionality of the following components: integrated tube assembly, patient table with a floating table top, high-voltage generator, dual-focus rotation anode X-Ray tube, manual beam limiting device, digital detector, wall stand and workstation for images post-processing, storage and viewing (See Table 1 comparing the **DigitalDiagnost C50** to the currently marketed and predicate DigitalDiagnost C50 (K163410, January 4, 2017) provided below).

The outcome of this comparison demonstrates that the minor differences in the technological characteristics do not affect the safety or effectiveness of the **DigitalDiagnost C50** when compared to the currently marketed and predicate DigitalDiagnost C50 (K163410, January 4, 2017).

The wireless portable detector of the **DigitalDiagnost C50** is identical to the wireless portable detector (SkyPlate E) of the currently marketed and reference device 1, MobileDiagnost WDR 2.2 (K191813-August 2, 2019) manufactured by SEDECAL SA. Therefore, both the wireless portable detector (Skyplate E) of the **DigitalDiagnost C50** and the currently marketed and reference device 1, MobileDiagnost WDR 2.2 employ identical fundamental scientific technology.

The **DigitalDiagnost C50** and the currently marketed and Reference Device 3, DigitalDiagnost C90 (K182973- January 11, 2019) manufactured by Philips Medical Systems DMC GmbH are provided with identical, fixed RAD detector (Pixium 4343RCE), UNIQUE 2 Post Processing software, embedded Windows 10 operating system. Therefore, **DigitalDiagnost C50** and the currently marketed and reference device 3, DigitalDiagnost C90 employ identical fundamental scientific technology in fixed RAD detector (Pixium 4343RCE), UNIQUE 2 Post Processing software, embedded Windows 10 operating system.

<b>Table 1 Comparison of Technological Characteristics of Currently marketed and Predicate DigitalDiagnost C50 versus the DigitalDiagnost C50</b>			
<b>Feature</b>	<b>Predicate Device: DigitalDiagnost C50 (K163410, January 4, 2017)</b>	<b>Device: DigitalDiagnost C50</b>	<b>Comment</b>
<b>Basic information</b>			
Product Code	KPR	Identical	No difference; thus, demonstrating SE.
Regulation No.	21 CFR 892.1680	Identical	No difference; thus, demonstrating SE.
Device Class	II	Identical	No difference; thus, demonstrating SE.
Electrical Requirement	Input voltage: 3-phase, 200/208/240/380/400/415/440/480/500Vac; Frequency:50/60Hz; Current: Short term: 112A (with generator M-CABINET CXA 50kW), 134A (M-CABINET CXA 65kW); Long term: 10A.	Identical	No difference; thus, demonstrating SE.
<b>Design characteristic</b>			
X-ray Tube	RO 1750 ROT 380 & SRO 33100 ROT 380	Identical	No difference; thus, demonstrating SE.
Max Tube Voltage	150 kV	Identical	No difference; thus, demonstrating SE.
Focal Spot Size	0.6mm/1.2mm	Identical	No difference; thus, demonstrating SE.
Tube Max power	50KW/100KW (250W equivalent anode input power)	Identical	No difference; thus, demonstrating SE.
Anode Type	Rotation	Identical	No difference; thus, demonstrating SE.
Generator	Philips Healthcare (Suzhou), M-CABINET CXA Pro 50kW, M-CABINET CXA Pro 65kW	Identical	No difference; thus, demonstrating SE.
Max Power	50KW/65KW	Identical	No difference; thus, demonstrating SE.
KV range	40-150	Identical	No difference; thus, demonstrating SE.
Milli ampere sec (mAs) product	0.4 mAs-600 mAs (with AEC control)	Identical	No difference; thus, demonstrating SE.
<b>Collimator</b>			
Operation Mode	Manual collimation	Identical	No difference; thus, demonstrating SE.
Shape of Beam	Rectangular	Identical	No difference; thus, demonstrating SE.
<b>Detector</b>			

Type	Digital Detector <ul style="list-style-type: none"> <li>Fixed RAD Detector</li> <li>Wireless Static Detector</li> </ul>	Identical	No difference; thus, Demonstrating SE.
Fixed RAD Detector	Pixium 4343RG	Pixium 4343RCE  (Note: this detector is identical to fixed RAD detector of the currently marketed and Reference Device 3, DigitalDiagnost C90 cleared under K182973).	The difference between 4343RCE and 4343RG is scintillator material and other minor differences of image area and image matrix as compared in this table. The differences don't affect the safety or effectiveness. And the fixed RAD detector Pixium 4343RCE of the <b>DigitalDiagnost C50</b> is identical to the fixed RAD detector of the currently marketed and Reference Device 3, DigitalDiagnost C90(K182973- January 11, 2019, Philips Medical Systems DMC). Thus, demonstrating SE
Wireless Static Detector	Varian PaxScan4336W	SkyPlate E (Trixell 3543DR)  (Note: this detector is identical to the Skyplate E Detector of the currently marketed and reference device 1, MobileDiagnost WDR 2.2 cleared under K191813).	The Skyplate E Detector of the proposed <b>DigitalDiagnost C50</b> is identical to the Skyplate E Detector of the currently marketed and Reference Device 1, MobileDiagnost WDR 2.2 (K191813- August 2, 2019). Thus, demonstrating SE
X-ray Scintillator Material	<ul style="list-style-type: none"> <li>GdOS (Fixed: Pixium 4343RG)</li> </ul>	<ul style="list-style-type: none"> <li>Cesium Iodide (Fixed: Pixium 4343RCE. )</li> </ul> <p>Note: Pixium 4343RCE is identical to fixed RAD detector of the currently marketed and Reference Device 3, DigitalDiagnost C90 cleared under K182973.</p>	The fixed RAD detector of the proposed <b>DigitalDiagnost C50</b> and currently marketed and Reference Device 3, DigitalDiagnost C90(K182973- January 11, 2019, Philips Medical Systems DMC) are both fabricated from identical Cesium Iodide material , thus demonstrating SE.
	<ul style="list-style-type: none"> <li>GdOS (Wireless Varian PaxScan4336W)</li> </ul>	<ul style="list-style-type: none"> <li>Cesium Iodide (Wireless SkyPlate E)</li> </ul>	The Wireless Detector of the proposed <b>DigitalDiagnost C50</b> and the currently marketed and Reference Device 1, MobileDiagnost WDR 2.2 (K191813- August 2, 2019) are both fabricated from identical Cesium Iodide material, thus demonstrating SE.



Image Area	<ul style="list-style-type: none"> <li>42.5cm x 42.5cm (Fixed: Pixium 4343RG)</li> </ul>	<ul style="list-style-type: none"> <li>42.03cm x 42.54cm (Fixed: Pixium 4343RCE)</li> </ul>	<p>The image area of the <b>DigitalDiagnost C50</b>, provided with fixed RAD detector is identical to the image area of the Fixed RAD Detector of the currently marketed and Reference Device 3, DigitalDiagnost C90(K182973- January 11, 2019, Philips Medical Systems DMC), thus demonstrating substantial equivalence</p> <p>The image area of the proposed Wireless Detector SkyPlate E is identical to the Wireless Detector SkyPlate E of the currently marketed and reference device 1, MobileDiagnost WDR 2.2 (K191813- August 2, 2019, Sedecal SA,). thus demonstrating SE.</p>
	<ul style="list-style-type: none"> <li>42.7 cm x 34.4 cm (Wireless: Varian PaxScan4336W)</li> </ul>	<ul style="list-style-type: none"> <li>34.5 cm x 42.5cm (Wireless: SkyPlate E)</li> </ul>	
Image Matrix	<ul style="list-style-type: none"> <li>2,874 x 2,869 (Fixed: Pixium 4343RG)</li> </ul>	<ul style="list-style-type: none"> <li>2,874 x 2840 (Fixed: Pixium 4343RCE)</li> </ul>	<p>The image matrix of the proposed <b>DigitalDiagnost C50</b>, provided with fixed RAD detector is similar to the image matrix of the currently marketed and Reference Device 3, DigitalDiagnost C90(K182973- January 11, 2019, Philips Medical Systems DMC). thus demonstrating substantial equivalence</p> <p>The image matrix of the proposed <b>DigitalDiagnost C50</b> provided with wireless detector SkyPlate E is similar to the currently marketed and Reference Device 1, MobileDiagnost WDR 2.2 (K191813- August 2, 2019, Sedecal SA). Infinitesimal change in the image size (X-ray field) does not impact clinical Image Quality. Therefore, they are equivalent and there is no impact on the safety and effectiveness of the device; thus, demonstrating SE.</p>
	<ul style="list-style-type: none"> <li>3,072 x 2,476 (Wireless: Varian PaxScan4336W)</li> </ul>	<ul style="list-style-type: none"> <li>2,156 x 2,653 (Wireless-SkyPlate E)</li> </ul>	

Pixel Size	148 µm (Fixed: Pixium 4343RG)	<ul style="list-style-type: none"> <li>148 µm (Fixed: Pixium 4343RCE)</li> </ul>	<p>The pixel size of the proposed <b>DigitalDiagnost C50</b> provided with fixed RAD detector is identical to the pixel size of the currently marketed and Reference Device 3, DigitalDiagnost C90(K182973- January 11, 2019,). thus demonstrating substantial equivalence.</p> <p>The pixel size of the proposed <b>DigitalDiagnost C50</b> provided with wireless detector SkyPlate E is identical to the currently marketed and Reference Device 1, MobileDiagnost WDR 2.2 (K191813- August 2, 2019, Sedecal SA), thus demonstrating SE.</p>
	139 µm (Wireless: Varian PaxScan4336W)	<ul style="list-style-type: none"> <li>160 µm (Wireless- SkyPlate E)</li> </ul>	
Analog / Digital (A/D) conversion	16 bits	Identical	No difference; thus, demonstrating SE.
<b>Table</b>			
Table type	Fixed and Height adjustment	Identical	No difference; thus, demonstrating SE.
Height adjustment	51.5 cm to 91.5 cm above floor, motorized adjustment	Identical	No difference; thus, demonstrating SE.
Tabletop longitudinal travel range	+/- 550mm	Identical	No difference; thus, demonstrating SE.
Tabletop Lateral travel	+/- 130mm	Identical	No difference; thus, demonstrating SE.
Loading (patient weight)	210 Kg	Identical	No difference; thus, demonstrating SE.
<b>Wall Stand</b>			
Vertical movement range	1500mm	Identical	No difference; thus, demonstrating SE.
Movement mode	Manual	Identical	No difference; thus, demonstrating SE..
<b>Tube Stand</b>			
Type	Ceiling suspension	Identical	No difference; thus, demonstrating SE.
Tube rotation	+/- 135 degree	Identical	No difference; thus, demonstrating SE.
Longitudinal movement range	1500 mm	Identical	No difference; thus, demonstrating SE.
<b>Source to Image Distance (SID)</b>			
SID	SID depends on different configurations, because	Identical	No difference; thus, demonstrating SE.

	the DigitalDiagnost C50 is a ceiling suspension X-ray system.		
<b>External Connectivity</b>			
DICOM	DICOM 3.0 compatible	Identical	No difference; thus, demonstrating SE.
<b>Software Platform</b>			
Software	Eleva WorkSpot	Identical	No difference; thus, demonstrating SE.
SkyFlow Software	No	Yes	The proposed <b>Digital Diagnost C50</b> includes the SkyFlow software used in the currently marketed and reference device Eleva Workspot with SkyFlow, (K153318- December 22, 2015, Philips Medical Systems DMC), thus, demonstrating SE.
Image Processing Algorithm	UNIQUE	UNIQUE 2	UNIQUE 2 image processing algorithm provided with <b>DigitalDiagnost C50</b> was previously cleared with the currently marketed and Reference Device 3, DigitalDiagnost C90 (K182973- January 11, 2019, Philips Medical Systems DMC). Upgrading to UNIQUE2 image processing algorithm does not alter the clinical workflow, hence no impact on the safety or effectiveness of the device; thus, demonstrating SE
Operating System	Windows 7 embedded	Windows 10 embedded	Introduction of operating system Windows 10 embedded does not impact clinical image quality. Therefore, there is no impact on the safety and effectiveness of the device; thus, demonstrating SE.

Based on the information provided above, the **DigitalDiagnost C50** is considered substantially equivalent to the currently marketed and predicate **DigitalDiagnost C50** ( K163410, January 4, 2017) in terms of fundamental scientific technology.

**Summary of Non-Clinical Performance data:**

This 510(k) premarket notification contains the technical documentation, which demonstrates that the **DigitalDiagnost C50** is substantially equivalent to the currently marketed and predicate **DigitalDiagnost C50** ( K163410, January 4, 2017). The technical documentation includes non-clinical verification / validation tests. These tests were performed on the DigitalDiagnost C50 according to the following international and FDA-recognized consensus standards:

- International and FDA-recognized consensus standards:
  - AAMI / ANSI ES60601-1: 2005/(R)2012 and C1:2009/(R)2012 and, A2:2010/(R)2012 (consolidated text) Medical electrical equipment –Part 1: General requirements for basic safety and essential performance. (Edition 3.1). FDA/CDRH recognition number 19-4.
  - IEC 60601-1-2, Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic Disturbances – Requirements and tests (Edition 4.0 2014). FDA/CDRH recognition number 19-8.
  - IEC 60601-1-3, Medical Electrical Equipment Part 1-3: General Requirements for Basic Safety and Essential Performance-Collateral Standard: Radiation Protection in Diagnostic X-Ray Equipment. (Edition 2.1 2013). FDA/CDRH recognition number 12-269.
  - IEC60601-2-28 Medical electrical equipment - Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis, (Edition 2.0 2010-03). FDA/CDRH recognition number 12-204.
  - IEC 60601-2-54, Medical Electrical Equipment-Part 2-54: Particular Requirements for the Basic

Safety and Essential Performance of X-Ray Equipment for Radiography and Radioscopy (Edition 1.1 2015). FDA/CDRH recognition number 12-296.

- IEC 60601-1-6, Medical Electrical Equipment Part 1-6: General Requirements for Basic Safety and Essential Performance- Collateral Standard: Usability (Edition 3.1 2013). FDA/CDRH recognition number 5-89.
- IEC 62304 Medical device software - Software life cycle processes (Edition 1.1 2015) FDA/CDRH recognition number 13-79
- IEC 62366-1 Medical devices - Part 1: Application of usability engineering to medical devices (Edition 1.0 2015). FDA/CDRH recognition number 5-114.
- ISO 14971 Medical devices – Application of risk management to medical devices (Edition 2.0, corrected version, 2007). FDA/CDRH recognition number 5-40.
- CFR 1020.30 Diagnostic x-ray systems and their major components.
- CFR 1020.31 Radiographic equipment.
- Device specific guidance document, entitled “Guidance for the Submission of 510(k)s for Solid State X-ray Imaging Devices – September 1, 2016”
- FDA’s Guidance document entitled, “Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices – May 11, 2005.”
- FDA’s Guidance document entitled, “Guidance for Industry and FDA Staff – Content of Premarket Submissions for Management of Cybersecurity in Medical Devices” ,issued October 2, 2014
- Pediatric Information for X-ray Imaging Device Premarket Notifications, issued November 28, 2017

Non-clinical verification and validation tests have been performed with regards to the intended use, the technical claims, requirement specifications, and the risk management results.

Non-clinical verification and validation test results demonstrate that **DigitalDiagnost C50** :

- Complies with the aforementioned international and FDA-recognized consensus standards and FDA guidance documents.
- meets the acceptance criteria and is adequate for its intended use.

Therefore, **DigitalDiagnost C50** is substantially equivalent to the currently marketed and predicate DigitalDiagnost C50 (K163410, January 4, 2017) in terms of safety and effectiveness.

**Summary of Clinical Data:**

The **DigitalDiagnost C50** does not require clinical study since substantial equivalence to the primary currently marketed and predicate device was demonstrated with the following attributes:

- Indication for use;
- Fundamental scientific technology;
- Non-clinical performance testing; and
- Safety and effectiveness.

Furthermore, the SkyPlate E detector and Pixium 4343RCE detector also has similar design, technology and Image acquisition workflow compared to the previously Pixium 4343RG and Varian PaxScan4336W detector used in the marketed and predicate DigitalDiagnost C50 (K163410, January 4, 2017). All technical detector characteristics that potentially have an influence on image quality are assessed and verified according to FDA Guidance for Industry and Food and Drug Administration Staff: Guidance for the Submission of 510(k)'s for Solid State X-ray Imaging Devices; issued on September 1, 2016.

**Substantial Equivalence Conclusion:**

The **DigitalDiagnost C50** is substantially equivalent to the currently marketed and predicate DigitalDiagnost C50 (K163410, January 4, 2017) in terms of design features, fundamental scientific technology, indications for use, and safety and effectiveness.

Additionally, substantial equivalence was demonstrated with non-clinical performance (verification and

validation) tests, which complied with the requirements specified in the international and FDA-recognized consensus standards, ISO 14971, IEC 60601-1, IEC 60601-1-2, IEC 60601-1-3, IEC 60601-2-54, IEC 60601-1-6, IEC 60601-2-28, IEC 62304, and IEC 62366-1.

The results of these tests demonstrate that **DigitalDiagnost C50** met the acceptance criteria and is adequate for its intended use.