



February 22, 2021

Surnic Corporation  
Helen Xie  
RA Assistant  
480 Apollo Street, Suite D.  
Brea, California 92821

Re: K203522

Trade/Device Name: Surn5 Electrosurgical Argon Beam Generator and Accessories  
Regulation Number: 21 CFR 878.4400  
Regulation Name: Electrosurgical Cutting and Coagulation Device and Accessories  
Regulatory Class: Class II  
Product Code: GEI  
Dated: December 29, 2020  
Received: December 31, 2020

Dear Helen Xie:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Long Chen, Ph.D.  
Assistant Director  
DHT4A: Division of General Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K203522

Device Name  
Surn5 Electrosurgical Argon Beam Generator and Accessories

### Indications for Use (Describe)

The Surn5 Electrosurgical Argon Beam Generator and Accessories are indicated for monopolar or bipolar surgery to achieve argon beam and cut or coagulation for the tissue. It is intended to be used with monopolar handpiece and dispersive electrode or bipolar handpiece and footswitches.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



# 510(k) Summary

Traditional 510(k) Submission

K203522

## Section 5

Version	1
Pages	3
Date	Feb-19-2021

### 1. Submitter Information

Company: Surnic Corporation  
Address: Suite D, 480 Apollo Street, Brea CA 92821  
Contact: Helen Xie  
Phone: 1(909) 859 - 9055  
Email: hx9055@gmail.com

### 2. Device Identification

Common name: Electrosurgical Generator and Accessories  
Classification name: Electrosurgical, Cutting & Coagulation & Accessories  
Trade name: Surn5 Electrosurgical Argon Beam Generator and Accessories  
Model: Surn5  
Device Class: Class II  
Product Code: GEI  
21CFR Regulation Number: 878.4400

### 3. Predicate Device

Product: HelixAR ABC Systems  
510(K): K172671

### 4. Device Description

Surn5 electrosurgical argon beam Generator and accessories have Monopolar Pure cut, Monopolar Bland Cut, Argon Beam, Bipolar functions. Surn5 electrosurgical argon beam Generator and accessories consists of HF generator and accessories. The device generates HF power, which is transmitted to the tissue through the output Pencil. Operator selects the output function and power with the front panel. The indicator is used to indicate the setup status and output status of the device. The output is controlled by foot or hand switch.

### 5. Indication for use

The Surn5 Electrosurgical Argon Beam Generator and Accessories are indicated for monopolar or bipolar surgery to achieve argon beam and cut or coagulation for the tissue. It is intended to be used with monopolar handpiece and dispersive electrode or bipolar handpiece and footswitches.

## 6. Substantial Equivalence Discussion

## Technological characteristics

Device	Subject device	Predicate device
Manufacturer	Surnic Corporation	Conmed Corporation
Model	Surn5 Electrosurgical Argon Beam Generator and Accessories	HelixAR ABC System
510(k) Number	K203522	K172671
Class	II	II
Product Code	GEI	GEI
Regulation Number	21 CFR 878.4400	21 CFR 878.4400
Mechanism of Action	The device generates high Frequency current 400-600 kHz	The device generates high Frequency current 385-600 kHz
System Voltage	100-240VAC- 50/60Hz	100-240VAC- 50/60Hz
4 Monopolar Cut Modes		
Max. Power	320 W (at 300Ω)	300 W (at 300Ω)
Wave Forms	Sinusoidal constant / modulated Cut / Coag Pulse Phases	Sinusoidal constant / modulated Cut / Coag Pulse Phases
8 Monopolar Coagulation Modes		
Max. Power	120 W (at 500Ω)	120 W (at 500Ω)
Wave Forms	Sinusoidal constant / modulated Cut / Coag Pulse Phases	Sinusoidal constant / modulated Cut / Coag Pulse Phases
Argon Beam		
Max. Power	150W (at 500Ω)	150W (at 500Ω)
4 Bipolar Modes		
Max. Power	100 W (at 100Ω)	90 W (at 300Ω)
Wave Forms	Sinusoidal constant / modulated Pause Phases	Sinusoidal constant / modulated Pause Phases

## 7. Non-Clinical Performance Data

To demonstrate the safety and effectiveness of Surn5 Electrosurgical Argon Beam Generator and Accessories show substantial equivalence to the predicate device (Surn5 Electrosurgical Argon Beam Generator and Accessories), Medrange completed successfully the following non-clinical performance tests:

General Safety Testing	IEC 60601-1:2005+C1+C2+A1:2012 Medical electrical equipment, Part 1 General requirements for basic safety and essential performance
EMC Safety Testing	IEC 60601-1-2:2014 (Edition 4.0) Medical electrical equipment, Part 1-2 General requirements for basic safety and essential performance Collateral Standard: Electromagnetic disturbances – Requirements and tests
HF Safety Testing	IEC 60601-2-2:2017 Medical electrical equipment, Part 2-2 Particular requirements for the basic safety and essential performance of high frequency - Surgical equipment and high frequency surgical accessories
Performance Testing	Applicable parts of FDA Guidance Document “Premarket Notification (510(k)) Submissions for Electrosurgical Devices for General Surgery”, May 2016
Usability Testing	IEC 60601-1-6:2010, AMD1:2013 Medical electrical equipment Part 1-6 General requirements for safety – Collateral Standard: Usability

## 8. Statement of Substantial Equivalence

The Surn5 Electrosurgical Argon Beam Generator and Accessories have the same intended use as the HelixAR, and the same technological characteristics. The non-clinical test results, such as performance data, software data, electrical safety and electromagnetic compatibility data have demonstrated that Surn5 Electrosurgical Argon Beam Generator and Accessories are as safe and effective as the predicate device. Therefore it is concluded, that the Surn5 Electrosurgical Argon Beam Generator and Accessories fulfill the requirements of a substantially equivalent device and that no new questions of safety and effectiveness were raised.