



April 8, 2022

Balt USA, LLC  
Michael Peters  
Manager, Regulatory Affairs  
29 Parker  
Irvine, California 92618

Re: K220699  
Trade/Device Name: Prestige Coil System  
Regulation Number: 21 CFR 870.3300  
Regulation Name: Vascular Embolization Device  
Regulatory Class: Class II  
Product Code: KR D  
Dated: March 8, 2022  
Received: March 10, 2022

Dear Michael Peters:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Finn Donaldson  
Acting Assistant Director  
DHT2C: Division of Coronary  
and Peripheral Intervention Devices  
OHT2: Office of Cardiovascular Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K220069

Device Name

Prestige Coil System

Indications for Use (Describe)

The Prestige Coil System is indicated for arterial and venous embolizations in the peripheral vasculature.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**K220069 510(k) Summary**

|                        |   |
|------------------------|---|
| <b>Applicant:</b>      | Balt USA, LLC<br>29 Parker<br>Irvine, CA 92618<br>Registration No.: 3014162263                                    |
| <b>Contact Person:</b> | Michael Peters<br>Manager, Regulatory Affairs<br>Telephone: (949) 788-1443<br>Email: michael.peters@baltgroup.com |

|                               |  |
|-------------------------------|--|
| <b>Date Summary Prepared:</b> | April 7, 2022  |
| <b>Trade Name:</b>            | Prestige Coil System   |
| <b>Common Name:</b>           | Vascular embolization device   |
| <b>Review Panel:</b>          | Cardiovascular   |
| <b>Product Code:</b>          | KRD  |
| <b>Regulation Number:</b>     | 21 CFR 870.3300  |
| <b>Regulation Name:</b>       | Device, Vascular, for Promoting Embolization   |
| <b>Device Classification:</b> | Class II   |
| <b>Predicate Device:</b>      | Prestige Coil System, 510(k)#: K200030<br>Originally cleared under trade name Optima Coil System |

**Device Description:**

The Prestige Coil System is a series of specialized coils that are inserted into the vasculature under angiographic visualization to embolize peripheral vascular anomalies. The system consists of an embolization coil implant comprised of platinum/tungsten, affixed to a delivery pusher to facilitate insertion into the hub of a microcatheter. The system is available in various shapes, lengths, and sizes. The devices are to be placed into anomalies to create blood stasis, reducing flow in the target vasculature, and inducing thrombosis. Upon positioning coils into the vasculature, the coils are thermally detached from the delivery pusher in serial manner until the target vasculature is occluded.

**Indications for Use:**

The Prestige Coil System is indicated for arterial and venous embolizations in the peripheral vasculature.

**Intended Use:**

The Prestige Coil System is intended for use in the peripheral vasculature to endovascularly obstruct or occlude blood flow in vascular abnormalities of the peripheral vessels.

**Device Comparison:**

|                            | <b>Predicate Device<br/>(K200030)</b>  | <b>Prestige Coil System<br/>(Subject Device, K220069)</b>  |
|----------------------------|--|--|
| Indications for Use        | The Optima Coil System is intended for the endovascular embolization of intracranial aneurysms and other neurovascular abnormalities such as arteriovenous malformations and arteriovenous fistulae. The Optima Coil System is also intended for vascular occlusion of blood vessels within the neurovascular system to permanently obstruct blood flow to an aneurysm or other vascular malformation and for arterial and venous embolizations in the peripheral vasculature. | Prestige Coil System is indicated for arterial and venous embolizations in the peripheral vasculature.<br><br>*These Indications for Use for the Prestige Coil System are a subset of those cleared for the Predicate.   |
| Intended Use               | The Optima Coil System is intended for use in the peripheral and neuro-vasculature to endovascularly obstruct or occlude blood flow in vascular abnormalities of the neurovascular and peripheral vessels.   | The Prestige Coil System is intended for use in the peripheral vasculature to endovascularly obstruct or occlude blood flow in vascular abnormalities of the peripheral vessels.<br><br>*This Intended Use for the Prestige Coil System is a subset of that cleared for the Predicate. |
| Anatomical Site            | Neurovasculature<br>Peripheral Vasculature   | Peripheral Vasculature<br><br>*The target anatomy for the Prestige Coil System is a subset of the cleared anatomy for the Predicate.   |
| Delivery to site           | Via delivery wire through microcatheter  | Same as K200030  |
| System Components          | Coil (implant)<br>Delivery System<br>Detachment Controller   | Same as K200030  |
| Method of Supply           | Sterile, single use  | Same as K200030  |
| <b>Coil (Implant)</b>      |  |  |
| Main Coil Material         | Platinum/Tungsten  | Same as K200030  |
| Primary Coil Wind Diameter | 0.010”-0.014”  | Same as K200030  |
| Coil Wire Diameter         | 0.00125”-0.0035”   | Same as K200030  |
| Coil Length                | 1-65cm   | Same as K200030  |
| Coil Secondary Diameter    | 1-24mm   | Same as K200030  |
| Number of Sizes Offered    | 209  | 267  |
| Secondary Shapes           | Complex, Helical   | Same as K200030  |
| Coil Types                 | Framing, Filling, Finishing  | Same as K200030  |

|  | <b>Predicate Device<br/>(K200030)</b>  |   | <b>Prestige Coil System<br/>(Subject Device, K220069)</b>   |   |
|--|--|---|---|---|
| No. & Size of Helical Loops on Distalmost End (Complex only) | 1.5 x 70%  |   | 1.5 x 70%<br>3 x 70%<br>3 x 100%  |   |
| Primary Wind (Coil OD) x Filar (Wire Diameter) combinations  | <u>Complex</u><br>.010" x .00125"<br>.011" x .0015"<br>.010" x .0015"<br>.011" x .00175"<br>.012" x .002"<br>.013" x .00225"<br>.014" x .003"<br>.014" x .0035"<br>.012" x .00125"<br>.014" x .0015"<br>.014" x .00175"<br>.014" x .002" | <u>Helical</u><br>.012" x .002"<br>.011" x .0015"<br>.010" x .00125"<br>.012" x .00125"<br>.014" x .0015"<br>.014" x .00175"<br>.014" x .002" | <u>Complex</u><br>.010" x .00125"<br>.011" x .0015"<br>.010" x .0015"<br>.011" x .00175"<br>.012" x .002"<br>.013" x .00225"<br>.014" x .003"<br>.014" x .0035"<br>.012" x .00125"<br>.014" x .0015"<br>.014" x .00175"<br>.014" x .002"<br>.014" x .00275"<br>.014" x .0025" | <u>Helical</u><br>.012" x .002"<br>.011" x .0015"<br>.010" x .00125"<br>.012" x .00125"<br>.014" x .0015"<br>.014" x .00175"<br>.014" x .002" |
| Stretch-resistance thread / attachment thread                | Polyolefin Engage Thread   |   | Same as K200030   |   |
| Coupler/Markerband   | Platinum/Iridium Alloy   |   | Same as K200030   |   |
| <b>Delivery System (Pusher)</b>                              |  |   |   |   |
| Construction/Design  | Body coil laser welded to hypotube   |   | Same as K200030   |   |
| Body Coil  | 4-part coil:<br>A. Heater Coil (92/8 Pt/W)<br>B. Distal Coil (SSTL)<br>C. Radio-opaque (RO, 92/8 Pt/W) Coil<br>D. Proximal Coil (SSTL)   |   | Same as K200030   |   |
| Hypotube   | SSTL Hypotube  |   | Same as K200030   |   |
| Connector  | Gold plated, SSTL Hypotube   |   | Same as K200030   |   |
| Adhesive   | Dymax 1128A-M-VT   |   | Same as K200030   |   |
| Jacket   | PET  |   | Same as K200030   |   |
| Fluorosafe Markers   | Pad Printed PET Shrink Tube  |   | Same as K200030   |   |
| Epoxy  | Epoxy 353 ND   |   | Same as K200030   |   |
| Lead Wires   | Polyimide coated silver lead wires   |   | Same as K200030   |   |

|            | <b>Predicate Device<br/>(K200030)</b> | <b>Prestige Coil System<br/>(Subject Device, K220069)</b> |
|------------|---------------------------------------|---|
|            | Other                                 |   |
| Detachment | Thermal via Detachment Controller     | Same as K200030   |
| Pouch      | Tyvek®                                | Same as K200030   |
| Shelf Life | 5 Years                               | Same as K200030   |
| Method     | Gamma Irradiation                     | Same as K200030   |

**Evaluations Leveraged from Predicate Device**

The following non-clinical device assessments were performed in support of the Predicate Device’s submission and clearance and have been evaluated as still valid based on the proposed changes to the Subject Prestige Coil System.

| <b>Test</b>             | <b>Acceptance Criteria</b>  | <b>Results</b> |
|-------------------------|---|----------------|
| Biocompatibility        | The samples shall be biocompatible for their intended use based on the requirements of ISO 10993-1.   | Pass           |
| Shelf Life              | The samples shall meet established acceptance criteria based on device specifications after simulated aging and transportation/distribution simulation. | Pass           |
| Packaging               | The samples shall meet established acceptance criteria for packaging performance.   | Pass           |
| Gamma Sterilization     | The samples shall meet established acceptance criteria for sterility.   | Pass           |
| Corrosion               | The samples shall meet established acceptance criteria for corrosion.   | Pass           |
| Detachment Zone Tensile | The samples shall meet established acceptance criteria for tensile strength.  | Pass           |

In addition to the above, the Magnetic Resonance Imaging Characterization undergone by the Predicate device has also been evaluated as still applicable to the Subject device and has not been reperformed.

**Performance Data – Bench:**

The following non-clinical bench testing was performed to demonstrate substantial equivalence of the Prestige Coil System to the Predicate device:

| <b>Test</b>                               | <b>Acceptance Criteria</b>  | <b>Results</b> |
|---|---|----------------|
| Visual and Dimensional Inspection         | The samples shall meet established acceptance criteria for visual physical damage and secondary diameter and length.  | Pass           |
| Simulated Use                             | The samples shall be prepared in accordance with the instructions for use and meet established acceptance criteria for device performance in a clinically relevant model.   | Pass           |
| Stretch-Resistance Thread Tensile Testing | The samples shall meet established acceptance criteria for tensile strength.  | Pass           |
| Usability Validation                      | The devices shall be prepared in accordance with their respective instructions for use and meet established acceptance criteria for device performance in a clinically relevant model. The labeling shall be clear and understandable by the intended user. | Pass           |

**Conclusion:**

There is no change to the risk profile, principles of operations or performance requirements of the Prestige Coil System in comparison to the Predicate Device (K200030). The successful completion of non-clinical bench testing demonstrates that the Prestige Coil System is substantially equivalent to the Predicate Device.