



Medacta International SA  
% Chris Lussier  
Director, Quality and Regulatory  
Medacta USA  
3973 Delp Street  
Memphis, Tennessee 38118

January 6, 2020

Re: K192827

Trade/Device Name: Quadra-P

Regulation Number: 21 CFR 888.3353

Regulation Name: Hip Joint Metal/Ceramic/Polymer Semi-Constrained Cemented Or Nonporous  
Uncemented Prosthesis

Regulatory Class: Class II

Product Code: LZO, MEH, KWY, LZY, LPH, JDI

Dated: September 30, 2019

Received: October 2, 2019

Dear Chris Lussier:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

FOR Vesa Vuniqi  
Assistant Director  
DHT6A: Division of Joint  
Arthroplasty Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K192827

Device Name

Quadra-P

Indications for Use (Describe)

The hip prostheses QUADRA-P and QUADRA-P collared are designed for cementless use in total or partial hip arthroplasty, for primary or revision surgery. The hip prosthesis QUADRA-P cemented is designed for cemented use in total or partial hip arthroplasty in primary or revision surgery. Hip replacement is indicated in the following cases:

- Severely painful and/or disabled joint as a result of arthrosis, traumatic arthritis, rheumatoid polyarthritis or congenital hip dysplasia
- Avascular necrosis of the femoral head
- Acute traumatic fracture of the femoral head or neck
- Failure of previous hip surgery: joint reconstruction, internal fixation, arthrodesis, partial hip arthroplasty, hip resurfacing replacement or total hip arthroplasty.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 2.0 510(k) Summary

### I. Submitter

Medacta International SA  
Strada Regina  
6874 Castel San Pietro (CH)  
Switzerland  
Phone (+41) 91 696 60 60  
Fax (+41) 91 696 60 66

Contact Person: Stefano Baj, Regulatory and Compliance Director, Medacta International SA  
Date Prepared: September 30, 2019

### II. Device

Device Proprietary Name:	Quadra-P
Common or Usual Name:	Femoral Stems
Classification Name:	Hip joint metal/ceramic/polymer semi-constrained cemented or nonporous un-cemented prosthesis
Primary Product Code:	LZO
Secondary Product Codes	MEH, KWY, LZY, LPH, JDI
Regulation Number:	21 CFR 888.3353, 21 CFR 888.3390, 21CFR 888.3360, 21 CFR 888.3358,
Device Classification	II

### III. Predicate Device

Substantial equivalence is claimed to the following devices:

Primary predicate device:

- Quadra-P, K181254, Medacta International SA

Secondary predicate device:

- Medacta Total Hip Prosthesis System – Quadra C (also referred to as “Quadra C”), K083558, Medacta International SA

The following devices are referenced:

- AMISem-H Proximal Coating, K121011, Medacta International SA
- AMISem-H Proximal Coating, AMISem-P and AMISem-P Collared, K173794, Medacta International SA

**IV. Device Description**

The Quadra-P implants are line extensions to the currently cleared Quadra-P implants (K181254).

The Quadra-P implants in this submission are comprised of the following products:

- Quadra-P STD (Stem size 00);
- Quadra-P Short Neck (STD Stem sizes 0 - 10 and LAT Stem sizes 0 – 10);
- Quadra-P Collared (STD Stem sizes 00 - 10 and LAT Stem sizes 0 – 10); and
- Quadra-P Cemented (STD Stem sizes 0 - 8 and LAT Stem sizes 0 – 8).

The Quadra-P implants are part of the Medacta Total Hip Prosthesis system. The Medacta Total Hip Prosthesis system consists of femoral stems, modular femoral heads, and acetabular components.

The Quadra-P, Quadra-P short neck, and Quadra-P Collared are cementless stems manufactured from Titanium-Niobium alloy with a Titanium plasma spray coating (MectaGrip) on the proximal area and HA coating on the shaft. The Quadra-P Cemented are cemented stems manufactured from high nitrogen stainless steel.

The acetabular components consist of metal cups and liners made of ultra-high molecular weight polyethylene (UHMWPE), or Highcross highly crosslinked ultra-high molecular weight polyethylene (HXUHMWPE). Acetabular components include the Mpace DM (K143453), VersafitCup (K083116 and K092265), VersafitCup CC Trio (K103352, K120521 and K122911), Mpace (K103721 and K132879), Mpace 3D Metal (K171966), and Medacta Bipolar Head (K091967).

The Quadra-P implants can be combined with the CoCr Ball Heads (K072857 and K080885), Endo Head (K111145), or MectaCer BIOLOX® Forte (K073337) or MectaCer BIOLOX® Delta Femoral Heads (K112115), or MectaCer Biolox Option Heads (K131518).

MectaCer Biolox Option Heads (K131518) cannot be combined with Quadra-P cemented stems, but only with the Quadra-P cementless stems.

**V. Indications for Use**

The hip prostheses QUADRA-P and QUADRA-P collared are designed for cementless use in total or partial hip arthroplasty, for primary or revision surgery. The hip prosthesis QUADRA-P cemented is designed for cemented use in total or partial hip arthroplasty in primary or revision surgery. Hip replacement is indicated in the following cases:

- Severely painful and/or disabled joint as a result of arthrosis, traumatic arthritis, rheumatoid polyarthritis or congenital hip dysplasia
- Avascular necrosis of the femoral head
- Acute traumatic fracture of the femoral head or neck
- Failure of previous hip surgery: joint reconstruction, internal fixation, arthrodesis, partial hip arthroplasty, hip resurfacing replacement or total hip arthroplasty.

## VI. Comparison of Technological Characteristics

The Quadra-P implants and the predicate devices share the following characteristics:

- CCD angle;
- materials of construction;
- coating;
- biocompatibility;
- device usage;
- sterility;
- shelf life; and
- packaging.

The Quadra-P femoral stems are technologically different from the predicate devices with respect to size and stem lengths.

## VII. Performance Data

Testing was conducted according to written protocols with acceptance criteria that were based on standards. The following mechanical studies were performed on worst-case stems in support of a substantial equivalence determination:

Non-Clinical Studies:

- Performance Tests
  - range of motion (ROM): *EN ISO 21535:2009 Non-Active Surgical Implants — Joint Replacement Implants — Specific Requirements for Hip-Joint Replacement Implants (ISO 21535:2007/Amendment 1:2016)*;
  - fatigue testing: *ISO 7206-4 Third Edition 2010-06-15 Implants For Surgery - Partial And Total Hip Joint Prostheses - Part 4: Determination Of Endurance Properties And Performance Of Stemmed Femoral Components [Including AMENDMENT 1 (2016)]*;
  - fatigue testing: *ISO 7206-6 Second Edition 2013-11-15 Implants For Surgery - Partial And Total Hip Joint Prostheses - Part 6: Determination Of Endurance Properties Of Head And Neck Region Of Stemmed Femoral Components*;
  - static fatigue testing: *ISO 7206-10:2003 Implants For Surgery -- Partial And Total Hip-Joint Prostheses -- Part 10: Determination Of Resistance To Static Load Of Modular Femoral Heads*; and
  - pull off force testing: ASTM F2009-00 (Reapproved 2011) *Standard Test Method For Determining The Axial Disassembly Force Of Taper Connections Of Modular Prostheses*.
- Pyrogenicity
  - Bacterial Endotoxin Test (LAL test) was conducted according to European Pharmacopoeia §2.6.14 (which is equivalent to USP chapter <85>) and pyrogen test according to USP chapter <151> for pyrogenicity determination.
  - the subject devices are not labeled as non-pyrogenic or pyrogen free.

Clinical Studies:

- No clinical studies were conducted.

**VIII. Conclusion**

The information provided above supports that the Quadra-P implants are substantially equivalent to the identified predicate devices.

Substantial equivalence has been demonstrated through a comparison of intended use, design and technological characteristics, as well as performance evaluations. The Quadra-P implants are as safe and effective as the predicate devices Quadra-P (K181254) and Quadra C (K083558).