



December 14, 2020

LimaCorporate S.p.A.
% Stephen Peoples
President
Peoples & Associates Consulting LLC
5010 Lodge Pole Lane
Fort Wayne, Indiana 46814

Re: K201084

Trade/Device Name: Physica system
Regulation Number: 21 CFR 888.3565
Regulation Name: Knee Joint Patellofemorotibial Metal/Polymer Porous-Coated Uncemented
Prosthesis
Regulatory Class: Class II
Product Code: MBH, JWH, HRY
Dated: November 9, 2020
Received: November 10, 2020

Dear Stephen Peoples:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Ting Song, Ph.D., R.A.C.
Assistant Director
DHT6A: Division of Joint
Arthroplasty Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K201084

Device Name
Physica system

Indications for Use (Describe)

Physica system is indicated for use in knee arthroplasty in skeletally mature patients with the following conditions:

- Non-inflammatory degenerative joint disease including
 - osteoarthritis
 - traumatic arthritis, and
 - avascular necrosis (not applicable to Physica TT Tibial Plate);
- Inflammatory degenerative joint disease including rheumatoid arthritis;
- Correction of functional deformity;
- Revision procedures where other treatments or devices have failed; and
- Treatment of fractures that are unmanageable using other techniques.

Additional indications for Physica LMC component are:

- Moderate varus, valgus, or flexion deformities.

In patients with preserved and well functioning collateral ligaments, Physica PS components are also indicated for:

- Absent or not-functioning posterior cruciate ligament;
- Severe antero-posterior instability of the knee joint.

Femoral, tibial and patellar components of the Physica system are intended for cemented use, with the exception of the TT Tibial Plate that is intended for uncemented use. Tibial liners can be used with cemented or uncemented tibial components.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Summary of Safety and Effectiveness

Date: December 14, 2020

Manufacturer:

Limacorporate S.p.A.
Via Nazionale, 52
33038 – Villanova di San Daniele
Udine - Italy

U.S. Contact Person:

Dr. Stephen J. Peoples
Principal Consultant
SPeoplesVMD@gmail.com
PEOPLES & ASSOCIATES
CONSULTING, LLC
411 Auditorium Blvd.
Winona Lake, IN 46590
Phone: 260-645-0327
FAX: +39 0432945512

Trade name: Physica system

Common name: Knee Prosthesis

Classification Name:

Product Code	Regulation and Classification Name
MBH	Knee joint patellofemorotibial metal/polymer porous-coated uncemented prosthesis per 21 CFR 888.3565
JWH	Knee joint Patellofemorotibial Polymer/Metal/Polymer Semi-Constrained Cemented Prosthesis per 21 CFR 888.3560
HRV	Knee joint Femorotibial Metal/Polymer Semi-Constrained Cemented Prosthesis per 21 CFR 888.3530

Description:

The Physica system is a total knee replacement system consisting of a Femoral component, Tibial plate, Tibial liner, Tibial stem and patella; the Tibial stem and Patella components are optional to be used as required for each individual patient.

Tibial liners and Patellar components made of conventional UHMWPE and cross-linked UHMWPE with Vitamin E are provided as part of the Physica system. The Tibial Plate is provided in cemented and uncemented versions.

Indications for use:

Physica system is indicated for use in knee arthroplasty in skeletally mature patients with the following conditions:

- Non-inflammatory degenerative joint disease including
 - osteoarthritis
 - traumatic arthritis, and
 - avascular necrosis (not applicable to Physica TT Tibial Plate);
- Inflammatory degenerative joint disease including rheumatoid arthritis;
- Correction of functional deformity;
- Revision procedures where other treatments or devices have failed; and
- Treatment of fractures that are unmanageable using other techniques.

Additional indications for Physica LMC component are:

- Moderate varus, valgus, or flexion deformities.

In patients with preserved and well functioning collateral ligaments, Physica PS components are also indicated for:

- Absent or not-functioning posterior cruciate ligament;
- Severe antero-posterior instability of the knee joint.

Femoral, tibial and patellar components of the Physica system are intended for cemented use, with the exception of the TT Tibial Plate that is intended for uncemented use. Tibial liners can be used with cemented or uncemented tibial components.

Predicate Devices:

Applicant	Device name	Cleared via
LIMACORPORATE S.P.A	PHYSICA SYSTEM	K141934, K193284
BIOMET MANUFACTURING CORP.	VANGUARD COMPLETE KNEE SYSTEM	K113550
ZIMMER, INC.	NEXGEN TRABECULAR METAL TIBIAL TRAY	K072160

Summary of technology comparison:

The intended use, design, and materials of the TT Tibial Plate and the tibial liners and patellar components manufactured from cross-linked UHMWPE with Vit. E are substantially equivalent to those of the predicate devices. Design Control Activities have been successfully completed.

Non-clinical testing:

Mechanical testing had demonstrated the device's ability to perform substantially equivalent to the predicate devices in:

- Wear test;
- Contact areas and pressures at tibio-femoral and patello-femoral interfaces;
- Test on the locking strength between the tibial plate and the tibial liner;
- Fatigue testing of the tibial plate;
- Fatigue testing of distal features of tibial plate.

Since the tibial liners and patellar components manufactured from cross-linked UHMWPE with Vit. E have an equivalent design and geometry to the predicate tibial liners and patellar components, the ROM of the subject devices when used with the previously cleared femoral components is the same as the ROM provided by the predicate Physica system tibial liners and patellar components.

Clinical testing:

Clinical testing was not necessary to demonstrate substantial equivalence of the subject devices to the predicate devices.

Conclusion:

Based upon a comparison of intended use, materials, summary of technological characteristics, and preclinical testing, the Physica system is substantially equivalent to the predicate devices identified in this premarket notification.