



March 25, 2021

SQ Medical Supplies Inc.  
% Shelly Li  
Official Correspondent  
Landlink Healthcare Technology (Shanghai) Co., Ltd.  
Room 703, 705, Building 1, West Guangzhong Road 555  
Shanghai, 200072  
China

Re: K203776

Trade/Device Name: Surgical Masks, Model SNN200640 & MN112  
Regulation Number: 21 CFR 878.4040  
Regulation Name: Surgical apparel  
Regulatory Class: Class II  
Product Code: FXX  
Dated: December 23, 2020  
Received: December 28, 2020

Dear Shelly Li:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Ryan Ortega, PhD  
Acting Assistant Director  
DHT4B: Division of Infection Control  
and Plastic Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K203776

Device Name  
Surgical Masks

### Indications for Use (Describe)

The Surgical Masks are intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single use, disposable device(s), provided non-sterile.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary: K203776

### I. Submitter

SQ Medical Supplies Inc.  
8889 West Olympic Blvd. Suite 1000 Beverly Hills, CA 90211.  
Establishment Registration Number: 3016757871

Contact person: Jonathan Lim  
Position: Chief Executive Officer  
Tel.: +1-888-912-8168  
E-mail: jonlim@sqmedicalsupplies.com

Preparation date: Feb. 26, 2021

### II. Proposed Device

Trade Name of Device: Surgical Masks  
Common name: Surgical Mask  
Regulation Number: 21 CFR 878.4040  
Regulatory Class: Class II  
Product code: FXX  
Review Panel: General Hospital

### III. Predicate Devices

510(k) Number: K160269  
Trade name: Surgical Face masks (Ear loops and Tie-on)  
Common name: Surgical Mask  
Classification: Class II  
Product Code: FXX  
Manufacturer: San-M Package Co., Ltd.

### IV. Device Description

The Surgical Masks are Flat Pleated type mask, utilizing Ear Loops way for wearing, and they all have Nose Piece design for fitting the face mask around the nose. The Surgical Masks are manufactured with three/four layers. The surgical mask has two models which are SNN200640 and MN112. They are basically the same, the only difference is the SNN200640 has the three layers and MN112 has the four layers.

The Surgical Masks are single use, disposable device, provided non-sterile.

**V. Indication for Use**

The Surgical Masks are intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single use, disposable device(s), provided non-sterile.

**VI. Comparison of Technological Characteristics with the Predicate Devices**

Table 10-1 General Comparison

Item		Proposed Device (K203776)	Predicate Device (K160269)	Decision
Trade Name		Surgical Masks	Surgical Face masks (Ear loops and Tie-on)	-
Product Code		FXX	FXX	Same
Regulation No.		21 CFR 878.4040	21 CFR 878.4040	Same
Class		Class II	Class II	Same
Mask Style		Flat-pleated, ear loops, 3/4 layers	Flat-pleated, ear loops or tie-on, 4 layers	Similar
Indication for Use		The Surgical Masks are intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single use, disposable device(s), provided non-sterile.	The Surgical facemasks are intended to be worn to protect the patient and healthcare personnel from transfer of microorganisms, blood fluid, and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluid. This is a single-use, disposable devices provided non-sterile.	Same
Material	Inner layer	White double spun-bond polypropylene	Polypropylene	Same
	Middle layer	Layer #1: Polypropylene Layer #2: Melt blown	Layer #1. Polypropylene spun-bond	Same

K203776 510(k) Summary

		polypropylene filter (Middle Layer #1 is not applicable SNN200640)	Layer #2: Polypropylene melt blown	
	Outer layer	Blue double spun-bond polypropylene	Polypropylene	Same
	Ear loops	Spandex+Nylon	Ear loops: Polyester, polyurethane; Side tapes: Polyester spun-bond (ear loops mask only) Ties tapes: Polypropylene spun-bond or polyester spun-bond	Different
	Nose piece	Steel wire coated by Polyethylene	Steel wire coated by Polyethylene	Same
	Color	Blue	Blue ,white	Similar
	Dimension	17.5cmx9.5cm	17.5cmx9.0cm 18.0cmx9.0cm	Same
	OTC Use	Yes	Yes	Same
	Sterility	Non-sterile	Non-sterile	Same
	For single Use	Yes	Yes	Same
	ASTM F2100 Level	Level 1: Pass at 80mmHg Level 3: Pass at 160mmHg	Level 1: Pass at 80mmHg Level 2: Pass at 120mmHg Level 3: Pass at 160mmHg	Similar
	Biocompatibility	Confirm to the requirements of ISO 10993 series standards	Confirm to the requirements of ISO 10993 series standards	Same

**VII. Non-Clinical Testing**

Non clinical performance tests were conducted to verify that the proposed device met all design specifications. The below table shows the test results of test article, which demonstrated that the proposed device complies with the standards of ASTM F2100-19:

Standard	Purpose	Acceptance Criteria		Results
		Level 1	Level 3	
ASTM F1862M-17	Fluid Resistance Performance	29 out of 32 pass at 80mmHg	29 out of 32 pass at 160mmHg	Pass
ASTM F2299	Particulate	≥95%	≥98%	Pass

---

K203776 510(k) Summary

---

	Filtration Efficiency			
ASTM F2101-19	Bacterial Filtration Efficiency	≥95%	≥98%	Pass
EN 14683:2019 Annex C	Differential Pressure	<5.0mmH <sub>2</sub> O/cm <sup>2</sup>	<6.0mmH <sub>2</sub> O/cm <sup>2</sup>	Pass
16 CFR 1610	Flammability	Class I non flammable		Pass

**VIII. Clinical Testing**

No clinical study is included in this submission.

**IX. Conclusion**

The conclusion drawn from the non-clinical performance testing data demonstrates that the subject device is as safe, as effective, and performs as well as or better than the predicate device, Surgical Face masks (K160269).