



June 7, 2021

Tuosheng Protective Products (Ningbo) Co., Ltd
% Charles Mack
Principal Engineer
Irc
2950 E Lindrick Drive
Chandler, Arizona 85249

Re: K210641

Trade/Device Name: Disposable Medical Mask
Regulation Number: 21 CFR 878.4040
Regulation Name: Surgical Apparel
Regulatory Class: Class II
Product Code: FXX
Dated: April 30, 2021
Received: May 5, 2021

Dear Charles Mack:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Clarence W. Murray III, Ph.D.
Acting Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K210641

Device Name

Disposable Medical Mask

Indications for Use (Describe)

The Disposable Medical Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. The face mask is intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single-use, disposable device(s) provided non-sterile. It is intended for adults only (greater than 21 years of age).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

K210641 510(k) SUMMARY

Preparation Date: June 3, 2021

Manufacturer's Name and Address: TUOSHENG PROTECTIVE
PRODUCTS (NINGBO) CO.,LTD
No 1728 Xiaolin Ave, North Industrial
Zone, Xiaolin Town, Cixi City, Zhejiang
Province, CHINA315321
Mr. Nanping Yuan
General Manager
Tel: 86-574-63795109
Fax: 86-574-63795109
Email: yuan@tuoshengprotective.com
FDA Establishment Registration No.:
3016844771

Designated Submission Correspondent and US Agent: IRC USA
2950 E Lindrick Dr., Chandler, Arizona
85249, USA
Mr. Charles Mack
PRINCIPAL ENGINEER
Tel: 931-6254938
Email: charliemack@irc-us.com

Telephone Number: 931-625-4938

Email Address: charliemack@irc-us.com

Trade Name: Disposable Medical Mask

Common Name(s): Disposable Medical Mask

Regulation Name(s): mask, surgical

Regulation Number(s): 21CFR878.4040

Product Code: FXX

Device Class: Class II

Predicate Device: Xiantao Rayxin Medical Products Co.,
Ltd.
Disposable Surgical Face Mask
K153496

Device Description:

The Disposable Medical Mask is a blue, flat pleated type mask, which utilizes an Earloop to Wear. It has a nose piece design to fit the facemask around the nose. The Disposable Medical Masks are manufactured with three layers and the inner and outer layers are made of spun-bond polypropylene, and the middle layer is made of melt-blown polypropylene filter. The model of the proposed device, earloop, is held in place over the user's mouth and nose by two elastic ear loops welded to the facemask. The elastic ear loops are not made with natural rubber latex. The nose piece in the mask is in the facemask layers to allow the user to fit the facemask around their nose, made of malleable aluminum wire. The Disposable Medical Masks are sold non-sterile and are intended to be a single-use, disposable device.

This product contains no components made from natural rubber latex.

Intended Use / Indications for Use

The Disposable Medical Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. The face mask is intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single-use, disposable device(s) provided non-sterile. It is intended for adults only (greater than 21 years of age).

Comparison of Technological Characteristics with the Predicate Device

Features	Subject Device	Predicate Device	Comparison
Device	Disposable Medical Mask	Disposable Surgical Face Mask	N/A
Model	Ear Loop	Ear Loop and Tie-On	N/A
Manufacturer	Tuosheng Protective Products (Ningbo) Co., Ltd.	Xiantao Rayxin Medical Products Co., Ltd.	N/A
510(k)	K210641	K153496	N/A
Indication for Use	The Disposable Medical Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids, and particulate material. The face mask is intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single-use, disposable device(s) provided non-sterile. It is intended for adults only (greater than 21 years of age).	The Disposable Surgical Face Masks are intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single-use, disposable device(s) provided non-sterile.	Similar
OTC or Prescription	OTC	OTC	Identical
Product Code	FXX	FXX	Identical
Classification	Class 2, CFR878.4040	Class 2, CFR878.4040	Identical
Material:			
Outer facing layer	Spun-bond polypropylene	Spun-bond polypropylene	Identical
Middle layer	Melt blown polypropylene filter	Melt blown polypropylene filter	Identical
Inner facing layer	Spun-bond polypropylene	Spun-bond polypropylene	Identical
Nose piece	Malleable aluminum wire	Malleable aluminum wire	Identical
Earloop	Polyurethane + Polyester	Polyester	Different
Colorant	Polypropylene (PP) master batch	Polypropylene (PP) master batch	Identical
Design features	Color: Blue Ear Loop	Color: Blue Ear Loop	Identical
Mask Style	Flat Pleated	Flat Pleated	Identical
Specification and Dimension	Length: 17.5 cm ± 0.85 cm Width: 9.5 cm ± 0.45 cm	Length: 17.5cm±1cm Width: 9.5cm±1cm	Identical

Features	Subject Device	Predicate Device	Comparison
Sterility	Non-sterile	Non-sterile	Identical
Usage	Disposable, Single Use	Disposable, Single Use	Identical
Performance:			
Performance Testing (ASTM F2100-19)	Level 2	Level 2	Identical
Fluid Resistance Performance ASTM F1862	32 out of 32 pass at 120mmHg	32 out of 32 pass at 120mmHg	Meets the ASTM F2100-19 Requirements for Level 2 Classification
Particulate Filtration Efficiency ASTM F2299	Average 99.37%	Average 98.46%	Meets the ASTM F2100-19 Requirements for Level 2 Classification
Bacterial Filtration Efficiency ASTM F2101	Average 99.54%	Average 98.70%	Meets the ASTM F2100-19 Requirements for Level 2 Classification
Differential Pressure (Delta P) EN 14683:2019, Annex C	4.00 mm H ₂ O/cm ²	4.2 mm H ₂ O/cm ²	Meets the ASTM F2100-19 Requirements for Level 2 Classification
Flammability 16 CFR 1610	Class 1 Non-Flammable	Class 1 Non-Flammable	Identical
Biocompatibility			
Cytotoxicity (ISO 10993-5)	Under the conditions of the study, not cytotoxic. Complies with ISO-10993-5.	Under the conditions of the study, not cytotoxicity effect	Identical
Sensitization (ISO 10993-10)	Under conditions of the study, not a sensitizer. Complies with ISO-10993-10.	Under conditions of the study, not a sensitizer.	Identical
Irritation (ISO 10993-10)	Under the conditions of the study, not an irritant. Complies with ISO-10993-10.	Under the conditions of the study, not an irritant	Identical

Discussion:

Although the materials used in the ear loops of the subject and predicate masks are not the same, the final subject masks are tested for biocompatibility and specification performance. Therefore, this difference does not raise any new questions about safety and effectiveness.

Although there is a slight difference in performance test results between the subject device and predicate device, they comply with the same performance standards, for ASTM F2100 Level 2. The minor differences in the technological characteristics do not raise issues on the safety and effectiveness of the subject device.

Performance Testing

Performance testing was provided to validate and verify that the Disposable Medical Mask, non-sterile, earloop met all requirements of related international standards, including biocompatibility and product specifications. These tests' results demonstrate compliance with the requirements of the consensus standards noted below.

Non-clinical Testing

Standards	Scope
ASTM F2100-19 Standard Specification for Performance of Materials Used in Medical Face Masks	Performance
ASTM F1862/F1862M-17 Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume at a Known Velocity)	Performance: Fluid Resistance Performance
ASTM F2101-19 Standard Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) of Medical Face Mask Materials, Using a Biological Aerosol of Staphylococcus aureus	Performance: Bacterial Filtration Efficiency (BFE)
ASTM F2299-2003 Standard Test Method for Determining the Initial Efficiency of Materials Used in Medical Face Masks to Penetration by Particulates Using Latex Spheres	Performance: Particulate Filtration Efficiency (PFE)
ASTM F2100-19 EN14683: 2019 Annex C	Performance: Differential Pressure (Delta P)
16 CFR 1610	Performance: Flammability
ISO10993-1:2009 Biological evaluation of medical devices--Part 1: Evaluation and testing	Biocompatibility
ISO10993-5:2009 Biological Evaluation of Medical Devices – Part 5 Tests for In Vitro Cytotoxicity.	Biocompatibility
ISO10993-10:2002/Amd. 1:2006(E) Biological Evaluation of Medical Devices – Part 10: Tests for Irritation and Delayed-Type Hypersensitivity	Biocompatibility
FDA Guidance: Surgical Masks - Premarket Notification [510(k)] Submissions; Guidance for Industry and FDA	Guidance

Test	Standards	Acceptance Criteria for Level 2 Mask	Test Result
Bacterial Filtration Efficiency (BFE, %)	ASTM F2101	≥ 98	Pass
Particulate Filtration Efficiency (PFE, at 0.1 μm, %)	ASTM F2299	≥ 98	Pass
Differential Pressure (Delta P, mm H ₂ O/cm ²)	ASTM F2100-19 EN 14683:2019, Annex C	< 6.0	Pass
Resistance to penetration by synthetic blood (minimum pressure in mmHg for pass result)	ASTM F1862/F1862M-17	≥ 120 mmHg	Pass
Flame Spread	16 CFR 1610	Class 1 Non-Flammable	Pass

Biocompatibility

The subject device is classified as a surface device and contact intact skin for limited contact duration.

We conducted the applicable tests noted below:

- In Vitro Cytotoxicity (ISO10993-5)
- Skin Sensitization (ISO10993-10)
- Skin Irritation (ISO10993-10)

Standard	Device Tests	Test Results
In Vitro Cytotoxicity (ISO10993-5: 2009)	Following the standard's defined conditions, the device is non-cytotoxic.	Pass
Skin Sensitization (ISO10993-10: 2010)	Following the standard's defined conditions, the device is non-sensitizing.	Pass
Skin Irritation (ISO10993-10)	Following the standard's defined conditions, the device is non-irritating.	Pass

All of the pre-determined acceptance criteria were met.

Clinical Test:

No clinical study is included in this submission.

Conclusions:

The conclusion drawn from the nonclinical tests demonstrates that the subject device in this 510(k) submission, K210641, Disposable Medical Mask, is as safe, as effective, and performs as well as or better than the legally marketed predicate device cleared under K153496.

END
