



Samsung Medison Co., Ltd.
% Scully KIM
Regulatory Affairs Specialist
3366, Hanseo-ro, Nam-myeon
Hongcheon-gun, Gangwon-do 25108
REPUBLIC OF KOREA

September 8, 2021

Re: K211945

Trade/Device Name: V8 Diagnostic Ultrasound System
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: Class II
Product Code: IYN, IYO, ITX
Dated: June 22, 2021
Received: June 23, 2021

Dear Scully KIM:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For

Thalia T. Mills, Ph.D.
Director
Division of Radiological Health
OHT7: Office of In Vitro Diagnostics
and Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K211945

Device Name

V8 Diagnostic Ultrasound System

Indications for Use (Describe)

The V8 diagnostic ultrasound system and probes are designed to obtain ultrasound images and analyze body fluids. The clinical applications include: Fetal/Obstetrics, Abdominal, Gynecology, Intra-operative, Pediatric, Small Organ, Neonatal Cephalic, Adult Cephalic, Trans-rectal, Trans-vaginal, Muscular-Skeletal (Conventional, Superficial), Urology, Cardiac Adult, Cardiac Pediatric, Thoracic, Trans-esophageal (Cardiac) and Peripheral vessel.

It is intended for use by, or by the order of, and under the supervision of, an appropriately trained healthcare professional who is qualified for direct use of medical devices. It can be used in hospitals, private practices, clinics and similar care environment for clinical diagnosis of patients.

Modes of Operation: 2D mode, Color Doppler mode, Power Doppler (PD) mode, M mode, Pulsed Wave (PW) Doppler mode, Continuous Wave (CW) Doppler mode, Tissue Doppler Imaging (TDI) mode, Tissue Doppler Wave (TDW) mode, ElastoScan Mode, Combined modes, Multi-Image mode(Dual, Quad), 3D/4D mode

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) Summary:

In accordance with 21 CFR 807.92 the following summary of information is provided:

1. Date Prepared – June 22, 2021
2. Manufacturer
SAMSUNG MEDISON CO., LTD.
3366, Hanseo-ro, Nam-myeon, Hongcheon-gun,
Gangwon-do, Republic of Korea
3. Primary Contact Person
Scully Kim
Regulatory Affairs Specialist
Phone: +82.2.2194.1312
Fax: +82. 2.2194.0273
Email: scully.kim@samsungmedison.com
4. Secondary Contact Person
Ninad Gujar
Vice President
Phone: +1.978.564.8632
Fax: +1.978.564.8677
Email: ngujar@neurologica.com
5. Proposed Device
 - Common/Usual Name : Diagnostic Ultrasound System and Accessories
 - Proprietary Name : V8 Diagnostic Ultrasound System
 - Regulation Name : Ultrasonic pulsed doppler imaging system
 - Regulatory Class : Class II
 - Product Code : IYN, IYO, ITX
 - Regulation Number : 21 CFR 892.1550, 892.1560, 892.1570
6. Predicate Devices
 - RS85 Diagnostic Ultrasound System (K192903) – Primary
 - HERA W9/ HERA W10 Diagnostic Ultrasound System (K192319) – Secondary
 - HS40 Diagnostic Ultrasound System (K210426) – Reference
 - LVivo Software Application (K200232) – Reference
 - UGEO PT60A Diagnostic Ultrasound System (K142466) - Reference
7. Device Description
The V8 is a general purpose, mobile, software controlled, diagnostic ultrasound system. Its function is to acquire ultrasound data and to display the data as 2D mode, Color Doppler mode, Power Doppler (PD) mode, M mode, Pulsed Wave (PW) Doppler mode, Continuous Wave (CW) Doppler mode, Tissue Doppler Imaging (TDI) mode, Tissue Doppler Wave (TDW) mode, ElastoScan Mode, Combined modes, Multi-Image mode(Dual, Quad), 3D/4D mode. The V8 also gives the operator the ability to measure anatomical structures

and offers analysis packages that provide information that is used to make a diagnosis by competent health care professionals. The V8 has real time acoustic output display with two basic indices, a mechanical index and a thermal index, which are both automatically displayed.

8. Intended Use

The V8 diagnostic ultrasound system and probes are designed to obtain ultrasound images and analyze body fluids.

The clinical applications include: Fetal/Obstetrics, Abdominal, Gynecology, Intra-operative, Pediatric, Small Organ, Neonatal Cephalic, Adult Cephalic, Trans-rectal, Trans-vaginal, Muscular-Skeletal (Conventional, Superficial), Urology, Cardiac Adult, Cardiac Pediatric, Thoracic, Trans-esophageal (Cardiac) and Peripheral vessel.

It is intended for use by, or by the order of, and under the supervision of, an appropriately trained healthcare professional who is qualified for direct use of medical devices. It can be used in hospitals, private practices, clinics and similar care environment for clinical diagnosis of patients.

Modes of Operation: 2D mode, Color Doppler mode, Power Doppler (PD) mode, M mode, Pulsed Wave (PW) Doppler mode, Continuous Wave (CW) Doppler mode, Tissue Doppler Imaging (TDI) mode, Tissue Doppler Wave (TDW) mode, ElastoScan Mode, Combined modes, Multi-Image mode(Dual, Quad), 3D/4D mode.

9. Technological Comparison to Predicate Devices

The V8 employs the same fundamental scientific technology as its predicate device RS85 (K192903) system.

10. Determination of Substantial Equivalence

Comparison to Predicate: The V8 is substantially equivalent to the predicate devices with regard to intended use, imaging capabilities, technological characteristics and safety and effectiveness.

- The systems are all intended for diagnostic ultrasound imaging and fluid flow analysis.
- The proposed V8 and predicate RS85 (K192903) have the same clinical intended use.
- The proposed V8 and predicate RS85 (K192903) have the same imaging modes and modes of operation.
- The Thoracic application has been included to system indications for use to support Lung scanning which is previously cleared in UGEO PT60A(K142466).
- The proposed V8 has migrated new SW feature (NerveTrack) based on AI and new transducer (CA4-10M) from predicate HS40 (K210426).
- The proposed V8 has included new SW feature (UterineAssist) based on AI. The UterineAssist is used for non-invasive processing of ultrasound images to detect, measure and calculate relevant medical parameters of uterus. It provides

similar indication for use of LVivo Software Application previously cleared (K200232).

- The proposed V8 has included two new transducers LA4-18A and LA3-22AI. The Biocompatibility and Image Performance Test have been conducted, and their test reports are included in this submission.
- The proposed V8 and predicate RS85 (K192903) have same capability in terms of performing measurements, capturing digital images, reviewing and reporting studies.
- The proposed V8 and predicate RS85 (K192903) have been designed in compliance with approved electrical and physical safety standards.
- The system is manufactured with materials which have been evaluated and found to be safe for the intended use of the device.
- The system has acoustic power levels which are below the applicable FDA limits.

11. Summary of Non-Clinical Testing

The device has been evaluated for acoustic output, biocompatibility, software function, cleaning and disinfection effectiveness as well as thermal, electrical, electromagnetic and mechanical safety, and has been found to conform with applicable FDA guidance and medical device safety standards. The V8 and its applications comply with the following FDA-recognized standards.

Reference No.	Title
IEC 60601-1	AAMI ANSI ES60601-1:2005/(R)2012 and A1:2012, C1:2009/(R)2012 and A2:2010/(R)2012 (Consolidated Text) Medical electrical equipment - Part 1: General requirements for basic safety and essential performance (IEC 60601-1:2005, MOD)
IEC 60601-1-2	IEC60601-1-2: 2014(4th Edition) , Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - EMC
IEC 60601-2-37	IEC 60601-2-37 Edition 2.0 2007, Medical electrical equipment – Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment
ISO10993-1	AAMI / ANSI / ISO 10993-1:2009/(R)2013, Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process
ISO14971	ISO 14971:2007, Medical devices - Application of risk management to medical devices
NEMA UD 2-2004	NEMA UD 2-2004 (R2009) Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment Revision 3

12. Summary of Clinical Tests

The proposed device V8 Ultrasound System did not require clinical studies to demonstrate substantial equivalence.

13. Conclusion

Since the predicate devices and subject device have a similar intended use and key technological features, the non-clinical data support the safety of the device and demonstrate that the V8 Ultrasound System should perform as intended in the specified use conditions. Therefore, SAMSUNG MEDISON CO., LTD. concludes that the performance of the subject device is as safe and effective, and is therefore substantially equivalent, to the primary predicate device that is currently marketed for the same intended use.

- **END of 510(k) Summary**